



South Carolina Electric & Gas Company
P.O. Box 88
Jenkinsville, SC 29065
(803) 345-4040

10CFR50.36

Ollie S. Bradham
Vice President
Nuclear Operations

September 12, 1990

08 SEP 18 P1:47

Mr. S. D. Ebnetter
Regional Administrator
U. S. Nuclear Regulatory Commission
Region II, Suite 2900
101 Marietta Street, N. W.
Atlanta, Georgia 30323

SUBJECT: Virgil C. Summer Nuclear Station
Docket No. 50/395
Operating License No. NPF-12
Special Report (SPR 90-011)

Dear Mr. Ebnetter:


This Special Report is being submitted by South Carolina Electric & Gas Company pursuant to the requirements of Technical Specification 3.3.3, Radiation Monitoring Instrumentation. On August 29, 1990, the Main Plant Vent High Range Process Monitor, RM-A13, was declared inoperable due to "spiking" of the monitor.

The Licensee has performed extensive troubleshooting on the unit, including the replacement of numerous electronic components, without resolving the spiking problem. A supplement report will be submitted when the problem is corrected detailing the deficiency and corrective action taken.

A preplanned alternate method of monitoring has been initiated as required by the Technical Specifications.

Should you require any additional information, please call at your convenience.

Very truly yours,


O. S. Bradham

RJB/OSB:lcd

c: O. W. Dixon, Jr./T. C. Nichols, Jr.
E. C. Roberts
R. V. Tanner
General Managers
J. J. Hayes, Jr.
C. A. Price
NRC Resident Inspector
J. B. Knotts, Jr.
Marsh & McLennan

Document Control Desk
INPO Records Center
ANI Library
NSRC
NPCF
RTS (ONO 880033)
File (818.05 & 818.07)

9010010042 900912
PDR ADUCK 05000395
S PDC

Teed
110