

DCB/DCB

TEL-FARMBROOK RADIOLOGY
29829 TELEGRAPH RD., SUITE L103
SOUTHFIELD, MI 48034
(313) 352-9525

September 12, 1990

United States Nuclear Regulatory Commission
Region III
730 Roosevelt Road
Great Ellyn, Illinois 60137

Attention: Roy J. Caniano, Chief
Nuclear Materials Safety

License No. 21-20152-01
Docket No. 030-18696

Response to 10 CFR 35.70 (e) violation:

CORRECTIVE STEPS: A weekly survey section has been added to our file of calibrations of the nuclear medicine room. This is in addition to the daily survey of the room which is done on the days scanning is performed.

Additional personnel have been trained to calibrate the nuclear medicine room when the regular technologist is on vacation and/or out of the office. This will assure that the room is monitored in accordance with 10 CFR 35.70 (e) guidelines.

Additional training of personnel in the department have been trained and informed of the requirements by NRC in monitoring the nuclear medicine room. Charts have been added to the calibration booklet used in daily calibration of the room when scans are performed.

Full compliance has been achieved as of this date, September 11, 1990.

Sincerely,



Roger B. Fenton, D.O./Medical Director



Sandra L. Berryman, R.T./R.D.M.S.
Supervising Technologist

SEP 14 1990

111
TE07



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
799 ROOSEVELT ROAD
GLEN ELLYN, ILLINOIS 60137

AUG 27 1990

Tel-Farmbrook Radiology
ATTN: Mrs. Vicki Royrke
Business Manager
29829 Telegraph Road
Southfield, MI 48034

License No. 21-20152-01
Docket No. 030-18696

Gentlemen:

This refers to the routine safety inspection conducted by J. W. Patterson and R. Ewing of this office on July 26, 1990, of activities authorized by NRC Byproduct Material License No. 21-20152-01 and to the discussion of our findings with you and Sandra Lee Berryman at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and with the conditions of your license. The inspection consisted of a selective examination of procedures and representative records, observations, independent measurements, and interviews with personnel.


During this inspection, certain of your activities appeared to be in violation of NRC requirements, as specified in the enclosed Notice. A written response is required.

In accordance with 10 CFR 2.790 of the Commission's regulations, a copy of this letter, the enclosure, and your response to this letter will be placed in the NRC Public Document Room.

The responses directed by this letter and the accompanying Notice are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

We will gladly discuss any questions you have concerning this inspection.

Sincerely,


Roy J. Caniano, Chief
Nuclear Materials Safety
Section 2

Enclosure: Notice of
Violation

cc w/enclosure:
DCD/DCB (RIDS)

~~900907223~~

NOTICE OF VIOLATION

Tel-Farmbrook Radiology
Southfield, MI 48034

Docket No. 030-18696
License No. 21-20152-01

As a result of the inspection conducted on July 26, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violation was identified:

10 CFR 35.70(e) requires that the licensee survey for removable contamination once each week in all areas where radiopharmaceuticals are routinely prepared for use, administered or stored.

Contrary to the above, no area surveys for removable contaminated were performed by the licensee in their imaging and preparation room for week's ending June 8 and 29 and July 6, 13 and 20, 1990.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

August 23, 1990
Dated

Roy J. Caniano
Roy J. Caniano, Chief
Nuclear Materials Safety
Section 2

9009100012