

ENDOCRINE DIAGNOSTICS

M. Rashid A. Khairi, M.D., F.A.C.P. r

Karen C. Rice, M.D.

7450 North Shadeland Avenue

Indianapolis, IN 46250

Telephone (317) 845-1006

August 13, 1990

George M. McCann, Chief
Nuclear Material Safety
Section 1

United States Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, IL 60137

Re: NRC letter dated 07/03/90
License No. 13-24511-01
Docket No. 030-28698

Item 1:

My medical assistant has recently attended an educational update in association with the Nuclear Med. Dept. at Community Hospital as follows:

Basic Radiation Physics (3 hrs)
Radiation Biology (2 hrs)
Radiation Protection (3 hrs)

We have made arrangements to be included in future updates with their facility.

Item 2:

Purchase of an extremity film badge. It should be noted that we were previously advised that this was not necessary for the low dose source we are using. However, in view of your current recommendation, a film badge has been ordered.

Item 3:

Will use the facility of Community Hospital to check for contamination prior to shipment of source and will be sure that our shipping papers include UN 2910.

I hope that the above complies with your letter of July 03, 1990.

Sincerely,

MRAK

M. Rashid A. Khairi, M.D.
MRAK/cip

RECEIVED

JUN 29 1990

REGION III

9010010129 900913
REG. LIC30
13-24511-01 ABC

AUG 2 1990