

APPENDIX A

NOTICE OF VIOLATION

City of Lynchburg

License No. 45-19651-01

As a result of the inspection conducted on August 9, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violation was identified.

License Condition 14 requires that each sealed source containing licensed material shall be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to the above, the sealed sources contained in your soil density gauge have not been tested for leakage and/or contamination since it was received in April 1981.

This is a Severity Level IV Violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) admission or denial of the alleged violation; (2) the reasons for the violation if admitted; (3) the corrective steps which have been taken and the results achieved; (4) corrective steps which will be taken to avoid further violations; and (5) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date: August 27, 1982