

Veterans
Administration

August 23, 1982

2 SEP 1 10:40



Mr. John A. Olshinski
United States Regulatory Commission
Region II
101 Marietta Street NW
Suite 3100
Atlanta, Georgia 30303

Dear Mr. Olshinski:


In reference to your letter of August 12, 1982, we wish to inform you that we have been in full compliance with the requirements of 10 CFR 35.14 since July 19, 1982.

On August 19, 1982 we telephoned the office and spoke with Mr. Potter confirming that we received the reprint of the inspection conducted by Mr. Chambliss. Corrective measures have been taken to ensure full compliance of the regulations.

We have performed the tests for leakage and/or contamination of the sealed sources used for calibration on July 19, 1982 and will continue at intervals not to exceed six months.

The reason for the violation is oversight. Corrective steps will be taken in the form of a master check-list of safety and quality control procedures in which we engage. Full compliance with this regulation has started and will continue.

Sincerely yours,


DAVID A. TULL, M. D.
Medical Center Director

8210010266 820916
NMS LIC30
01-03042-02 PDR

In Reply Refer To: 680/115

Official Copy