

APPENDIX A  
NOTICE OF VIOLATION

CIBA-GEIGY Corporation  
556 Morris Avenue  
Summit, New Jersey 07901

Docket No. 030-09359  
License No. 29-00459-03

As a result of the inspection conducted on August 8-9, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (Enforcement Policy) (1988), the following violation was identified:

Condition 19 of License No. 29-00459-03 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in letters dated October 6, 1983 and October 25, 1983, application dated October 26, 1988 and letter dated October 11, 1989.

Your letter dated October 6, 1983 indicates that your survey instrument calibration vendor (Teledyne Isotopes) will furnish certificates of calibration of your survey instruments on a yearly basis.

Contrary to the above, recent calibration records for two survey instruments (a mini monitor Series 900 and a Nuclear Chicago Model 2650) were not available nor by observation of instrument calibration stickers did it appear that either instrument had been calibrated since September 16, 1983.

This is a Severity Level IV violation. (Supplement IV)

Pursuant to the provisions of 10 CFR 2.201, CIBA-GEIGY Corporation is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY

ML DL CIBA-GEIGY - 0003.0.0

08/20/90

ML 10

RETURN ORIGINAL TO  
REGION 1

IE:07

9010030305 900904  
REG1 LIC30  
29-00459-03 PNU