

**Dennison**

September 16, 1982

**Technical  
Papers**

Mr. D. J. Sreniawski, Chief  
Materials Radiation Protection  
United States Nuclear  
Regulatory Commission  
Region III  
799 Roosevelt Road  
Glen Ellyn, Illinois 60137

Reference: Your letter of August 18, 1982

Subject: Notice of Violation  
License No. 21-18534-01 and General License

Dear Mr. Sreniawski:

Since the routine safety inspection conducted by Messrs. W. J. Slowinski and T. J. Ploski on August 10th, 1982, we have taken measures to achieve compliance with the NRC regulations concerning radioactive devices used by us in gauging applications.

I. With regard to Item 1 of the notice of violation:

- A. On August 19th, 1982, AccuRay Corporation tested our sealed source containing cesium-137 for contamination or leakage. The test results were negative; an informational copy is enclosed.
- B. Mr. Ronald Koglin has been appointed Radiation Safety Officer. He will have administrative control over the use of our radioactive gauging devices. The testing and record keeping programs required by our NRC license will be maintained by Mr. Koglin. Arrangements have been made with AccuRay Corporation to test this unit at 6-month intervals.
- C. The testing of the unit is complete and an on-going test program has been established. We expect administrative compliance to be accomplished within sixty days.


II. With regard to Item 2 of the notice of violation:

- A. The krypton-85 sealed source which was shipped to Foxboro Manufacturing, Foxboro, Mass., was received for disposal. Corrective action regarding the shipping papers and package labeling for this device does not seem appropriate.
- B. Mr. Ronald Koglin has been appointed Radiation Safety Officer. As such, he will have administrative control over the shipment and receipt of our radioactive gauging devices. Compliance with all NRC and DOT shipping requirements of our license will be maintained through his offices.
- C. We are not anticipating any shipments of radioactive material at the present time. Administrative compliance is expected to be complete within sixty days.

In the future, we will use the appropriate governmental regulations and assistance from the Nuclear Regulatory Commission, rather than advice from vendors, as our operating guidelines concerning radioactive devices.

If you should have any questions regarding the above information, please do not hesitate to contact me at your earliest convenience.

Very truly yours,



A. H. Hupp  
Senior Vice President  
Technical Operations



RADIOLOGICAL INSPECTION REPORT

This is a report of the inspection made of your radioisotope device and should be retained in a permanent file along with all other records of licensing or registration, receipt, installation, servicing and transfer of your radioactive material. Your regulatory authority may wish to review this information. Check your license or local regulations carefully.

17  
DUNN PAPER CO  
218 RIVERVIEW ST  
PORT HURON MI 48060

REPORT DATE: 820830

LAB TEST DATE: 820824

PERFORMED BY: K.J. HARRIS

ATTN. RON KOGLIN  
PLANT SITE: PORT HURON MI

Device Model	Device Serial Number	Source Serial Number	Isotope	Quan. (mCi)	Field Inspection Result				Lab Test Result
					Source	Shutter	Performed By	Date	
NR/LS-101	872307	CS-2150NR	CS137	25	NEG	OK	W. GERBER	820819	NEG

RECEIVED  
SEP 7 1982  
TECH. OPERS.

NOTES

1. NanoCurie (nCi) = .001 microCurie ( $\mu$  Ci) =  $10^{-6}$  milliCurie (mCi).
2. The entry "Neg" in the source column means less than 0.5 nanoCurie of removable contamination.
3. Any amount of detected activity greater than 0.5 nanoCurie is expressed in nanoCuries.
4. The entry "OK" in the shutter column means the shutter mechanism and indicators, if any, are operating properly, labeling is in proper condition, and the external radiation levels are consistent with those specified for the device. Discrepancies are detailed in appropriate notes.

5. The presence of 5 nanoCuries ( $005 \mu$  Ci) or more of removable contamination is considered evidence that the source is leaking. Refer to your regulatory requirements regarding leakage or malfunction.

*Eugene A. Stephens*  
DONALD C. STEPHENS  
RADIOLOGICAL OPERATIONS MANAGER