Dacket No. 030-03370 License No. 47-00404-02 Official Reins

Cabell Huntington Hospital ATTN: Mr. Don Smith, President 1340 Hal Greer Boulevard Huntington, WV 25701

Gentlemen:

SUBJECT: CONFIRMATION OF ACTION LETTER

This refers to a telephone conversation on September 28, 1990, between Mr. Paul Smith, Vice President, Legal Services and Risk Management, and Mr. William Cline of this office, concerning actions you will take to assure the safe use of byproduct material at Cabell Huntington Hospital.

With regard to the matters discussed, it is our understanding that you will take the following actions:

- Review your radiation safety procedures with all of your nuclear medicine personnel. The review will include but not be limited to the proper procedures for handling contamination incidents. You will ensure that personnel have a thorough understanding of the procedures.
- Implement procedures for dose calibrator constancy tests and area radiation 2. surveys to be conducted when nuclear medicine procedures are performed on weekends.
- Assure the counting system and method used to analyze wipe samples have sufficient sensitivity to detect 2000 disintegrations per minute per 100 square centimeters.

If your understanding of our discussion is different from that above, please contact this office promptly. This Confirmation of Action Letter does not preclude the NRC from issuing an Order, which could require additional specific action related to this matter.

Sincerely,

ORIGINAL SIGNED BY J. PHILLIP STOHR J. Philip Stohr, Director Division of Radiation Safety and Safeguards

CAL-030-03370/90-01

cc: State of West Virginia bcc: Document Control Desk

J. Lieberman, OE G. Jenkins, EICS

RII: DRSS RII: DRSS SWaldron CHosey CMEST 9/25/90 965/90

9010230198 901001 REG2 LIC30 47-00404-02 PD

RIM: DRSS

PDC

RII: DRSS **BMallett** NOT 9/\$ /90

RII: EICS