



DEPARTMENT OF VETERANS AFFAIRS  
Medical Center  
4500 South Lancaster Road  
Dallas TX 75216

In Reply Refer To: 549/141B

April 4, 1994

Docket No. 030-00503  
License No. 42-00220-08  
EA 93-217

Director, Office of Enforcement  
U.S. Nuclear Regulatory Commission  
ATTN: Document Control Desk  
Washington, DC 20555

**RE: ADDENDUM TO REPLY TO A NOTICE OF VIOLATION**

In your letter, dated March 9, 1994, you requested the following information to supplement our letter, dated October 25, 1993, in response to a Notice of Violation.

1. Our response to Violations I.A, I.B, and I.C, did not specify the date when full compliance was or will be achieved.

**RESPONSE:** Full compliance will be implemented by April 30, 1994, and prior to any clinical application of the Cobalt T-1000.

2. Our response to Violations I.B and I.C.1 did not describe the mechanism by which we will assure that training in provisions of our QM program is given to both current and future employees as appropriate, nor did it indicate the date by which full compliance will be achieved.

**RESPONSE:** Appropriate training in the QM program will be provided to current employees, including all authorized users and physics and technical staff, by April 30, 1994, and prior to any clinical application of the Cobalt T-1000. Appropriate training will be provided to future employees, including all authorized users and physics and technical staff, prior to directing, overseeing, or performing any clinical application of the Cobalt T-1000. Written evidence of such training will be retained indefinitely.

Alan G. Harper  
Medical Center Director

cc: Regional Administrator  
U.S. Nuclear Regulatory Commission, Region IV  
611 Ryan Plaza Drive, Suite 400  
Arlington, TX 76011

100063

940414027B 940404  
PDR ADOCK 03000503  
C PDR

IF07  
1/0