

Appendix

NOTICE OF VIOLATION

Bureau of Engineering
City of Akron, Ohio

License No. 34-20183-01

As a result of the inspection conducted on September 16, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violations were identified:

1. Condition No. 13 of your license requires that you test your sealed sources containing byproduct material for leakage and/or contamination at intervals not to exceed six months.

Contrary to this requirement, as of the day of the inspection, you have not tested your sealed source(s) containing cesium-137 and americium-241-beryllium for contamination or leakage since they were initially tested by the manufacturer in September 1981.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 18 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in application dated June 8, 1981, including the sample Radiation Safety Program.

The above referenced application, Item 15(b) supplement, states that locks shall be maintained on equipment to prevent accidental exposure of the sealed source when not under the direct supervision of authorized personnel.


Contrary to the above, on the day of the inspection, the source rod on your Troxler density gauge was not padlocked nor was the gauge under the direct surveillance of authorized personnel. Licensee representatives stated the source rod has not been padlocked since the unit was repaired in late October 1981.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

9/29/82


D. J. Sreniawski, Chief
Materials Radiation Protection
Section 2