

# CANSPEC

MATERIALS ENGINEERING AND TESTING

Docket No. 030-32380

License No. 29-28659-01

April 13, 1994

U.S. Nuclear Regulatory Commission  
Document Control Desk  
Washington, DC 20555

Re: Reply to Notice of Violation - Routine Inspection  
No. 030-32380/94-001

To Whom it May Concern:

Enclosed please find our responses pursuant to the provisions of 10CFR 2.201.

#### Violation A:

1. It was my misunderstanding that the audit conducted by the NRC would suffice as an internal audit of these individuals.
2. The individuals involved have since been successfully audited.
3. Notice to file that all individuals involved with radiography for Canspec will be audited internally regardless of NRC involvement.
4. Compliance was achieved immediately.

#### Violation B:

1. The meters in question were units shipped to me from Canada the dates on the meters as written in Canada was 11/8/93. American workers would see this date as November 8, 93. Thus fully believing they were using a calibrated instrument.
2. Any instrument received from Canada in the future will be fitted with a new calibration date written as it would be understood in the US.
3. The violation was discussed with all employees and they were asked to be aware of the possible misunderstanding and if there is any question about the correct date to contact a supervisor to confirm date.
4. Compliance was achieved immediately.

9404150201 940413  
PDR ADDCK 03032380  
C PDR


1507  
1/0  
CANSPEC TESTING INC.

Violation C:

1. In the transfer of documentation from our PA location to the office in New Jersey the training document was misplaced, and until this audit had gone undetected.
2. The violation was discussed with the individual involved and he concurred that the training had been given. The individual will write a quick note to his file confirming that the proper training had been given and a new form will be produced and signed by RSO and individual. This new form will be put in file.
3. Future documentation will be confirmed into the individuals file upon completion of training.
4. Full compliance was achieved immediately.

Thank you for your concern in these matters. We will make every attempt possible to follow NRC regulation in the future.

Sincerely,



Steven J. McCool  
RSO/Manager  
Canspec Testing Inc.

cc: Regional Administrator, Region I  
Barry Voit.

SJM/dmm