



CONVERSATION RECORD

NAME OF PERSON(S)/TITLE CONTACTED OR IN CONTACT WITH YOU		DATE OF CONTACT	TYPE OF CONVERSATION	
Alan M. Jackson, MS, CHP		2/07/2020	<input type="checkbox"/> E-MAIL <input checked="" type="checkbox"/> TELEPHONE <input type="checkbox"/> INCOMING <input checked="" type="checkbox"/> OUTGOING	
E-MAIL ADDRESS		TELEPHONE NUMBER		
AlanJ@rad.hfh.edu		(313) 916-2739		
ORGANIZATION		DOCKET NUMBER(S)		
Henry Ford Allegiance Health		030-01990		
LICENSE NAME AND NUMBER(S)		MAIL CONTROL NUMBER(S)		
Henry Ford Allegiance Health License No. 21-00258-06		617316		
SUBJECT				
Henry Ford Allegiance Health Amendment Request - Additional Information Required				
SUMMARY AND ACTION REQUIRED (IF ANY)				
<p>This is a summary of the conversations between Laura Cender and Alan Jackson of Henry Ford Allegiance Health regarding the amendment request dated December 20, 2019.</p> <p>Per our discussion today, additional information is required to continue processing the submitted amendment request. Please address each of the following items and provide your response by no later than Friday, February 28, 2020.</p> <ol style="list-style-type: none"> <li>1. Please confirm that PET isotopes will not be used at the locations of use requested with this license amendment.</li> <li>2. Please submit an additional copy of Diagram 1 that includes the following details:                     <ol style="list-style-type: none"> <li>a.) Clearly indicate the scale used for the diagram.</li> <li>b.) Indicate the direction North on the diagram.</li> <li>c.) Clearly indicate the use of areas adjacent to, above, and below the requested areas of use.</li> <li>d.) Provide room numbers, if available, for the locations requested.</li> </ol> </li> </ol>				
NAME OF PERSON DOCUMENTING CONVERSATION				
Laura B. Cender				
SIGNATURE			DATE OF SIGNATURE	
			02/07/2020	

**CONVERSATION RECORD (continued)**

LICENSE NAME AND NUMBER(S)

Henry Ford Allegiance Health  
License No. 21-00258-06

MAIL CONTROL NUMBER(S)

617316

**SUMMARY AND ACTION REQUIRED (IF ANY) (Continued)**

3. Please submit an additional copy of Diagram 1 that includes the following details:

- a.) Indicate the direction North on the diagram.
- b.) Clearly indicate the use of areas adjacent to, above, and below the requested areas of use.
- c.) Provide room numbers, if available, for the locations requested.

4. Please confirm that a separate license amendment request to decommission the existing SPECT imaging and nuclear stress rooms will be submitted once licensed activities cease in these areas.

**Cender, Laura**

---

**From:** Cender, Laura  
**Sent:** Monday, February 10, 2020 9:02 AM  
**To:** Jackson, Alan (RSO)  
**Subject:** RE: RE: Pending Amendment Request for NRC License No. 21-00258-06 - Additional Information Required

Hello Alan,

You can either fax the files to our regional office, or use a password to protect the document prior to sending it to me electronically. You can either send me the password in a separate email or give me a call if you choose to send the file electronically. I can be reached at 630-829-9712 or on my cell at 217-493-0429.

Please let me know if you have any questions.

Thank you,  
Laura

Laura Cender  
U.S. Nuclear Regulatory Commission  
Materials Licensing Branch  
E-mail: [Laura.Cender@nrc.gov](mailto:Laura.Cender@nrc.gov)  
Phone: (630) 829-9712

**From:** Jackson, Alan (RSO) <[AlanJ@rad.hfh.edu](mailto:AlanJ@rad.hfh.edu)>  
**Sent:** Monday, February 10, 2020 8:26 AM  
**To:** Cender, Laura <[Laura.Cender@nrc.gov](mailto:Laura.Cender@nrc.gov)>  
**Subject:** [External\_Sender] RE: Pending Amendment Request for NRC License No. 21-00258-06 - Additional Information Required

Laura,

I appreciate you sending this e-mail because then I can respond more quickly. I could send you scaled diagrams electronically as pdf files but I assume that the NRC IT people would object due to cyber security concerns.

-Alan

*Alan M. Jackson, MS. CAP*  
Radiation Safety Officer  
Henry Ford Hospital  
2799 West Grand Boulevard  
Detroit, MI 48202  
(313) 916-2739 Voice  
(313) 916-8456 Fax  
(734) 657-4133 Cell  
E-mail: [AlanJ@rad.hfh.edu](mailto:AlanJ@rad.hfh.edu)

**From:** Cender, Laura [mailto:[Laura.Cender@nrc.gov](mailto:Laura.Cender@nrc.gov)]

**Sent:** Friday, February 7, 2020 3:41 PM

**To:** Jackson, Alan (RSO) <[AlanJ@rad.hfh.edu](mailto:AlanJ@rad.hfh.edu)>

**Subject:** Pending Amendment Request for NRC License No. 21-00258-06 - Additional Information Required

Hello Alan,

Attached is a record detailing the additional information required to move forward with the requested license amendment.

To summarize the request please provide the following:

1. Confirmation that PET isotopes will not be used at the proposed areas of use.
2. Please resubmit Diagram 1 and Diagram 3 to include the additional details requested in accordance with NRC Guidance Document NUREG 1556 Vol. 9 Rev. 3.
3. Please confirm that a separate amendment request will be submitted to decommission the existing SPECT room and existing stress room.

Please submit your response by no later than Friday, February 28, 2020. Please ensure that your response is dated and signed by either licensee management or by the radiation safety officer named on the license. Please feel free to contact me at 630-829-9712 or send me an email if you have any questions.

Thank you,

Laura

Laura Cender

U.S. Nuclear Regulatory Commission

Materials Licensing Branch

E-mail: [Laura.Cender@nrc.gov](mailto:Laura.Cender@nrc.gov)

Phone: (630) 829-9712