



CONVERSATION RECORD

| | | | | |
|--|------------------------|-----------------|---|--|
| NAME OF PERSON(S)/TITLE CONTACTED OR IN CONTACT WITH YOU | | DATE OF CONTACT | TYPE OF CONVERSATION | |
| Cynthia Myers | | 01/28/2020 | <input type="checkbox"/> E-MAIL | <input type="checkbox"/> INCOMING |
| E-MAIL ADDRESS | TELEPHONE NUMBER | | <input checked="" type="checkbox"/> TELEPHONE | <input checked="" type="checkbox"/> OUTGOING |
| cynthia.r.myers@dupont.com | | (989) 636-4110 | | |
| ORGANIZATION | DOCKET NUMBER(S) | | | |
| DDP Specialty Electronic Materials US, Inc. | 030-39163 | | | |
| LICENSE NAME AND NUMBER(S) | MAIL CONTROL NUMBER(S) | | | |
| DDP Specialty Electronic Materials US, Inc. 21-35537-01 | 617225 | | | |
| SUBJECT | | | | |
| Request to Amend NRC License No. 21-35537-01 - Additional Information Required | | | | |
| SUMMARY AND ACTION REQUIRED (IF ANY) | | | | |
| <p>This is a record of the conversation between Laura Cender and Cindy Myers regarding the amendment request dated December 10, 2019. Please provide your signed and dated response by no later than Friday Feb. 14, 2020.</p> <ol style="list-style-type: none"> The letter dated December 10, 2019 described the byproduct source material as being tritium bound as Titanitrid with protective silicone seals. Please confirm that this Titanitrid product is in the form of a solid foil, and is not a gas. In Section 4 of your request you state that the device is designed for minimal user maintenance, but in Section 9 of your application you state that routine and non-routine analytical maintenance will be performed by in-house technicians. <ol style="list-style-type: none"> Please describe the activities that may be performed as part of routine and non-routine analytical maintenance. Please describe any maintenance, source replacement, cleaning, calibration, or other maintenance activities that would require technicians to interact directly with the source. Please note that Condition 22 of your license currently limits such activities to the device manufacturer or other licensed individuals. Please confirm that leak tests will be performed prior to receipt and disposal of the devices. | | | | |
| NAME OF PERSON DOCUMENTING CONVERSATION | | | | |
| Laura B. Cender | | | | |
| SIGNATURE | | | DATE OF SIGNATURE | |
| <i>Laura B. Cender</i> | | | 1/28/2020 | |

Cender, Laura

From: Cender, Laura
Sent: Tuesday, January 28, 2020 12:05 PM
To: Myers, Cynthia R
Subject: Pending Amendment of NRC License No. 21-35537-01 - Additional Information Requested
Attachments: Conversation Record to DDP Specialty Electronic Materials 1.28.2020.pdf

Hello Cindy,

Thank you for taking time out of your day to discuss your pending NRC license amendment request. As we discussed, a record of our conversation is attached.

Please provide your signed and dated response to the requested items by no later than Friday, February 14, 2020.

Thank you,
Laura

Laura Cender
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712