



FAIRVIEW GENERAL HOSPITAL

13101 LORAIN AVENUE / CLEVELAND, OHIO 44111 / (216) 476-7000

September 9, 1982

J. R. Miller, Chief
Technical Inspection Branch
United States Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Re: License No. 34-0216201
Visit by W. J. Slawinski on July 28, 1982

Dear Mr. Miller:

We have thoroughly reviewed the findings of Mr. Slawinski's visit, concerning the high film badge readings on Mrs. Betty Paul, with the Radiation Safety Office, the Administrator and Administrative Assistant of the Department of Radiology and Nuclear Medicine, and the Nuclear Medicine Technician. The findings were also the subject of an extended discussion during the most recent Radioisotope Committee meeting.

The following actions have been taken to preclude recurrence of an oversight of unusual or excessive badge readings:

1. Badge reports by R. S. Landauer will be initialed by the Assistant Radiology Administrator in charge of Nuclear Medicine, the Chief Technician in Nuclear Medicine, and the Radiation Officer before they are filed.
2. R. S. Landauer, Jr. & Co. customer service has been requested by letter to automatically alert us to high badge readings. (copy enclosed)
3. Our current arrangement involves a quarterly visit of this department by a physicist to review procedures, records and film badge readings, and to check our compliance with ALARA. The physicist then discusses his findings at the Radioisotope Committee Meeting, which is scheduled the same day. This procedure virtually precludes that any unusual badge readings would escape our notice. Unfortunately our association with this group of physicists, Nuclear Medicine Associates, Inc., and this review system had not yet been instituted in early 1980.

8210060407 821001
NMS LIC30
34-02162-01 PDR

1982 — OUR 90TH YEAR OF SERVICE — 1982

SEP 27 1982


The following actions have been taken concerning your notice of violation found during your July 28, 1982 inspection, regarding Condition Number 16 of our license:

D.O.T. Yellow II labeled packages are being surveyed at the package surface and at three feet from the package upon receipt in the department, and a permanent record is kept on the survey results.

A Record Form was devised. Surveys have been made and records have been kept and are available for inspection as of August 2, 1982.

Please contact me if any additional information is required.

Sincerely,



Kenneth T. Misener
Associate Administrator

KTM/sr



FAIRVIEW GENERAL HOSPITAL

18101 LORAIN AVENUE / CLEVELAND, OHIO 44111 / (216) 476-7000

R.S. Landauer, Jr. & Co.
Glenwood Science Park
Glenwood, Illinois 60425

July 29, 1982

Dear Customer Service:

I talked to Mrs. Margaret Voytina in regards to one of your features on your dosimetry reports.

That is you reporting directly to us on individuals who have an over-exposure on one of their badge inserts (GI badge).

Conferring with Dr. Juergen Schapiro, we decided that we would like to have a letter sent to us when someone has a badge reading exceeding 200 millirem. On individuals who exceed 1000 millirem we would like a cablegram sent to us.

The individuals to contact for any of the above incidents would be:

Mr. Rick Lange
Director of Radiology Services
or
Mr. Raymond Pellow
Assistant Director of Radiology Services

Thank you for your cooperation.

Raymond C. Pellow, R.T.
Assistant Director of
Radiology Services

RP.