
P.O. BOX 201 / TULSA, OKLAHOMA 74102 / (918) 599-2000 / TWX 910-845-2106

November 20, 1990

U.S. Nuclear Regulatory Commission Attn: Document Control Desk Washington, D.C. 20555

Reply to a Notice of Violation (Notice)

Docket No: 030-13633/90-07 License No: 35-07038-02



Corrective Action:

As soon as the violation was brought to our attention, a leak test was performed. The results were: Beta Gamma - O. Alpha - O. analysis dated 9-4-90. by Troxler Electronics Laboratory, P.O. Box 12057, Research Triangle, N.C. 27709.

The problem has been discussed with the manager of fuel, the supervisor at Northeast 3 & 4 and the Radiation Safety Officer. The following procedure has been established to prevent future violations.

- 1. The lock on the storage cabinet where the instrument is will be changed.
- 2. The only key will be kept in the possession of the Radiation Safety Officer.
- 3. The Radiation Officer upon request for the key, will go to site to determine if leak test is required. If testing is required, it will be tested and a key issued to the cabinet only after the results of the leak test analysis indicates it safe to use the instrument.
- 4. The key may only be checked out for one week intervals and must be returned to the Radiation Safety Officer at end of each check-out period.
- 5. The lock will be changed on or before December 14, 1990.
- 6. The only keys to the lock will be in the possession of the Radiation Safety Officer by December 14,1990.

Sincerely,

Stephen Davenport

Supervisor, Central Lab, Public Service of Okla.

CENTRAL AND SOUTH WEST SYSTEM

xc:Regional Administrator, Region IV Nuclear REgulatory Commission

Suite 1000

012050021 901120 164 LIC30 5-07038-02 PD

Central Power and Light Public Service Company of Oklahoma Tulsa Oklahoma

| Radionuclide: CS 137 | |
|---|----|
| Source Serial: CC 5363 | |
| Inst Model. | |
| Inst Serial: 9-19-00 | |
| Individual's Name: Stephen Davengon | et |
| Inst Serial: Date of Wipe: 8-29-90 Individual's Name: Stephen Daver por Telephone: 918-599-2327 | |
| PLEASE TYPE OR PRINT LEGIBLY - | |

THIS IS YOUR RETURN ADDRESS LABEL

Leak Test Analysis Removable Activity Alpha Beta Gamma

- Public Service OKLAhomA
- Box 201
- Tuis. 1, OK 74102
 - ATTN: Stephen Davenport

tokier Electronic Exporatories, Inc. P.O. Box 12057, Research Triangle Park, N. C. 27709-919-545-8661 Terix 5814502 TROXI, U.O.

NOTES

- 1. Follow procedures as defined in your leak test kit instructions
- 2. Fill out this form and the bad label with required information where applicable. Seat the filter paper in the plastic bag. Place the plastic bag and this form in the pre-addressed envelope
- 3. Removable activity will be reported in Ci A value of 0 indicates less than 00005 Ci
- 4. Federal and state regulations require that sealed sources be removed from service and reports filed if removable activity is greater than .005 . Cr
- 5. Due to the potential hazard. Troxler recommends that an additional wipe be made if removable activity exceeds .0005 LCi.
- 6. You will be notified by telephone collect if the test yields greater than .001 LCi removable activity.