

APR 8 1994

Milwaukee Diagnostic Services  
ATTN: Mary J. Mommaerts  
President  
3033 South 27th Street  
Milwaukee, WI 53215

License No. 48-26493-01  
Docket No. 030-33145

Dear Ms. Mommaerts:

This refers to the routine safety inspection conducted by W. P. Reichhold of this office on February 23 and 25, 1994 and continuing NRC review through March 23, 1994, of activities authorized by NRC Byproduct Material License No. 48-26493-01, and to the discussion of our findings with Mr. Thomas Jacobson at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and with the conditions of your license. The inspection consisted of a selective examination of procedures and representative records, observations, independent measurements, and interviews with personnel.

During this inspection, certain of your activities were found to be in violation of NRC requirements, as specified in the enclosed Notice. A written response is required.

The following three violations were also identified: (1) quarterly sealed source inventory and linearity checks were not signed by the Radiation Safety Officer; (2) documents required by 10 CFR 19.11 were not posted; and (3) the survey meter was not capable of measuring 1,000 millirem per hour. However, we are using discretion as allowed by 10 CFR Part 2, Appendix C, Section VIIB in the NRC Enforcement Policy and are not including these items in the Notice of Violation because you committed to take appropriate corrective actions, the items are categorized as Severity Level V violations, and the remaining criteria in the aforementioned Enforcement Policy are satisfied.

In accordance with 10 CFR 2.790 of the Commission's regulations, a copy of this letter, the enclosure, and your response to this letter will be placed in the NRC Public Document Room.

APR 8 1994

Milwaukee Diagnostic Services -2-

The response directed by this letter and the accompanying Notice are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

We will gladly discuss any questions you have concerning this inspection.

Sincerely,

Original Signed by B. J. Holt

B. J. Holt, Chief  
Nuclear Materials Inspection  
Section 1

Enclosure: Notice  
of Violation

bcc w/enclosure:  
PUBLIC

RIII <sup>YES</sup> 4/14/94  
Reichhold/bt  
03/ /94

RIIB  
Holt  
4/14/94