



DEPARTMENT OF THE ARMY

LOS ANGELES DISTRICT, CORPS OF ENGINEERS  
P.O. BOX 2711  
LOS ANGELES, CALIFORNIA 90053-2325

April 15, 1994

REPLY TO  
ATTENTION OF:

Safety & Occupational  
Health Office

Mr. Gregory P. Yuhas  
Chief, Nuclear Regulatory Commission Region V  
Radioactive Material Safety Branch  
Attention: Gregory P. Yuhas, Chief  
1450 Maria Lane  
Walnut Creek, CA 94596-5368

Dear Mr. Yuhas:

In reference to your letter dated 28 Mar 94, Notice of Violation Docket Number 030-20427, regarding item 9 of the license renewal dated 22 Jan 92, the "CAUTION" label has been placed on the device and is in legible condition.


The device was not in use at the time of inspection and had not been used for the 5 months prior to the inspection. The contractor was not on the site on a regular basis during this period of time. The device was basically in storage.

The "CAUTION" label is in place and in legible condition. Additional "WARNING RADIOACTIVE MATERIALS" signs have been placed on the device, and a new check list to include shuttle position has been established between the contractor and the Radiation Safety Officer. Full compliance has been achieved as of 15 Mar 94.

If I can be of further assistance, please feel free to contact me at (213) 894-5328 or Dr. Ronald J. Hawley, at (602) 640-2076.

Sincerely,

FOR THE COMMANDER:

  
Jerome J. Dittman  
Lieutenant Colonel  
Corps of Engineers  
Deputy District Engineer

CF:  
US Nuclear Regulatory Commission  
ATTN: Document Control Desk  
US Nuc Reg Comm Reg V  
ATTN: Mr. Kent M. Prendergast

9404200179 940415  
PDR ADOCK 03020427  
C PDR

IF07  
1/0