

NOTICE OF VIOLATION

Emma L. Bixby Medical Center  
Adrian, MI 49221

License No. 21-03194-01  
Docket No. 030-02027

As a result of the inspection conducted on November 6, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violations were identified:

1. 10 CFR 35.22(a)(1) require that the membership of the Radiation Safety Committee consist of at least three individuals and must include an authorized user of each type of use permitted by the license, the Radiation Safety Officer, a representative of the nursing service and a representative of management who is neither an authorized user nor the Radiation Safety Officer.

Contrary to the above, since the first quarter of 1990, the membership of the Radiation Safety Committee has failed to include a representative of management.

This is a repeat violation.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 14 states that the license is based on the licensee's statements and representations listed. The listed application dated October 18, 1978 (received October 3, 1978) states in Item No. 15 that individuals are required not to eat, drink, smoke or apply cosmetics in any area where radioactive materials are used or stored.

Contrary to the above, on the date of this inspection, the inspector observed an individual drinking coffee in the nuclear medicine room, an area where radioactive materials are used and stored.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Date

November 18, 1990

Roy J. Canzano  
Roy J. Canzano, Chief  
Nuclear Materials Safety  
Section 2

9012050160 901120  
REGS LIC30  
21-03194-01 PDC