

Quality Assurance Laboratories, Inc.

AN INDEPENDENT NON-DESTRUCTIVE TESTING LABORATORY

80 PLEASANT AVENUE

SOUTH PORTLAND, MAINE 04106

Telephone (207) 799-8911 or (207) 799-7251

October 10, 1990

USNRC
Region 1
475 Allendale Road
King Of Prussia, Pennsylvania 19406

Attn: John R. White

Docket No. 030-16028
License No. 18-19078-01


Subject: Response to routine inspection No. 90-001

In reference to appendix A-Notice of violation, it was documented that we were not in compliance with 10CFR 34.29(b) because of the fact that we had the gamma alarm on a switch that was energized outside of the gamma cell. Please review our corrective action in this matter.

1. As of 10/9/90, we have hard wired the gamma alarm into the circuitry so it is on constantly and can't be switched off.

We hope this meets your approval. If there are any questions, please don't hesitate to call.

Sincerely,



Arthur S. Gallant, R.P.O.

X-RADIOGRAPHY . GAMMA RADIOGRAPHY . ULTRASONICS . MAGNETIC PARTICLE . DYE PENETRANT . EDDY CURRENT

ON SITE

OFF SITE

9012100140 901128
REG 1 LIC 30
18-19078-01 FDC