

June 14, 1990

U.S. Nuclear Regulatory Commission
Region 1
475 Allendale Road
King of Prussia, PA 19406
ATTN: John R White, Chief
Nuclear Materials Safety Section C
Division of Radiation Safety and Safeguards

Subject: Docket No. 030-19419
License No. 18-20564-01
Routine Inspection No. 030-19419/89-01

Dear Mr. White:

This is to advise you that, belatedly, we have made the corrections required by your report of inspection, Appendix A. Specifically:

Item A

We have applied to NRC for a license amendment naming the Morrison Geotechnical Engineering training program in addition to the manufacturer's training program. I will send you a copy of the amended license when it is received.

Item B

We are maintaining records of leak tests and physical inventory of gages, both accomplished at 6 months intervals.

Item C

We have provided each licensed material source with a transparent envelope containing:

- 1) Shipping paper naming the source material types and stating their activities
- 2) Bill of lading showing the category of label applied to the package, the transport index, and the hazard class number
- 3) Additionally; package certification and emergency procedures

The envelopes are located in plain sight on either the dashboard or the driver's door of the transport vehicle.

9012100144 901128
REG1 LIC30
18-20564-01 PIC

207/774-0983


PORTLAND, MAINE 04102

THOMPSON'S POINT

PAGE TWO
June 14, 1990
John R White

I believe that once our license amendment is received we will be in full compliance with your requirements. I expect to receive the amended license by August 15, 1990

Yours Very Truly,



Decatur H. Cousins
President

DHC:emb

June 14, 1990

U.S. Nuclear Regulatory Commission
Region 1
475 Allendale Road
King of Prussia, PA 19406
ATTN: John R White, Chief
Nuclear Materials Safety Section C
Division of Radiation Safety and Safeguards

Subject: Docket No. 030-19419
License No. 18-20564-01
Routine Inspection No. 030-19419/89-01

Dear Mr. White:

This is to advise you that, belatedly, we have made the corrections required by your report of inspection, Appendix A. Specifically:

Item A

We have applied to NRC for a license amendment naming the Morrison Geotechnical Engineering training program in addition to the manufacturer's training program. I will send you a copy of the amended license when it is received.

Item B

We are maintaining records of leak tests and physical inventory of gages, both accomplished at 6 months intervals.

Item C

We have provided each licensed material source with a transparent envelope containing:

- 1) Shipping paper naming the source material types and stating their activities
- 2) Bill of lading showing the category of label applied to the package, the transport index, and the hazard class number
- 3) Additionally; package certification and emergency procedures

The envelopes are located in plain sight on either the dashboard or the driver's door of the transport vehicle.

207/774-0983


PORTLAND, MAINE 04102

THOMPSON'S PCANT

PAGE TWO
June 14, 1990
John R White

I believe that once our license amendment is received we will be in full compliance with your requirements. I expect to receive the amended license by August 15, 1990

Yours Very Truly,



Decatur H. Cousins
President

DHC:emb