

APPENDIX A

NOTICE OF VIOLATION

Osteoporosis Center of Maine, Inc.
Portland, Maine 04102

Docket No. 030-29315
License No. 18-27858-01

As a result of the inspection conducted on November 9, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (Enforcement Policy) (1990), the following violation was identified:

License Condition 13.A. of NRC License No. 18-27858-01 requires that the sealed source be tested for leakage and/or contamination at intervals not to exceed one year.

Contrary to the above, as of November 9, 1990, a one curie Gadolinium-153 sealed source had not been tested for leakage and/or contamination since August 4, 1989, a time interval exceeding one year.

This is a Severity Level IV violation. (Supplement IV)

Pursuant to the provisions of 10 CFR 2.201, Osteoporosis Center of Maine is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY

ML DL OSTEOPOROSIS - 0003.0.0
11/23/90

9012100178 901128
REG1 LIC30
18-27858-01 PDC