

NOTICE OF VIOLATION

Southwestern Indiana Cancer Center
Evansville, Indiana

Docket No. 030-30712
License No. 13-25945-01

As a result of the inspection conducted on November 13 and 16, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violation was identified:

10 CFR 20.203(c)(1) requires that each high radiation area be conspicuously posted with a sign or signs bearing the radiation caution symbol and the words: "Caution High Radiation Area."

Contrary to the above, on November 13 and 16, 1990, the cobalt-60 teletherapy room, a high radiation area, was not posted as required.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

NOV 30 1990

Dated _____

William H. Schultz
William H. Schultz, Chief
Nuclear Materials Safety
Section 1

9012110045 901130
REG3 LIC30
13-25945-01 PDC