



80 SEYMOUR STREET
P.O. BOX 5037
HARTFORD, CT 06102-5037
203/545-5555

To: USNRC, Region I
Jenny M. Johansen, Medical Inspections
475 Allendale Road
King of Prussia, PA 19406

March 4, 1994

From: Mr. Peter J. Mas, M. S., Health Physicist & RSO

Re: Reply to a Notice of Violation; Routine Inspection # 030-01239/93-001

The following letter is a reply to the followup letter from Ms. Jenny M. Johansen, dated Feb. 9, 1994. Your letter requests that we address incident response and radiation safety oversight.

Incident Response:

The Radiation Safety Office will provide for a waste monitoring station, equipped with an alarming ratemeter, for the automatic monitoring of the routine, non-biomedical wastes that exit the hospital. This device should identify the contaminated wastes that would become a potential problem if removed off-site. Additionally, the hospital's Security Office has a "Call List" of physicists that will handle radiation incidents on or off site.

Radiation Safety Oversight:


The Health Physicist & RSO job description will be reevaluated. Also, a time study of the office and RSO's workload may be conducted for this reevaluation.

If you should have any further questions, please do not hesitate to contact us. We appreciate your efforts and we look forward to a safe and effective program with radioactive materials.

Respectfully submitted,

Mr. Peter J. Mas, M. S.
Health Physicist & RSO

9404260083 940412
PDR ADOCK 03001239
C PDR



80 SEYMOUR STREET
P.O. BOX 5037
HARTFORD, CT 06102-5037
203/545-5555

To: USNRC, Region I
Jenny M. Johansen, Medical Inspections
475 Allendale Road
King of Prussia, PA 19406

March 4, 1994

From: Mr. Peter J. Mas, M. S., Health Physicist & RSO

Re: Reply to a Notice of Violation; Routine Inspection # 030-01239/93-001

The following letter is a reply to the followup letter from Ms. Jenny M. Johansen, dated Feb. 9, 1994. Your letter requests that we address incident response and radiation safety oversight.

Incident Response:

The Radiation Safety Office will provide for a waste monitoring station, equipped with an alarming ratemeter, for the automatic monitoring of the routine, non-biomedical wastes that exit the hospital. This device should identify the contaminated wastes that would become a potential problem if removed off-site. Additionally, the hospital's Security Office has a "Call List" of physicists that will handle radiation incidents on or off site.

Radiation Safety Oversight:

The Health Physicist & RSO job description will be reevaluated. Also, a time study of the office and RSO's workload may be conducted for this reevaluation.

If you should have any further questions, please do not hesitate to contact us. We appreciate your efforts and we look forward to a safe and effective program with radioactive materials.

Respectfully submitted,

A handwritten signature in cursive script that reads 'Peter J. Mas'.

Mr. Peter J. Mas, M. S.
Health Physicist & RSO