

OCT 5 1982

APPENDIX A

NOTICE OF VIOLATION

Columbia Hospital for Women
Washington, D. C. 20037

Docket No. 30-10159
License No. 08-15994-01

As a result of the inspection conducted on August 16, 1982, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

- A. 10 CFR 19.12 requires that all individuals working in restricted areas be instructed in the applicable provisions of the Commission's regulations and your specific license.

Contrary to the above, as of August 16, 1982, a technologist working in the restricted area had not been instructed in the applicable provisions of the regulations and the conditions of the license, including the requirement to wear gloves while handling radioactive materials.

This is a Severity Level IV violation. (Supplement VI)

- B. Condition 15 of License No. 08-15994-01 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in an application dated May 11, 1979.

1. Item 15 of this application requires that disposable gloves be worn at all times while handling radioactive materials.

Contrary to the above, as of August 16, 1982, personnel in the Nuclear Medicine Radioimmunoassay Laboratory have never worn disposable gloves when handling radioactive materials.

This is a Severity Level IV violation. (Supplement VI)

2. Item 10 of this application requires that survey meters be calibrated at least annually and following repair.

Contrary to the above, as of August 16, 1982, the survey meter had not been calibrated since April 6, 1981, a period of more than one year.

This is a Severity Level IV violation. (Supplement VI)

- C. 10 CFR 30.51(a) requires that each licensee keep records showing the receipt, transfer, export, and disposal of licensed material.

Contrary to the above, as of August 16, 1982, the required records of licensed material received and disposed of by transfer to an authorized recipient had not been maintained since May 16, 1982.

This is a Severity Level V violation. (Supplement IV)

OFFICIAL RECORD COPY

OCT 5 1982

Pursuant to the provisions of 10 CFR 2.201, Columbia Hospital for Women is hereby required to submit to this office within thirty days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violation; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

OFFICIAL RECORD COPY