

NOTICE OF VIOLATION

University of Wisconsin-Parkside
Kenosha, Wisconsin

License No. 48-15342-02
Docket No. 030-12999

As a result of the inspection conducted on October 26, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violations were identified:

1. License Condition No. 20 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

The referenced letter dated April 19, 1989, states on pages 13 and 14 that non-technical personnel shall be instructed, upon hire and annually, in the precautions and procedures to minimize exposure to radioactive materials.

Contrary to the above, as of the date of this inspection, non-technical personnel have not received the required instruction upon hire and/or annually.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 12.C requires that records of leak test results to be kept in units of microcuries and maintained for inspection by the Commission.

Contrary to the above, results of leak tests performed in January and September 1990 were not in units of microcuries. Further, a record of the leak test performed in May 1990 was not maintained for review during this inspection.

This is a Severity Level V violation (Supplement VI).

3. 10 CFR 19.11(c) requires the licensee to post Form NRC-3, Notice to Employees.

License Condition No. 20 states that the licensee shall conduct its program in accordance with statements representations and procedures contained in certain referenced applications and letters. The referenced letter dated April 19, 1989 states that Safety Instructions (Appendix A) and Emergency Procedures (Appendix B) must be posted in each laboratory where radioactive Materials are used.


Contrary to the above, on the date of this inspection, the Form NRC-3, Safety Instructions and Emergency Procedures were not posted in laboratories 105, 375 or 373.

This is a Severity Level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

NOV 28 1990

Dated _____



William H. Schultz, Chief
Nuclear Materials Safety
Section 1