

Mr. Jerry A. Frohlich
President
Mrs. Geraldine A. Siero
Vice President of Patient Services
Mr. Donald R. Nagy
Vice President of Human Resources

North Detroit General Hospital ^{DCD}

3105 CARPENTER AVENUE
DETROIT, MICHIGAN 48212

TELEPHONE: (313) 369-3000

December 4, 1990

Nuclear Regulatory Commission
799 Roosevelt Road
Glen Ellyn, Illinois 60137

ATTENTION: Mr. A. Bert Davis
Regional Administrator

RE: License No.: 21-10578-02

Dear Mr. Davis:

The following is in clarification of our reply to "Notice of Violation" as requested by Mr. Roy Caniano in our conversation this afternoon.

- (1) Item #2 a question regarding the door closure arose. The door to the Nuclear Area is on an automatic door closure which is self locking. The RSO will survey the area on a regular basis to check to ensure that when no one is in the area, the door is closed and locked.
- (2) Item #3 pertains to eating and smoking policies within the department. The employees were retrained and to the smoking and eating policies at their in-service meeting. The RSO will continuously observe to see whether the policy is being adhered to during his daily visits to review the log books.
- (3) Item #5 refers to the dose calibrator and its usage. We are currently using daily check sheets that will immediately tell us plus or minus 10% of normal requirements. The RSO will check on his daily visits to the department to make sure these forms are used. If the dose calibrator shows a reading of plus or minus 10%, they will immediately cease using the equipment. This was stressed in the in-service meeting recently held with the department.
- (4) Item #6 - the RSO will review the calibration sheets obtained from the physicist. It will be the responsibility of the RSO to ensure that the worksheets are received from the physicist for our files.

I trust that the above clarification will assist you in your


9012180075 901204
REGG LIC30
21-10578-02 PDR

17.08

IE07
110
DEC 10 1990

Nuclear Regulatory Commission
Page 2

Thank you very much.



Jerry A. Frohlich
President

JAF/bjj

cc: Director, Office of Enforcement
U.S. Nuclear Regulatory Commission
ATTENTION: Document Control Desk
Washington, D.C. 20555