

SAINT VINCENT HOSPITAL, INC. WORCESTER, MASSACHUSETTS 01604 - TELEPHONE 508-798-1234

Docket # 030-01954

License # 20-12869-01

October 2, 1990

Mohamed M. Shanbaky, Chief Nuclear Materials Safety Section A Division of Radiation Safety and Safeguards Nuclear Regulatory Commission Region 1 475 Allendale Road King of Prussia, Pennsylvania 19406

Dear Mr. Shanbaky:

Your letter of September 25, 1990 indicates a need to address preventative actions regarding the inspection on June 27, 1990. The day following the inspection an In-Service meeting was held to instruct all of the nuclear medicine technologists. A record of this meeting with attendance is retained in our In-Service log.

To insure continued compliance the Radiation Safety Officer routinely inspects the records of the necessary site monitoring, dose calibrator constancy checks and byproduct storage. Records of the inspection are kept. I am unaware of further preventative action that could be taken and hope this meets the Commission's requirements.

Sincerely;

w. fil

deffrey Kelly Associate Vice President Professional Services

Melvin H. Farmelant, M.D., Chairman Radiation Safety Committee & RSO

mb

901	219	0063	901210
REG 20-	128	1030	PDR