

NOTICE OF VIOLATION

Indiana University School of Medicine

Docket No. 030-09673
License No. 13-15734-01

As a result of the inspection conducted on November 14, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violation was identified:

License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced applications and letters.

Item 1. of the referenced letter dated July 9, 1986 states that survey meters are calibrated annually.

Contrary to the above, the licensee failed to calibrate a survey meter, Serial No. 1002, from January 13, 1986 to March 28, 1989, and failed to calibrate a survey meter, Serial No. 2134, from January 28, 1986 to March 12, 1990. Both meters were periodically used during those intervals.

This is a repeat violation.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

DEC 7 1990

Dated _____

William H. Schultz

William H. Schultz, Chief
Nuclear Materials Safety
Section 1

9012190110 901207
REG3 LIC30
13-15734-01 PDR