

NOTICE OF VIOLATION

Caylor-Nickel Hospital
Bluffton, Indiana

License No. 13-01629-04
Docket No. 030-00192

As a result of the inspection conducted on November 1, 1990, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1989) (Enforcement Policy) the following violations were identified:

1. a. 10 CFR 35.632(b)(4) states that the full calibration will include a check of the timer linearity.
- b. 10 CFR 35.634(a)(1) states that the monthly spot checks will include a check of timer linearity.

Contrary to the above:

- a. The full calibration performed on January 26, 1990, did not include a check of the timer linearity.
- b. The monthly spot checks from February to July 1990 did not include a check of the timer linearity.

This is a Severity Level IV violation (Supplement VI) and a repeat violation.

2. License Condition No. 14 states that the license is based on statements and representations contained in certain referenced applications and letters.

The referenced letter dated February 21, 1986, states that personnel will be trained as described in Appendix H of NRC Regulatory Guide TM 608-4. Regulatory Guide TM 608-4 states that personnel who work in the vicinity of the teletherapy unit should receive training on the hazards of radioactive material and the safety procedures appropriate to their respective duties.

Contrary to the above, the housekeeping personnel did not receive training on the hazards of radioactive material and the safety procedures appropriate to their duties since the license was issued on February 25, 1988.

This is a Severity Level IV violation (Supplement VI).

9012200032 901211
REG3 LIC30
13-01629-04 PDR

Notice of Violation

2

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

DEC 11 1990

Dated _____

William H. Schultz

William H. Schultz, Chief
Nuclear Materials Safety
Section 1