

DCD/DLB

NOV 07 1990

Dodge Health Services
ATTN: John Landdeck
President
709 University Avenue
Beaver Dam, WI 53916

Gentlemen:

As a result of the inspection conducted on October 24, 1990, a NRC Form 591, SAFETY INSPECTION, is issued. The enclosed form sets forth the violation noted. Please acknowledge receipt of this form by signing and dating in the appropriate space on all copies. You are requested to retain the original and return four signed and dated copies to this office within ten days.

I wish to express my appreciation for the cooperation extended to me during the inspection.

Sincerely,

James L. Cameron
Radiation Specialist

Enclosure: NRC Form 591

RIII
JLC
Cameron/dsv
11/2/90

RIII
WPS
Schultz
11-6-90

9012280080 901212
REG3 LIC30
48-24610-01 PDR

IE07

SAFETY INSPECTION

1. LICENSEE Dodge Health Services 709 University Avenue Beaver Dam, WI 53916	2. REGIONAL OFFICE U.S. NUCLEAR REGULATORY COMMISSION REGION III 799 ROOSEVELT ROAD GLEN ELLYN, IL 60137	
3. DOCKET NUMBER(S) 030-18673	4. LICENSE NUMBER(S) 48-24610-01	5. DATE OF INSPECTION 10/24/90

Licensee:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission's (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The findings as a result of this inspection are as follows:

- 1. Within the scope of this inspection, no violations were observed.
- 2. The inspector also verified the steps you have taken to correct the violations identified during the last inspection. We have no further questions on these actions at this time.
- 3. During this inspection certain of your activities, as checked below, were in violation of NRC requirements.
THIS IS A NOTICE OF VIOLATION which is required to be posted in accordance with 10 CFR 19.11.

A. _____ was not properly posted to indicate the presence of _____, 10 CFR 20.203(b), (c), (d), (e) or 34.42.

B. Containers located in _____ were not properly labeled to indicate the presence of radioactive material. 10 CFR 20.203(f)(1), or (f)(2).

C. _____ of sealed sources were not performed at the proper frequencies. 10 CFR _____ License Condition Number _____

D. Records of _____ were not properly maintained. 10 CFR _____ or License Condition Number _____

E. Documents were not properly posted or otherwise made available. 10 CFR 19.11.

F. Reports or notifications of _____ were not made in accordance with 10 CFR _____ or License Condition Number _____

H. Shipping papers for the transport of licensed material were not properly completed as required by 10 CFR 71.56

I. Shipping papers did not include the proper shipping name of the material transported (i.e. Radioactive Material, N.O.S.)

J. Nor the Emergency Telephone Number.

K. _____

I hereby state that within 30 days the actions described by me to the inspector will be taken to correct the violations identified in the items checked above. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201. No further response will be submitted unless required by the NRC.

Golda BDDTB
10

James L. Cameron

10/31/90