

APPENDIX
NOTICE OF VIOLATION

Oklahoma City - County Health Department
Oklahoma City, Oklahoma

Docket No. 030-09350/90-01
License No. 35-15563-01

During an NRC inspection conducted on November 29, 1990, violations of NRC requirements were identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (1990), the violations are listed below:

1. License Condition 12.A requires, in part, that sealed sources shall be tested for leakage and/or contamination at intervals not to exceed 6 months.

Contrary to the above, the chromatograph detector cell (Serial No. 4591) containing a Ni-63 foil source was not tested for leakage and/or contamination at intervals not to exceed 6 months between May 1983 and December 1988.

This is a Severity Level IV violation. (Supplement VI)

2. 10 CFR 30.51(a) requires, in part, that each licensee keep records showing the receipt, transfer, export, and disposal of licensed material.

Contrary to the above, as of the date of this inspection, a record documenting the receipt of Ni-63 detector cells (Serial Nos. 4591 and 5249) had not been maintained.

This is a Severity Level IV violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Oklahoma City - County Health Department, is hereby required to submit to this office, within 30 days of the date of the letter transmitting this Notice, a written statement or explanation in reply, including for each violation: (1) the reason for the violation if admitted, (2) the corrective steps which have been taken and the results achieved, (3) the corrective steps which will be taken to avoid further violations, and (4) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending the response time.

Dated at Arlington, Texas,
this 24th day of December 1990

9101020409 901224
REG4 LIC30 PDR
35-15563-01