

OCT 26 1982

The Blood Center of Central Iowa
ATTN: Mr. Larry Weldy
Laboratory Director
1050 Seventh Street
Des Moines, IA 50314

License No. 14-18765-01
30-14146

Gentlemen:

Thank you for your letter dated October 11, 1982, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated September 28, 1982.

This letter also refers to the telephone conversation between you and Mr. W. P. Reichhold of my staff on October 21, 1982, concerning your corrective actions.

It is our understanding that the following additional actions will be taken:


1. Additional individuals will be added to your license to ensure that your program always has an authorized user.
2. A quality control check list will be used to ensure that the ash residue surveys are performed.
3. An audit of the waste surveys will be performed weekly.

If our understanding of the above is incorrect, please contact this office.

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,



D. G. Wiedeman, Chief
Materials Radiation Protection
Section 1

IE HQ FILE COPY

8211060134 821026
NMS LIC30
14-18765-01 PDR

cc w/ltr dtd 10/11/82:
DMB/Document Control Desk (RIDS)

OFFICE	RIII	RIII	RIII		
SURNAME	Douglas:rr	Reichhold	Wiedeman		
DATE	10/25/82	10/24/82			

DMB:IE07