

OCT 2 2 1982

Docket Nos. 30-02959
70-01362

License Nos. 37-00467-34
SNM-1369

Hahneman Medical College and Hospital
ATTN: Lewis C. Mills, M.D.,
Vice President for Health Affairs
230 N. Broad Street
Philadelphia, Pennsylvania 19102

Gentlemen:

Subject: Inspection No. 82-01

This refers to the routine safety inspection conducted by J. Johansen and Mr. R. Ladun of this office on August 12, and 13, 1982 of activities authorized by NRC License Nos. 37-00467-34 and SNM-1369 and to the discussions of our findings held by Ms. Johansen with yourself and members of your staff at the conclusion of the inspection, and to a subsequent telephone discussion between Mrs. Sharon Steinberg of your staff and Ms. Johansen on August 16, 1982.

The inspection was an examination of activities conducted under your licenses as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your licenses. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, measurements made by the inspector, and observations by the inspector.

Based on the results of this inspection, it appears that certain of your activities were not conducted in full compliance with NRC requirements, as set forth in the Notice of Violation, enclosed herewith as Appendix A. These violations have been categorized by severity level in accordance with the NRC Enforcement policy (10 CFR 2, Appendix C) published in the Federal Register Notice (47 FR 9987) dated March 9, 1982. You are required to respond to this letter; and, in preparing your response, you should follow the instructions in Appendix A.

From the telephone discussion on August 16, 1982, it is our understanding that you will assign a TLD ring badge for the left and right hand of your Nuclear Medicine technologist who is ambidextrous in order to determine which dosimeter most adequately evaluates the technologist's extremity exposure.

The responses directed by this letter and the accompanying Notice are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter and your reply will be placed in the Public Document Room.

8211060150 821022
PDR ADDCK 07001362
C PDR

OFFICIAL RECORD COPY

1E:07

OCT 22 1982

Your cooperation with us in this matter is appreciated.

Sincerely,

Original Signed By:
John D. Kinneman

for Thomas T. Martin, Director
Division of Engineering and Technical
Programs

Enclosure: Appendix A, Notice of Violation

cc:
John Day, Ph.D., Radiation Safety Officer
Public Document Room (PDR)
Nuclear Safety Information Center (NSIC)
Commonwealth of Pennsylvania

bcc:
Region I Docket Room (with concurrences)
Chief, Operational Support Section (w/o encl.)

for RE: DETP
Johansen/jpf
10/7/82

for RI: DETP
Ladun
10/21/82

for RI: DETP
Kinneman
10/21/82
OFFICIAL RECORD COPY