



CENTRAL STEEL DIVISION

GARY WORKS
1 NORTH BROADWAY
GARY, INDIANA 46402

October 8, 1982

Mr. James R. Miller, Chief
Technical Inspection Branch
United States Nuclear
Regulatory Commission
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Re: Notice of Violation
License No. 37-02375-17
License No. 13-07964-05
Dated: 8-27-82

Dear Mr. Miller:

Pursuant to the telephone conversation of October 6, 1982, between Mr. J. B. Masaitis, Chairman of United States Steel Corporation Radiation Committee, and Messrs. Lasuk and Weidman of your office, this letter clarifies the additional action we are taking regarding Items 3 and 4 of the Notice of Violation, dated August 27, 1982.

The corrective action that we are taking to avoid future non-compliance with the violations discussed in Items 3 and 4 is as follows:

Item 3:

We will maintain a log-record of survey instrument calibrations which will include:

1. Dates when instruments are next due for calibration.
2. Dates when instruments were calibrated.

Timely notification to the responsible departments will be made approximately 30 days prior to the next due date for survey instrument calibrations.

Item 4:

We will maintain a log-record of wipe tests for our devices which will include:

1. Dates when devices are next due for wipe test.
2. Dates when wipe tests are completed.

Timely notification to responsible personnel will be made approximately 30 days prior to the due dates for the next wipe tests.

8211060184 821028
NMS LIC30
37-02375-05 PDR

OCT 12 1982



J. R. Miller

October 8, 1982

In addition to the above, Plant and Corporation Radiation Committees will monitor the plant radiation program by conducting periodic audits.

Very truly yours,

E. W. Sieger RM

E. W. Sieger
General Manager