

WA. BOADE M.D.,
LIMITED

Diagnostic Medical Imaging

REC'D
DEC 27 1990

December 19, 1990


Mr. A. Bill Beach
Director Division of Radiation Safety & Safeguards
United States Nuclear Regulatory Commission
Region IV
611 Ryan Plaza Drive
Suite 1000
Arlington, TX 76001

Re: Notice of violation response. Docket #30-30273/90-02
License #40-26908-01

Pursuant to provision 10CFR 2.201 please allow this to
constitute written reply to violation severity level V
violation.

Prior to October 26, 1990, inspection by your office it
was our organizations policy to record swipe test results
in CPM's contrary to your regulation stating they should
be in DPM's. Following receipt of your notice of
violation, we instituted a change in our format and all
swipe tests of our mobile Tomo units are now recorded in
DPM's per 100 cm. squared per regulatory requirements.

It is my hope that through this action the foresighted
violation has been corrected.


Roger M. Rae
Radiation Safety Officer
W.A. Boade M.D., Ltd.

9101070166 901219
REG4 LIC30
40-26908-01 PDR

1700 South Euclid Avenue • Sioux Falls, South Dakota 57117 • (605) 335-1455

IC 453

IE-07