

Official

DEC 31 1990

32-30003

Hospital Alejandro Otero Lopez
ATTN: Mr. Milton Maldonado
Executive Administrator
P. O. Box 1142
Manati, PR 00701

Gentlemen:

SUBJECT: NRC REPORT NO: 52-24916-01/90-01

Thank you for your response of December 3, 1990, to our Notice of Violation issued on November 9, 1990, concerning activities conducted under NRC Byproduct Material License No. 52-24916-01. We have evaluated your response and found that it meets the requirements of 10 CFR 2.201.

With regard to Violation B, your response indicated that the Radiation Safety Committee meeting for the third quarter of 1989 was suspended because of Hurricane Hugo. However, your response failed to address the cause for missing the Radiation Safety Committee meeting during the third quarter of 1990. Nonetheless, your proposed corrective action for this violation is acceptable.

With regard to Violation C, we believe, based on interviews with hospital personnel during the inspection, that the leak tests for June and December 1989 were not performed and that the violation occurred as stated in the Notice of Violation.

We will examine the implementation of your corrective actions during future inspections.

We appreciate your cooperation in this matter.

Sincerely,

W. E. Cline

for

William E. Cline, Chief
Nuclear Materials Safety and
Safeguards Branch
Division of Radiation Safety
and Safeguards

cc: Commonwealth of Puerto Rico

bcc: Document Control Desk

RII:DRSS

RII:DRSS

HB
HBermudez
12/1/90

W. E. Cline
WCline
12/1/90

9101140011 901231
REG2 LIC30
52-24916-01 PDR

1/0
TC 07