

NOV 05 1982

03002027
03011519

Emma L. Rixby Hospital
ATTN: Jay E. Kreuzer
Executive Vice President
818 Riverside Avenue
Adrian, MI 49221

License No. 21-03194-01
License No. 21-03194-04

Gentlemen:

Thank you for your letter dated October 15, 1982 informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated September 20, 1982. In addition to the correction actions described in your letter, we understand, from a telephone discussion between Dr. R. Taylor and Ms. E. Matson that the following actions have been taken or will be taken for the following items of non-compliance described and numbered in our letter:

- Item 1. The Medical Isotope Committee met on September 15, 1982, and will meet December 15, 1982, and quarterly thereafter.
- Item 2. The Housekeeping Department personnel were given the required training on October 14, 1982, and all nurses and new employees will be given instructions before January 1983.
- Item 4. The leak test of the barium-133 source was performed on August 30, 1982. The results were at background as shown by a well counter.
- Item 5. A linearity test was conducted on August 26, 1982. The results were within $\pm 5\%$ accuracy.
- Item 6. Weekly surveys were begun September 9, 1982.
- Item 7. Records of disposals were begun September 9, 1982.
- Item 8. Misadministrations will be reported to the NRC quarterly or immediately upon occurrence.
- Item 10. A teletherapy dry run will be conducted by November 1, 1982.

8211110279 821105
NMS LIC30
21-03194-01 PDR

IE H.Q. File Copy
DMB IE-07

OFFICE	RILL	RILL	RILL			
SURNAME	Matson/cw	Sreniawski	Miller			
DATE	11/2/82					

Emma L. Bixby Hospital


- 2 -

Item 11. The dosimetry system has been calibrated by the NBS the week of October 25, 1982, and a full calibration will be performed using this system by November 30, 1982.

If you do not have the same understanding of the items discussed, please notify this office. We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,



J. R. Miller, Chief
Technical Inspection Branch

cc w/ltr dtd 10/15/82:
DMB/Document Control Desk (RIDS)

OFFICE	R111 <i>sem</i>	R111 <i>RS</i>	R111 <i>RS</i>				
SURNAME	Matson/cw	Sreniawski	Miller <i>fn</i>				
DATE	11/1/82	11/2/82	11/3				