

REPORT DETAILS

Report No. 70-1113/84-09

1. Key Persons Contacted

Licensee Employees

[REDACTED] Manager, Regulatory Compliance
[REDACTED] Manager, Fuel Manufacturing
[REDACTED] Acting Manager, Licensing and Nuclear Materials Management
[REDACTED] Manager, Fuel Chemical Operations
[REDACTED] Manager, Projects and Planning
[REDACTED] Acting Manager, Manufacturing Technology and Engineering
Operations
[REDACTED] Acting Manager, Fuel Fabrication
[REDACTED] Senior Engineer, Measurements and Statistics
[REDACTED] Engineer, Automation Technology
[REDACTED] Manager, Fuel Chemical Quality Assurance
[REDACTED] FCQA, Process Control Engineer
[REDACTED] Senior Engineer, Automation Technology
[REDACTED] Analyst, Licensing and Nuclear Materials Management
[REDACTED] Analyst, Licensing and Nuclear Materials Management
[REDACTED] Specialist, Licensing and Nuclear Materials Management and
SS Representative
[REDACTED] Supervisor, Fuels Support

The inspector also interviewed several other licensee employees.

*Denotes those present at the exit interview.

2. Exit Interview

The inspection scope and findings were summarized on July 20, 1984, with those persons indicated in Paragraph 1 above.

3. Licensee Action on Previous Inspection Findings

In the area of NDA measurements, no previous inspection findings were pending at the time of this inspection.

4. Measurements and Statistical Controls (85206)

a. Incinerator Active Box Monitor (ABM)

(1) System Description and Scope

The General Electric Company, Wilmington, North Carolina, incinerates and measures the SNM content of combustible waste materials generated in various plant processing areas. The combustible waste materials to be incinerated are

segregated from noncombustible waste materials in the facility decontamination room. The segregated combustible and noncombustible waste materials are measured for accountability purposes by the Elephant Gun measurement system. The incinerator facility operates [redacted]

The facility incinerator measurement system is comprised of [redacted] measurement systems that are controlled by a single dedicated computer, called ICAMS (Incinerator Criticality Accountability Measurement System).

The following measurement systems use independent measurements to control the quantity of SNM processed by the incinerator.

At the time of this inspection,

RATHER THAN QUANTITY the ABW measurement system for accountability measurements, the licensee had decided to place operating emphasis upon reducing the backlog of waste boxes to be incinerated.

Currently the facility is generating [] boxes of scrap/waste per week and is incinerating [] boxes per week. With a backlog of [] boxes on hand as of June 9, 1984, it is anticipated that [] will be required to process the backlog of waste boxes. The licensee has no immediate plans to qualify the ABM measurement system until the backlog of waste boxes has been reduced to an acceptable level.

(b) The In Incinerator Monitor (IIM) consists

[] For Criticality Controls, the maximum acceptable control limits indicated in paragraph 4.a(1) above, were established by testing prior to startup of the facility.

(c) The Active Can Monitor (ACM) is comprised of

[] At this time, the ACM is the only Incinerator NDA measurement system qualified for accountability measurements.

(2) ABM Standards and Calibrations

At the time of this inspection, the licensee was using twelve NDA standards that had been fabricated, tested and approved for use in performing NDA accountability measurements with ABM measurement system. The matrix materials used in fabricating these standards contained []

[] that had been withdrawn from the facility stores especially for this project. The SNM used in fabricating these standards was well characterized UO₂ powder that had been analyzed by the facility Chemet and Ledoux laboratories. Traceability of these standards to a National Measurement System was accomplished through analytical techniques employed by the measuring laboratories to measure NBL standard UO₂ material. A listing of all facility NDA standards used to perform measurements by the ACM is presented in Table 1. Several of these standards contain similar quantities of SNM whose densities vary over a wide range.

For the ABM measurement system, calibration curves for passive [U-238] measurements were being produced semiannually using five NDA standards that covered the expected measurement range of the system. The data from these calibration standards were best fitted to a calibration curve using [redacted]. The quality of this calibration curve is tested daily by measuring a single NDA standard. To the extent observed during their inspection, no deviations from approved operating procedures for standards fabrication and calibration measurements were detected.

(3) ABM Measurement Techniques

At the time of this inspection, the ABM measurement system was operating three shifts per day for five days per week as directed by Temporary Operating Instructions (TOI). The ABM measurement system is designed to minimize operator influence on the measurements produced. After entering a minimum number of operating parameters,

[redacted] Since the ABM measurement system had not been qualified for accountability measurements, plus the fact that the facility was in the midst of physical inventory preparations and had no boxes on hand that had been previously measured, the system was not independently tested for measurement quality during this inspection.

Replicate measurements of process waste materials and NDA standards normally used to calculate random and systematic error components were not being performed and are not required since the system has not been approved for accountability measurements.

b. SAM-II Enrichment Verification

(1) System Description and Scope

The licensee routinely employs SAM-II enrichment analyzer measurement systems to verify enrichments of previously made measurements of UO₂ production powder from blend area operations and non-tampersealed items on hand at the time of physical inventories. At the time of this inspection, [redacted] enrichment analyzer systems, employing [redacted] units, were being used to perform enrichment verification measurements.

[redacted] units were being used in the [redacted] unit was being used in the [redacted] area; [redacted] units were being used in the [redacted] area and [redacted] units

REF ID: A64742
were being used in the [REDACTED] ROOM.

[REDACTED] unit, which is qualified for accountability measurements and located in the facility [REDACTED] area, was not examined during this inspection.

Each enrichment measurement verification system is comprised of a SAM-II Unit equipped with 2"x0.5" NaI(Tl) detector. Some units were equipped with

(2) SAM-II Standards and Calibrations

At the time of this inspection, a set of [REDACTED] HDA standards were being used by the [REDACTED] SAM-II units performing verification of previously made enrichment measurements. This set of standards had been fabricated from production UO₂ powder and were contained in 3-gallon buckets. The UO₂ powder contained in these standards had been analyzed by the facility Chemet and Ledoux laboratories and was determined to be traceable to a National Measurement System through analytical techniques employed by the analyzing laboratories. All standards had been tested and approved prior to their use in performing calibrations for HDA measurements. For units operating in the MK-III microprocessor mode, calibrations were performed with three standards fitted to a calibration curve by linear regression techniques. For units operating as SAM-II units, calibrations were performed by point calibration technique employing one HDA standard. All calibrations were being performed at least quarterly or more often, for point calibration units, as changes in enrichments dictated. A listing of approved HDA standards used by the SAM-II enrichment verification units is presented in Table-II. No deviations from approved procedures were noted.

(3) SAM-II Measurement Techniques

Measurement techniques employed by the SAM-II units routinely performing enrichment verification of previously made measurements were not examined during this inspection. Although these SAM-II units were operating [REDACTED] shifts per day for [REDACTED] days per week, the enrichment verification measurements being performed were not being used for accountability purposes. For this reason, the quality of measurement performance was not tested during this inspection.

c. License Conditions Review

Docket 70-1113 License Conditions to Materials and Plant Protection Amendment MPP-3 revised and expanded December 20, 1983 for License No. SNM-1097 were reviewed during this inspection. The licensee was determined to be complying with the six license conditions applicable to SNM measurements by the ABM and SAM-II verification measurement systems.

d. FNMC Plan Review

A review of the facility FNMC plan applicable to NDA measurement by the ABM and SAM-II verification measurement systems was conducted during this inspection. Sections 1.0 Organization, 3.0 Measurements, 4.0 Measurement Control Program, 8.0 Management and Appendix C-3, Specially Accepted Safeguards systems, provide descriptions and methods used to determine the U-235 content of SNM measured by NDA techniques at the facility. To the extent of this review and followup inspection observations, the licensee was determined to be following their approved FNMC plan as required.

e. Standard Operating Procedures (SOP) Review

Several facility operating groups are continuously developing, implementing, altering, and updating SOPs for satisfactory operations, measurement controls, and maintenance functions associated with the ABM and SAM-II verification measurement systems. To the extent reviewed and operational conditions observed, the licensee was judged to be complying with the requirements of the following SOPs.

(1) ABM Measurement System Procedures

- (a) Report No. AT-82-667, Standards, Active Box Monitor-Active Can Monitor, dated February 1983.
- (b) Temporary Operating Instructions (TOI) No. A-837, Extension 1, Incinerator Operation, dated July 8, 1984.
- (c) Quality Notice (QN) No. F-G-1155, Revision 0, Qualification of the Active Box Monitor (ABM), dated December 9, 1982.
- (d) Q.N. No. F-G1073, Revision 0, Qualification Test for the Active Box Monitor and the Active Can Monitor, dated February 1, 1982.

(2) SAM-II Verification Measurement System

- (a) Quality Control Inspection Instructions (QCII) No. 2.5.2.7, Revision 0, SAM-II System and Associated Equipment, dated June 18, 1984.

- (b) Product/Process Quality Plan (PPQP) No. 2.6.1, Revision 13, Enrichment Verification Via the SAM-II System and Associated Equipment, dated June 18, 1984.
- (c) Process Requirement Operating Document (PROD) No. 20.1, U02 Warehouse Powder Receiving Verification Via SAM-II, dated June 22, 1984.
- (d) Quality Control Operating Requirements (QCOR) No. 2.3.1.5, Revision 4, U02 Warehouse Powder Receiving and Enrichment Verification Via SAM-II, dated June 22, 1984.
- (e) QCOR No. 2.1.1.14, Revision 19, Powder Screen and Pack, dated February 28, 1983.
- (f) Quality Information Equipment Plan (QIEP) No. B-20.3, Revision 3, Operation and Maintenance of SAM-II Assay Meter, dated November 9, 1982.
- (g) Q.N. No. 1199, Revision 0, SAM-2, MK-3, Microcontroller 3 Point Calibration Qualification, dated June 24, 1980.

(3) Training Procedure

- (a) Practices and Procedures (P&P) No. 70-32, Revision 9, Qualification of QC Inspection, Examination and Test Personnel, dated February 9, 1983.

f. Operator Training and Qualifications

(1) ABM Measurement System

Since the ABM measurement system has not been qualified for accountability measurements, each facility operating group associated with the system provides operator training guidance. [REDACTED] operators from the Fuels Shop Operations Unit are used to operate the system and are not required to be officially certified as qualified at this time.

(2) SAM-II Verification Measurement System

Training of operators for SAM-2 enrichment verification measurements is based upon the criteria specified by ANSI/ASME 45.2.6-1978, Qualification of Inspection, Examination, and Testing Personnel for Nuclear Power Plants Requirements. Each operator receives special training on the operation of the equipment and must show a thorough understanding of measurements, operations and maintenance procedures associated with enrichment verification measurements. At the time of this inspection, [REDACTED] operators were qualified to perform enrichment verification measurements.

Training tests results and operator certifications were documented.

5. Facility Organization (85202)

The licensee's management organization provides a system of checks and balances for accountability measurements. The facility operations, measurements and accounting groups have different reporting avenues to responsible management. The custody, measurements and accounting functions for SNM control do not report directly through the same supervisory line. The Licensing and Nuclear Materials Management (L&NMM) group coordinates the overall functions for accounting for SNM at the site. At the time of this inspection, the following groups were responsible for measurement and accounting operations associated with the facility incinerator ABM and ACM measurement systems and the SAM-2 enrichment verification measurement systems.

a. ABM and ACM Measurement Systems

b. SAM-2 Enrichment Verification Measurement Systems

(1)

(2)

(3)

PPD U-235
10 CFS 27/20/1982

(4)

6. Physical Inventory Preparations(85212)

On August 23, 1984, the licensee has scheduled their annual physical inventory to account for all SNM materials at the site. In preparation for the physical inventory exercise, the licensee developed a schedule of events to assure that no component essential to the preparations, listing, and accounting for SNM had been overlooked. Inventory preparations were discussed with the licensee and observations of completed scheduled preparation activities was conducted during this inspection. Approximately [REDACTED] physical inventory teams will be used to list the physical inventory. The inspector attended a training session for inventory teams and to the extent observed, the licensee was determined to be following approved inventory instructions written especially for the August 13, 1984 physical inventory.

7. Independent Inspection Effort (92713)

A tour of the manufacturing process areas was conducted during this inspection. Observations and examinations of procedural requirements in the areas of physical inventory preparations, nuclear safety, and criticality items were emphasized during the tour. To extent observed and examined, no deviations from approved procedures were detected.

TABLE 1
ABM MEASUREMENT STANDARDS
(70-113/84-09)

<u>ABM Standard No.</u>	<u>Net Weight (lbs)</u>	<u>Enrichment U-235 (%)</u>	<u>U-235 (Gm's)</u>	<u>U-238 (Gm's)</u>
0(1)	670	-----	---	---
1(2)	700			
2(4)	1185			
3	743			
4(2)	701			
5(2)	700			
6	602			
7	798			
8	695			
9(2)(3)	719			
10	---			
11(4)	---			
12(4)	---			
26(4)	---			

(1) Contains no SNM.

(2) Used as calibration standards.

(3) Standards 9 and 10 are combined and used as standard 910.

(4) These standards are considered developmental standards and are not used for IIM balance measurements.

TABLE 2
SAM-II ENRICHMENT VERIFICATION STANDARDS
(70-113/84-09)

<u>SAM-II Standard No.</u>	<u>Net Weight (Gm's)</u>	<u>Enrichment U-235 (%)</u>	<u>Uranium (Gm's)</u>	<u>U-235 (Gm's)</u>
SAMSTD0101	14671			
0102	13831			
0103	14511			
0104	14411			
0105	15341			
0106	15811			
0107	14581			
0108	14371			
0109	15791			
0116	14730			
0117	14690			
0122	16323			
0123	15013			
0124	5013			
0125	4213			

LHCS TRANSACTION DESCRIPTIONS

TRANSACTION

100 - [REDACTED]

2.790 d

FUNCTION:

MODIFICATIONS:

- 11 Modified the code to eliminate the repetitive input of header information for multiple requests from the same user.

104 - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

105 - [REDACTED]

FUNCTION:

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions _____
FOIA: 87-88

Rutman

Lin 6.1

N-8

LMCS TRANSACTION DESCRIPTIONS

TRANSACTION

MODIFICATIONS:
None.

110 - [REDACTED]

FUNCTIONS:

MODIFICATIONS:

1: Modified T110 to print special OOA/OOC
report to printer and disk file.

111 - [REDACTED]

FUNCTIONS:

MODIFICATIONS:
None.

TRANSACTION

LMDC TRADES... ELECTIONS

112 -

FUNCTION:

MODIFICATIONS:

1:

114 -

FUNCTION:

MODIFICATIONS:

1: Changed nondescriptive list names to the actual
tab name.

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions
FOIA: 87-88

N-8

LHOB TRANSACTION DESCRIPTIONS

TRANSACTION

118

FUNCTION:




MODIFICATIONS:

- 1: Changed nondescriptive list names to the actual list name.
- 2: Added raw data printout routines.
- 3: Added raw data archive routines.

119

FUNCTION:



MODIFICATIONS:

None.

120

FUNCTION:



MODIFICATIONS:

No.

LMOS TRANSACTION DESCRIPTIONS

TRANSACTION

121

FUNCTIONS:

MODIFICATIONS:

None.

127

FUNCTIONS:

MODIFICATIONS:

None.

130

FUNCTIONS:

MODIFICATIONS:

None.

15

15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
66
67
68
69
70
71
72
73
74
75
76
77
78
79
80
81
82
83
84
85
86
87
88
89
90
91
92
93
94
95
96
97
98
99
100
101
102
103
104
105
106
107
108
109
110
111
112
113
114
115
116
117
118
119
120
121
122
123
124
125
126
127
128
129
130
131
132
133
134
135
136
137
138
139
140
141
142
143
144
145
146
147
148
149
150
151
152
153
154
155
156
157
158
159
160
161
162
163
164
165
166
167
168
169
170
171
172
173
174
175
176
177
178
179
180
181
182
183
184
185
186
187
188
189
190
191
192
193
194
195
196
197
198
199
200
201
202
203
204
205
206
207
208
209
210
211
212
213
214
215
216
217
218
219
220
221
222
223
224
225
226
227
228
229
230
231
232
233
234
235
236
237
238
239
240
241
242
243
244
245
246
247
248
249
250
251
252
253
254
255
256
257
258
259
260
261
262
263
264
265
266
267
268
269
270
271
272
273
274
275
276
277
278
279
280
281
282
283
284
285
286
287
288
289
290
291
292
293
294
295
296
297
298
299
300
301
302
303
304
305
306
307
308
309
310
311
312
313
314
315
316
317
318
319
320
321
322
323
324
325
326
327
328
329
330
331
332
333
334
335
336
337
338
339
339
340
341
342
343
344
345
346
347
348
349
349
350
351
352
353
354
355
356
357
358
359
359
360
361
362
363
364
365
366
367
368
369
369
370
371
372
373
374
375
376
377
378
379
379
380
381
382
383
384
385
386
387
388
389
389
390
391
392
393
394
395
396
397
398
399
400
401
402
403
404
405
406
407
408
409
409
410
411
412
413
414
415
416
417
418
419
419
420
421
422
423
424
425
426
427
428
429
429
430
431
432
433
434
435
436
437
438
439
439
440
441
442
443
444
445
446
447
448
449
449
450
451
452
453
454
455
456
457
458
459
459
460
461
462
463
464
465
466
467
468
469
469
470
471
472
473
474
475
476
477
478
479
479
480
481
482
483
484
485
486
487
488
489
489
490
491
492
493
494
495
496
497
498
499
500
501
502
503
504
505
506
507
508
509
509
510
511
512
513
514
515
516
517
518
519
519
520
521
522
523
524
525
526
527
528
529
529
530
531
532
533
534
535
536
537
538
539
539
540
541
542
543
544
545
546
547
548
549
549
550
551
552
553
554
555
556
557
558
559
559
560
561
562
563
564
565
566
567
568
569
569
570
571
572
573
574
575
576
577
578
579
579
580
581
582
583
584
585
586
587
588
589
589
590
591
592
593
594
595
596
597
598
599
600
601
602
603
604
605
606
607
608
609
609
610
611
612
613
614
615
616
617
618
619
619
620
621
622
623
624
625
626
627
628
629
629
630
631
632
633
634
635
636
637
638
639
639
640
641
642
643
644
645
646
647
648
649
649
650
651
652
653
654
655
656
657
658
659
659
660
661
662
663
664
665
666
667
668
669
669
670
671
672
673
674
675
676
677
678
679
679
680
681
682
683
684
685
686
687
688
689
689
690
691
692
693
694
695
696
697
698
699
700
701
702
703
704
705
706
707
708
709
709
710
711
712
713
714
715
716
717
718
719
719
720
721
722
723
724
725
726
727
728
729
729
730
731
732
733
734
735
736
737
738
739
739
740
741
742
743
744
745
746
747
748
749
749
750
751
752
753
754
755
756
757
758
759
759
760
761
762
763
764
765
766
767
768
769
769
770
771
772
773
774
775
776
777
778
779
779
780
781
782
783
784
785
786
787
788
789
789
790
791
792
793
794
795
796
797
798
799
800
801
802
803
804
805
806
807
808
809
809
810
811
812
813
814
815
816
817
818
819
819
820
821
822
823
824
825
826
827
828
829
829
830
831
832
833
834
835
836
837
838
839
839
840
841
842
843
844
845
846
847
848
849
849
850
851
852
853
854
855
856
857
858
859
859
860
861
862
863
864
865
866
867
868
869
869
870
871
872
873
874
875
876
877
878
879
879
880
881
882
883
884
885
886
887
888
889
889
890
891
892
893
894
895
896
897
898
899
900
901
902
903
904
905
906
907
908
909
909
910
911
912
913
914
915
916
917
918
919
919
920
921
922
923
924
925
926
927
928
929
929
930
931
932
933
934
935
936
937
938
939
939
940
941
942
943
944
945
946
947
948
949
949
950
951
952
953
954
955
956
957
958
959
959
960
961
962
963
964
965
966
967
968
969
969
970
971
972
973
974
975
976
977
978
979
979
980
981
982
983
984
985
986
987
988
989
989
990
991
992
993
994
995
996
997
998
999
1000

FUNCTION:
MODIFICATIONS:
None.

FUNCTION:
MODIFICATIONS:
None.

FUNCTION:
MODIFICATIONS:
None.

8

1100 TRANSACTION DESCRIPTION

TRANSACTION
DESCRIPTION

210

[REDACTED]
FUNCTION:

[REDACTED]
MODIFICATIONS:

- 1: CHANGED nondescriptive list names to the actual job name.

210

[REDACTED]
FUNCTION:

[REDACTED]
MODIFICATIONS:

None.

240

[REDACTED]
FUNCTION:

[REDACTED]
MODIFICATIONS:

1:

LHCS TRANSACTION DESCRIPTION

TRANSACTION

210

FUNCTIONS:

MODIFICATIONS:

None.

270

FUNCTIONS:

MODIFICATIONS:

None.

330

FUNCTIONS:

MODIFICATIONS:

None.

340

FUNCTIONS:

MODIFICATIONS:

None.

LHCS TRANSACTION DESCRIPTIONS

TRANSACTION

360

FUNCTION:

MODIFICATIONS:

None.

361

FUNCTION:

MODIFICATIONS:

None.

363 - VERIFIES INITIALIZATION OF GEOMETRIC DFNSITY SAMPLES

FUNCTION:

MODIFICATIONS:

None.

380

FUNCTION:

MODIFICATIONS:

None.

LXCS TRANSACTION DESCRIPTIONS

TRANSACTION

670 -

FUNCTION:

MODIFICATIONS:

None.

672 -

FUNCTION:

MODIFICATIONS:

None.

669 -

FUNCTION:

MODIFICATIONS:

None.

701A -

FUNCTION:

MODIFICATIONS:

Non

1. LINE TRANSACTION DESCRIPTIONS

TRANSACTION

700

FUNCTION:

MODIFICATIONS:

None.

705

FUNCTION:

MODIFICATIONS:

None.

705A

FUNCTION:

MODIFICATIONS:

None.

706

FUNCTION:

MODIFICATIONS:

None.

LHCS TRANSACTION DESCRIPTIONS

TRANSACTION

707

FUNCTION:

MODIFICATIONS:

None.

710

FUNCTION:

MODIFICATIONS:

None.

710A

FUNCTION:

MODIFICATIONS:

None.

714

FUNCTION:

MODIFICATIONS:

None.

LNCS TRANSACTION DESCRIPTIONS

TRANSACTION

700

FUNCTION:

MODIFICATIONS:
None.

700

FUNCTION:

MODIFICATIONS:
None.

730

FUNCTION:

MODIFICATIONS:
None.

730A

FUNCTION:

MODIFICATIONS:
None.

D LM03 TRANSACTION DESCRIPTIONS

TRANSACTION

771

FUNCTION:

MODIFICATIONS:
None.

740

FUNCTION:

MODIFICATIONS:
None.

740A

FUNCTION:

MODIFICATIONS:
None.

750

FUNCTION:

MODIFICATIONS:
None.

LHCB TRANSACTION DESCRIPTIONS

TRANSACTION

760 - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

760 - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

760A - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

763 - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

LAC'S TRANSACTION DESCRIPTION

TRANSACTION

712A

FUNCTION:

MODIFICATIONS:

None.

769

FUNCTION:

MODIFICATIONS:

None.

759A

FUNCTION:

MODIFICATIONS:

None.

775

FUNCTION:

MODIFICATIONS:

None.

D

LADS TRANSACTION DESCRIPTIONS

TRANSACTION

775A - [REDACTED]

FUNCTION:

[REDACTED]

MODIFICATIONS:

None.

776 [REDACTED]

FUNCTION:

[REDACTED]

MODIFICATIONS:

None.

776A - [REDACTED]

FUNCTION:

[REDACTED]

MODIFICATIONS:

None.

777 - [REDACTED] 7

FUNCTION:

[REDACTED]

MODIFICATIONS:

None.

LHCO TRA-B SECTION DESCRIPTION

TRANSACTION
DESCRIPTION

780

FUNCTION:

MODIFICATIONS:

None.

7804

FUNCTION:

MODIFICATIONS:

None.

790

FUNCTION:

MODIFICATIONS:

None.

799

FUNCTION:

MODIFICATIONS:

None.

LHCS TRANSACTION DESCRIPTIONS

TRANSACTION

820 - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

821 - [REDACTED]

FUNCTION:

T621 is no longer in use.

MODIFICATIONS:

None.

830 - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

840 - [REDACTED]

FUNCTION:

MODIFICATIONS:

None.

D LMDS TRANSACTION DESCRIPTIONS

TRANSACTION

2400

FUNCTION:

MODIFICATIONS:

None.

842

FUNCTION:

MODIFICATIONS:

None.

847

FUNCTION:

MODIFICATIONS:

None.

860

FUNCTION:

MODIFICATIONS:

None.

D LMCS TRANSACTION DESCRIPTIONS

TRANSACTION
DESCRIPTIONS

810 -

FUNCTION:

MODIFICATIONS:

None.

870 -

FUNCTION:

MODIFICATIONS:

1: Per lab instructions, this transaction is no longer used.

871 -

FUNCTION:

MODIFICATIONS:

None.

872 -

FUNCTION:

MODIFICATIONS:

None.

LHCG TRANSACTION DESCRIPTIONS

TRANSACTION
DESCRIPTIONS

170A - [REDACTED]

FUNCTION:

MODIFICATIONS:
None.

272 - [REDACTED]

FUNCTION:

MODIFICATIONS:
None.

874 - [REDACTED]

FUNCTION:

MODIFICATIONS:
None.

881 - [REDACTED]

FUNCTION:

MODIFICATIONS:
None.

3
1403 TRANSACTION DESCRIPTIONS

TRANSACTION

874 -

FUNCTION:

MODIFICATIONS:

None.

875 -

FUNCTION:

This transaction does not work.

MODIFICATIONS:

None.

902 -

FUNCTION:

MODIFICATIONS:

None.

903 -

FUNCTION:

MODIFICATIONS:

None.

D

LNCS TRANSACTION DESCRIPTIONS

TRANSACTION

941

FUNCTION:

MODIFICATIONS:

None.

945

FUNCTION:

MODIFICATIONS:

None.

946

FUNCTION:

MODIFICATIONS:

None.

950

FUNCTION:

MODIFICATIONS:

None.

LNCC TRANSACTION DESCRIPTIONS

TRANSACTION

952 -

FUNCTION:

MODIFICATIONS:

None.

953 -

FUNCTION:

MODIFICATIONS:

None.

953 -

FUNCTION:

MODIFICATIONS:

None.

970 -

FUNCTION:

MODIFICATIONS:

None.

LMCS TRANSACTION DESCRIPTIONS

TRANSACTION

970 - MESSAGE LIST REPORT

FUNCTION:

MODIFICATIONS:

None.

972A - MESSAGE LIST DUMP

FUNCTION:

MODIFICATIONS:

None.

975 - UPDATE LIST RECORDS

FUNCTION:

MODIFICATIONS:

None.

976 -

FUNCTION:

MODIFICATIONS:


None.

B
• 4003 TRANSACTION DESCRIPTIONS

TRANSACTION

927 -

FUNCTION:




MODIFICATIONS:

None.

928 -

FUNCTION:




MODIFICATIONS:

None.

929 -

FUNCTION:



MODIFICATIONS:

None.

992 -

FUNCTION:

This transaction is not used.

MODIFICATIONS:

None.

LHOC TRANSACTION DESCRIPTIONS

TRANSACTION

990

FUNCTION:

MODIFICATIONS:

None.

999 - REQUEST ASSISTANCE

FUNCTION:

T999 is no longer in use.

MODIFICATIONS:

None.

INMET LIBRARY & SAMPLE REPORT

SAMPLE TO: K-10-C-1
EXX-EXX-EXX-EXX-EXX-EXX-EXXSAMPLE # 1 260480 NM, OF THIS 1 2
SERIES 2SAMPLES FOR # 260480
REQ. COMPONENT : ANT
DOCUMENT NO. : FEGO 80-44
CONTAINER/LOT #: 1216E008APAGE: 1
DATE: 24-AUG-82
TIME: 22:22NAME SAMPLE: 1 260480
TIME SAMPLE: 1 08100
PEN: 1 FFU
MATERIAL CLASS: 1 H-2
SOURCE: 1 2+708

ANALYST METHOD:

CODE EX NAME

40 00 FEGO 1EOTOFITB
205 00 O/H IN URESULT 10-15 TECH FORMERED
1.142 2.6-2.8 [REDACTED] 2-1
1.142 0.6-1.1 [REDACTED] 2-1
[REDACTED]

Please Add this figure -

O/L Supper To be 2.053

P7ch = 0.42
@ 2280() this is $\frac{1}{2} \times 0$
And just throw this
out4/24/82 Sun - 9:21:3
150-2330

OK

They are Super. Post w

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions 6 + 7C
FOIA: 87-88

FOIA-87-88 - GRC

O-1

CALIBRATION/VERIFICATION LOG
ENRICHMENT ANALYZERS

TIME
DATE 8-19-82
ANALYZER # 4

TIME	VERIFICATION/CALIBRATION												M	K	
	BLANK		.715 VERIF./CAL.		1.454		2.223 VERIF./CAL.		2.999		3.978 VERIF./CAL.				
	U	B	Enr.*	U	B	U	B	Enr.*	U	B	U	E	Enr.*	U	B
2300															
24															
0145															
0221	—	—	.715												
0337	—	—													
0412	—	—	.714												
0515															
060550	—	—	.712												
07															
0832	17.10		70%												
09															
10	—		warn w/o alarm mode Time 3 min												
11															
12															
13															
14															
15															
16															
17															
18															
19															
20															
21															
22															

* CIRCLE OUT OF ALARM VALUES

CL-338 R1 Ref: SCP-401 Issued: 5/7/82

Approved:

TIME	VOLTS	COARSE	FINE
2300			

Information in this record was deleted
in accordance with the Freedom of Information
Act, Exemptions
FOIA, 82-88

O. 2

CALIBRATION/VERIFICATION LOG

ENRICHMENT ANALYZERS

Friday

DATE 8-20-82ANALYZER # 4

TIME	VERIFICATION/CALIBRATION												H	K		
	BLANK		.715 VERIF./CAL.		1.454		2.223 VERIF./CAL.		2.999		3.978 VERIF./CAL.					
	U	B	Enr.*	U	B	U	B	Enr.*	U	B	U	B				
2300																
24																
01																
02																
03																
04																
05																
06																
07																
08																
09																
10	Gashen	@ 10 min	T: Run 6	Hi < 100	for My Count.	by	[REDACTED]									
1121																
1252																
13																
1406	61:															
1523																
1557																
17																
18																
19																
20																
2145																
2247																

* CIRCLE OUT OF ALARM VALUES

AMP SETTING CHANGES:

TIME	VOLTS	COARSE	FINE
2300			

CALIBRATION/VERIFICATION LOG

ENRICHMENT ANALYZERS

DATE 8-21-82
ANALYZER # 4

VERIFICATION/CALIBRATION									
TIME	BLANK	.715 VERIF./CAL.	1.454	2.223 VERIF./CAL.	2.999	3.978 VERIF./CAL.			
	U	B	Enr.*	U	B	Enr.*	U	B	Enr.*
~230422									
2404									
01									
*0243									
03									
04									
0522									
06									
07									
08									
09									
10									
11									
12									
13									
14									
15									
16									
17									
18									
19									
20									
21									
22									

* CIRCLE OUT OF ALARM VALUES

CL-320 RI Ref: SCP-401 Issued: 5/7/82 Approved [REDACTED]

AMP SETTING CHANGES:

TIME	VOLTS	COARSE	FINE
2304			

CALIBRATION/VERIFICATION LOG

ENRICHMENT ANALYZERS

500

DATE 8-22-82ANALYZER # 4

TIME	VERIFICATION/CALIBRATION												M	K	
	BLANK		.715 VERIF./CAL.		1.454		2.223 VERIF./CAL.		2.999		3.978 VERIF./CAL.				
	U	B	Enr.*	U	B	U	B	Enr.*	U	B	U	B	Enr.*	U	B
2300															
24															
01															
02															
03															
04															
05															
06															
07															
08															
09															
10															
11															
12															
13															
14															
15															
16															
17															
18															
19															
20															
21															
22															

* CIRCLE OUT OF ALARM VALUES

CL-328 RI

Ref: SCP-401

Issued: 5/7/82

Approved:

AMP SETTING CHANGES:

TIME	VOLTS	COARSE	FINE
2300			



8/24/82 (C) 10:20
T8e wood second Tag.

The attempt to legally
new wood went to

put in qualification of
Al. Alativity -

very careful reading was

done in

at this time
very cold winds

Careful line of big game!

What will happen to

the one b

The species called in d
before about Al. Nelson
of Sample (which first
was not qualified). Will
be documented on disc
marked? Why did
not do it




TABLE FOR SCP-401

ANALYZER #	(NP-BU Series)	
	REGRESSION MODEL COEFFICIENT	
	SLOPE	INTERCEPT
011	.999535	-.015054
012	1.000361	-.011828
013	1.003814	-.014472
014	—	Down —

ANALYZER #	(NP-BU Series)		
	SLOPE	INTERCEPT	MINIMUM "U" COUNT
011	1.00593	-.00412	279269
012	1.00906	-.00522	289827
013	1.00544	-.00599	290003
014	—	Down —	287181 8/18/81

NOTE: Check each calibration. The "B" counts for the lowest enrichment must have a "B" count greater than the "B" count for the highest enrichment.

APPROVED BY

CALIBRATION/VERIFICATION LOG

ENRICHMENT ANALYZERS

M

DATE 3-23-82ANALYZER #

TIME	VERIFICATION/CALIBRATION												M	K	
	BLANK		.715 VERIF./CAL.		1.454		2.223 VERIF./CAL.		2.999		3.978 VERIF./CAL.				
	U	B	Enr.*	U	B	U	B	Enr.*	U	B	U	B	Enr.*	U	B
2300															
24															
01															
02															
03															
04															
05															
06															
07															
08															
09															
10															
11			C												
12															
13															
14															
15															
16															
17															
18															
19															
20															
21	75			70				2216				5255			
22															

* CIRCLE OUT OF ALARM VALUES

CL-338 R1 Ref: SCP-401 Issued: 5/7/82 Approved: [Redacted]

AMP SETTING CHANGES:

TIME	VOLTS	COARSE	FINE
2300			

568

PRODUCTION g/g ISOTOPIC Δ O/U RATIO

17
15
AY 1

8U/8 ISOTOPIC
6 O/U RATIO

100% Recovery on 8U and 100% recovery
on 6U

CONTAINER #	CRUC POS.	TARE WGT.	UO ₂ WGT.	TARE + U ₃ O ₈	TARE + SOLUTION	TUBE #	ANA #	MEAS. U ₂₃₅	CALC. U ₂₃₅	O/U RATIO
3300B1B0033	1	27.0874	4.4872	31.6196	53.9708	7604	X-1	4.182	3.258	2.731
3000B1B0250	2	26.7727	4.4871	33.4117	55.7552	7946	X-1	3.892	3.002	2.096
3000B1P0067	3	26.6157	4.5076	33.2107	51.6256	1770	4.015	3.004	2.062	2.047
3000B1D0251	4	29.4245	4.4854	34.0033	50.0000	7		2.056	2.056	
3300B1M00292	5	29.3505	4.5037	34.0255	52.4852	7672	4120	3.192	3.027	
3300B1M00292	6	26.9062	4.4925	31.0162	53.1635	7746	4.078	3.101	2.056	
3300B1D0084	7	27.1420	4.5054	36.9127	54.3840	7326	4.199	3.235	2.048	
3300B1M00296	8	29.2105	4.4515	33.2532	55.2552	7376	4.380	3.222	2.031	
3000B1C60935	9	25.5568	4.4524	30.3119	49.3251	7033	4.491	3.614	2.037	
3000B1P009663	10	30.5070	4.5735	34.0002	56.3522	7225	4.078	3.973	2.067	
3000B1P009684	11	29.6620	4.5002	37.3204	50.0002	7		2.0		2.046
39500B1D0372	12	26.4577	4.5005	31.0129	51.0625	7328	5.647	3.939	2.093	
								714	2.724	3.911
8-20-82		0% IN <u>6/20</u>	OX. OUT <u>5/20</u>		9/10 <u>5/25</u>		-2.21 <u>5.230</u>		5-21 <u>11/6</u>	<u>5-21-02</u>

CL-245 Rev.1

Ref: SCP 400 Issued: 5/4/82

Approved:

CALIBRATION/VERIFICATION LOG
ENRICHMENT ANALYZERS

DATE 8-25-82
ANALYZER # 4

TIME	VERIFICATION/CALIBRATION												M	N	
	BLANK		.715 VERIF./CAL.		1.454		2.223 VERIF./CAL.		2.999		3.978 VERIF./CAL.				
	U	B	Enr.*	U	B	U	B	Enr.*	U	B	U	B	Enr.*	U	B
/2304	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/2402	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/0058	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/0210	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/0246	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/0454	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
05	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
06	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
J 07 : 0	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
08	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
J 09 : 0	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
J 10 : 0	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
J 11 : 0	61	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
12	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
13	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
14	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
15	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
16	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/173,	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/1852	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/1928	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
20	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/2148	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500
/2251	—	—	—	—	—	—	—	—	—	—	—	—	Pos. 51000	—700	—500

* CIRCLE OUT OF ALARM VALUES

SL-338 RI Ref: SCP-401 Issued: 5/7/82 Approved [Redacted]

AHP SETTING CHANGES:

TIME	VOLTS	COARSE	FINE
2300			

CALIBRATION/VERIFICATION LOG
ENRICHMENT ANALYZERS

T-444
DATE 8-24-82
ANALYZER # 4

TIME	VERIFICATION/CALIBRATION												M	K	
	BLANK		715 VERIF./CAL.		1.454		2.223 VERIF./CAL.		2.999		3.978 VERIF./CAL.				
	U	B	Enr.*	U	B	U	B	Enr.*	U	B	U	B	Enr.*	U	B
✓ 2308 25	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
✓ 2401	-	-	-	7.24	-	-	-	-	-	-	-	-	-	-	7425 1.204
01	-	-	-	-	-	-	-	2.223	-	-	-	-	-	-	-
✓ 0208	10/0	-	-	7.10	-	-	-	2.211	-	-	-	-	-	-	-
03	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
✓ 0400	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
✓ 050436	-	-	-	7.16	-	-	-	2.222	-	-	-	-	-	-	70617 1.209
06	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
✓ 0706	10/0	-	-	7.21	-	-	-	(2.176) ^{out}	-	-	-	-	-	-	-
✓ 0826	-	-	-	-	-	-	-	-	-	-	-	-	2.993 ^{out}	Required	-
✓ 0903	-	-	-	7.07	-	-	-	2.211	-	-	-	-	-	-	70649 1.209
10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
11	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
✓ 1205	16/0	-	-	7.13	-	-	-	2.225 ^{out}	-	-	-	-	-	-	-
✓ 1350	-	-	-	-	-	-	-	-	-	-	-	-	3.959	Permitted	-
✓ 1426	-	-	-	7.11	-	-	-	2.222	-	-	-	-	-	-	70617 1.209
✓ 1537	-	-	-	-	-	-	-	-	-	-	-	-	3.959	-	-
✓ 1611	-	-	-	7.01	-	-	-	2.222	-	-	-	-	-	-	70649 1.209
17	-	-	-	-	-	-	-	-	-	-	-	-	3.959	-	-
18	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
✓ 1947	%	-	-	7.16	-	-	-	2.215	-	-	-	-	-	-	-
20	-	-	-	-	-	-	-	-	-	-	-	-	5.222	-	-
21	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-
22	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-

* CIRCLE OUT OF ALARM VALUES

SL-338 R1 Ref: SCP-401 Issued: 5/7/82 Approved: [Redacted]

AMP SETTING CHANGES:

TIME	VOLTS	COARSE	FINE
2300	-	-	-
-	-	-	-
-	-	-	-
-	-	-	-

CHEMET LABORATORIES SAMPLE REPORT

 * MAIL TO: KB7 CR

SAMPLE #: 303083 NO. OF TESTS : 2

PAGE: 1
 DATE: 17-AUG-82
 TIME: 22:135

SAMPLER'S FAY #: 23875
 REQ. COMPONENT : 987
 DOCUMENT NO. : FROD 10.50
 CONTAINER/LOT #: 3305RM0554

DATE SAMPLED : 16-AUG-82
 TIME SAMPLED : 04:41
 AREA : FFU
 MATERIAL CLASS : 1 U02
 SOURCE : SLUGGER

ANLS METHOD

CODE EX NAME

RESULT

UNITS

TECH

COMPLETI

CODE	EX	NAME	RESULT	UNITS	TECH	COMPLETI
60	00	FROD ISOTOPICS	3.443	X U-235	[REDACTED]	17-AUG-82 22
205	00	O/U in U	3.443	O/U RATIO	[REDACTED]	17-AUG-82 22

Should be

2.046

[] bind thru []

Miss Point
 of LNICS - from this page
 in not the permanent type

Information in this record was deleted
 in accordance with the Freedom of Information
 Act, exemptions
 FOIA

67-644

0.3

6
[] did say [] gave results to []
] to him. Actually, what was wrong?
Were the results [] gathered for []
Where did [] come up with his six standa

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions 6
FOIA: 87-88

O. 4

TABLE FOR SCT-401

ANALYZER #	(HP-BU Series)	
	SLOPE	INTERCEPT
011	.999535	-.015054
012	1.000361	-.011828
013	1.003814	-.014472
014	—	Down —

ANALYZER #	(HP-BU Series)		
	SLOPE	INTERCEPT	MINIMUM "U" COUNT
011	1.00593	-.00422	279269
012	1.00906	-.00522	289827
013	1.00544	-.00589	290003
014	—	Down	287181 ^{84%}

NOTE: Check each calibration. The "B" counts for the lowest enrichment must have a "B" count greater than the "B" count for the highest enrichment.

ANALYZER 3

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions 6 & 7C
FOIA 87-88

O. 5

• OUT-OF-LIMIT values

EQUIPMENT ANALYZER CALIBRATION/VERIFICATION LOG

DATE: 6-25-62
ANALYZER: 4

SAMPLES CTD/RUSD		VERIFICATION *			CALIBRATION		
PROD.	G218	.715 enr	2.223	3.978	REPK'D	3.978 counts	H K
/	/						POS. DM,SL
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						
/	/						

* OUT-OF-LIMIT values.

ANALYZER: 3

SAMPLES CTD/R LSD		VERIFICATION *			CALIBRATION			POS.BK.SL	
PROD.	G218	.715 enr	2.223	3.978	REP K'D	3.976 counts	H	K	
/	/								
/	/								
/	/								
/	/								
/	/								
/	/								
/	/								
/	212	.712	2.220	3.975	✓	3.975	6.74	1.57	
/	/								
/	212	.710	2.218	3.973	✓	3.973	6.72	1.55	
/	/								
/	212	.712	2.220	3.975					
/	/								
/	212	.710	2.218	3.973					
/	/								
/	212	.712	2.220	3.975					
/	/								
/	210	.711	2.216	3.974	✓	3.974	6.73	1.56	
/	/								
/	210	.710	2.215	3.973	✓	3.973	6.72	1.55	
/	/								
/	210	.711	2.216	3.974					
/	/								
/	210	.715	2.220	3.976					
/	/								
/	/								

* OUT-OF-LIMIT values.

ENVIRONMENTAL ANALYZER CALIBRATION VERIFICATION LUB

DATE: 10-22
ANALYZER: 3

SAMPLES CTD/RISD				VERIFICATION *				CALIBRATION			
PROD.	G218	.715 entr	2.223	3.978	REPK'D	3,978 counts	N	K	PCT. B.E. SL		
1	z / o	705	z 207	z 517	-	-	-	-	-		
+	74				-	-	-	-	-		
-1-	-2				-	-	-	-	-		
1	1	721	z 227	3 554							
1	1										
1	2 / 2	709	z 223	3 993							
1	1										
1	1										
1	1										
1	1										
1	1	720	z 237	3 991							
1	1	711	z 221	3 998							
1	1										
1	1										
1	1	710	z 212	3 976	99.75	3,978	100.00	100.00	100.00		
1	1	711	z 200	3 977	99.75	3,978	100.00	100.00	100.00		
1	1										
1	1										
1	212	713	z 371	3 572	99.75	3,573	100.00	100.00	100.00		
1	111	711	z 377	3 574	99.75	3,575	100.00	100.00	100.00		
1	1										

• OUT-OF-LIMIT values.

ANALYZER: 3

SAMPLES CTD/RLSD		VERIFICATION *			CALIBRATION				
PROD.	G218	.715' entr	2.223	3.978	REPK'D	3.978 counts	M	K	POS. MK.SL
/	/								
/	1.15	7.12	(2 : 7.6)	3.556					
/	/	7.22	2.2.5	3.968					
/	/								
/	/								
/	/								
/	0	1	1.735	2.277	3.972				
/	/		→						
/	1	1	1.711	2.217	3.977				
717	1	1.711	2.213	3.984					
/	/								
/	/								
/	/								
/	/								
/	/								
/	/								
/	/								
/	/								

* OUT-OF-LIMIT values.

TABLE PAGE 109-001

6

STANDARD VERIFICATION LIMITS (9825A)			
ANALYZER #011	ANALYZER #012	ANALYZER #013	ANALYZER #014
.719 ± .009	.717 ± .012	.714 ± .009	.715 ± .008
2.223 ± .013	2.231 ± .015	2.220 ± .012	2.224 ± .014
3.982 ± .019	3.982 ± .024	3.976 ± .021	3.966 ± .021

REGRESSION MODEL COEFFICIENT (IP-80)		
ANALYZER	SLOPE	INTERCEPT
011	1.003968	- .006642
012	1.005995	- .009013
013	1.003739	- .002528
014	1.005886	- .006204

DTAS ADJUSTMENT FACTOR AND MINIMUM "U" COUNT LIMITS (9825A)			
ANALYZER	SLOPE	INTERCEPT	MINIMUM "U" COUNT
011	1.00593	- .00422	285048
012	1.009012	- .00522	289827
013	1.005714	- .00599	290003 283060
014	1.00569	- .01038	287744

6-2

NOTE: Prior to calibration, the "P" counts for the lowest enrichment level were a "P" count greater than the "R" count for the highest enrichment.

TABLE FOR SCP-501

STANDARD VERIFICATION LIMITS (9825A)			
ANALYZER #011	ANALYZER #012	ANALYZER #013	ANALYZER #014
.719 ± .009	.717 ± .012	.714 ± .009	.715 ± .008
2.223 ± .013	2.231 ± .015	2.220 ± .012	2.224 ± .014
3.982 ± .019	3.982 ± .024	3.976 ± .021	3.966 ± .021

REGRESSION MODEL COEFFICIENT (HP-80)		
ANALYZER	SLOPE	INTERCEPT
011	1.003968	-0.006642
012	1.005995	-0.009013
013	1.003739	-0.002528
014	1.005886	-0.006204

BIAS ADJUSTMENT FACTOR AND MINIMUM "U" COUNT LIMITS (9825A)			
ANALYZER	SLOPE	INTERCEPT	MINIMUM "U" COUNT
011	1.00593	-0.00422	285048
012	1.00906	-0.00522	289827
013	1.005744	-0.00599	290003 283060
014	1.00569	-0.01038	287744

NOTE: Check each calibration. The "B" counts for the lowest enrichment must have a "B" count greater than the "B" count for the highest enrichment.

6-28

ATTN: I AM RUFFY@RDF
LMCR - STANDARD OUT OF LIMITS NOTICE: REFERENCE # 9
INPUT VERIFICATION REPORT TR
OUT-OF-ALARM

DATE: 12-FEB-81
TIME: 0615R114

TECH PAY #: [REDACTED] TECH NAME: [REDACTED] T#111
METHOD CODE: A000 METHOD NAME: OPTOD ISOTOPIC
ANALYZER #: 12 ANALYZER NAME: ATIV ENRICHMFNT #2
STD TYPE #: 901 STANDARD BATCH NAME: .715 ISO. VERIF.
INIT CONC: 60 LINT NAME: % H-235

UPPER CTL: 0.73100
UPPER ALRM: 0.72400
NOMINAL: 0.71500
LOWER ALRM: 0.70600
LOWER CTL: 0.69900
>OF SIGMA: 0.00480

VERIF STD #: 200011 RESULT: 0.72500

Was value input PROPERLY? YES ... NO ...
(If answer is NO, attach Transaction R20 REPORT.)

What do control chart and trend data show?

Have there been any changes or UNUSUAL EVENTS?

YES ... NO ...

If YES, what?

Are there any distinctions with this analysis? (Do other standards show a
analyze different? Preparation, Tech?, Calibration?) YES ... NO ...
If YES, what?

Did investigation show a most likely cause? YES ... NO ...

If YES, what?

What corrective action was taken?

When? ... By whom? ...

Chemist approval: Date:

Released out on t790 DATE:

Weekend of 1/15-16/83

I was working day shift when I found the following 902 transactions:

1902 - 15/1/83 at 0840 #13675
1902 - " " 0841 #13677
1902 " " 1257 "
(Can't delete)
1902 " " 1259 "
(Can't delete)
1902 " " "
1902 " " 1301 "
1902 " " 1414 "

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions

FOIA: 6 7. 84 6 7. 84

ICC(2)

4/22/83 @ 8 PM To 8 AM
in front of GAD INOpriate

intended about 6 6
putting a wrong STD
in lines that shut down
the system! Very few
Pass Word - Come in the
next day & very body
wondered why she didn't
& did not catch it! She
told me she didn't know
what lines put in - if
it was her! Found
out that the department
in effect the other that
told her - she didn't have
time - said no time to
go home! He put it in!
etc etc

LEAD SUPERVISOR

LHCS - STANDARD OUT OF LIMITS NOTICE

DATE: 18-DEC-82
TIME: 12116124
71110

+ CLK +
Fakat
b1

Y #: [REDACTED]
CODE #: 6070
R #: 12
L #: 903 STANDARD BATCH NAME: 3.978 ISO. VERIF.

TECH NAME: [REDACTED]
METHOD NAME: FROD ISOTOPICS

ANALYZER NAME: ADV ENRICHMENT #2

TL : 4.02200
LRM : 4.00600
LRM : 3.97800
LRM : 3.96000
TL : 3.94400

STATION STD #: 200008

HECK FAILURE. RESULT = 2.22300

Not in line
Shift at trip
line.

long come in
Co 1500
12/8/82

6700 - 1580 Shift
12/8/82.

F-2 PERSONNEL ACCOUNTING USE ONLY		F-3 PERSONNEL ACCOUNTING USE	
SDO	SDO	SDO	SDO
SDO	SDO	SDO	SDO
FR 6135667	FR 6135667	FR 6135667	FR 6135667
FOR SUPERVISOR USE ONLY	FOR SUPERVISOR USE ONLY	FOR SUPERVISOR USE ONLY	FOR SUPERVISOR USE ONLY

GENERAL ELECTRIC COMPANY
ATTENDANCE RECORD

51

GENERAL ELECTRIC COMPANY
ATTENDANCE RECORD

51

CHEMET LABORATORIES SAMPLE REPORT

RECEIVED
BY MAIL TO I KH-C/R

AMPLE #1 [REDACTED] NO. OF TESTS : 2

PAGE: 1
DATE: 24-APR-82
TIME: 22:22

SAMPLER'S PAY #: 22496
REQ. COMPONENT #: 987
DOCUMENT NO. : PROD 80.96
CONTAINER/LOT #: 1200E500R6

DATE SAMPLED : 22-APR-82
TIME SAMPLED : 08:00
AREA : PFU
MATERIAL CLASS : 1 H02
SOURCE : 26708

ANLS METHOD

CODE	EX NAME	RESULT	UNITS	TECH	COMPLETED
60 00	FROD ISCTOPICS	1.191	% U-235	[REDACTED]	24-APR-82 22:11
205 00	O/U in U	1.191	O/U RATIO	[REDACTED]	24-APR-82 22:21

Please add this figure -
O/U Suppose To be 2.053 ✓

4/24/82
@ 2230

This is fixed
throw this
AND dist:
out
Deleted
They are Super. pos. w/
4/24/82 sun - 902/903
150-2330
OK

D 4
PAGE: 1
MAIL TO: ER2
DATE: 30-DEC-83
TIME: 21:21

SAMPLE #: 540914 NO. OF PTS : 2

SUPERSEDES ALL PREVIOUS REPORTS

SAMPLER'S ID: [REDACTED]
REQ. COMPONENT: 982
DOCUMENT NO.: 1 PROD 90.43
CONTAINER/LOT #: 220RHF0350

DATE SAMPLED : 29-DEC-83
TIME SAMPLER : 01:40
AREA : SEC
MATERIAL CLASS : 1 U02
SOURCE : SLUGGER

ANALYSIS

CODE ANZ	NAME	RESULT	UNITS	UACH	COMPLETE
60 11 PROD ISOTOPICS		2.148	% U-235	[REDACTED]	30-DEC-83
205 701 O/U in H	6	2.076	O/U ratio	[REDACTED]	30-DEC-83

② Sig'd a LHS
by Owner print-out
for Sample 90.43 Trace #360
* 540914 Sampled Trace #360
that another chemistry was
added by mistake. (rc)

540914

PLATE PAY NO.

COMPONENT 98.7NO. 90163AINER/LOT NO. 220.RHP 0350

PLTS TO

CODE K8.7

NAME

KS/SPECIAL INSTR

8 DATE SAMPLED

12 2983
MO DAY YR
01 40

7 TIME SAMPLED

8 AREA GECO

9 SOURCE

STORY ITEM

MATERIAL CLASSIFICATION

FUELS	GAD SHOP W/O	ENVIRONMENTAL	ALL METALS
UO ₂	15 Gd UO ₂	25 RIVER SAMPLES	39 B/C
SP	16 Gd SP	26 LIQUID WASTES	40 ZIRC 2
COORDINATED GECO	17 Gd-LW	27 SOIL SAMPLES	41 ZIRC 4
UO ₂	18 GdO ₂	28 STACK FILTERS	42 CS
RAD LIQUIDS/WASTE	20 Gd Smear	29 BIOASSAY	
SCRAP	21 Gd Rod	31 NH ₄ OH	
FUEL ROD	22 Gd Scrap	38 DI/H ₂ O	50 SS
UF ₆			
HF ACID			52 XM19
UNH			75 INCONEL
UF			
GAS			
FILTER			84 ALUMINUM
B&W			85 ZIRCONIUM

OB [] OTHER

CHEM LAB (WET SPEC GAD)			ENVIRONMENTAL LAB		
METAL IMPURITY	036 SM	200 TITRATION	501 ALKALINITY		
Cd	039 Th	201 U GRAVIMET	503 BIOASSAY		
Fe	040 W	202 UDDM	607 C		
B	050 UNDISS RESIDUE	203 GU/G	609 C		
Si	051 Zn-230	204 GU/I	510 C		
Mn	060 PROD ISO	205 O/U-O/M	611 Fe		
Mg	061 G218 ISO	210 POROSITY	615 F		
Po	062 SCRAP ISO	215 WET FE	625 NH ₃		
Cr	080 P in Zr	250 B/C WT LOSS	626 W + NH ₃		
Al	090 ARCHIVE	252 W/O BORON	630 N		
V	151 C	254 B/C PART DENS	635 NO _x		
Mo	153 S	256 B/C O ₂	636 NO _x		
Sn	155 Cl	296 B/C ROD HT	540 P in Si + Cr		
Cu	156 F	300 OTHER	645 pH		
Na	160 N		647 EC		
Zn	162 H		655 SO ₄ , SO ₃		
Ag	163 H-GUMBALL		SUSPENDED		
Ni	165 O		DISSOLVED		
Co	170 NH ₃		TOTAL		
Ca	171 NH ₄ 3				
Ba	174 pH		660 SO ₄		
Be	175 NORMALITY		665 U		
Hf	176 % HF		670 Zr		
Nb	177 WEIGHT		797 Ti		
Ta	180 OUTGAS		798 Si		
Tl	181 ROD ATM		799 Mn		
P	182 RIBBON HT		800 DINH		

FBI-GYD

183 @ 2/19 TA 1674 (No Supervision on wk end)

540914 showed on 360 Trans

I30

Fe

O/u

Did not print off from LMC S for the original w/
showed Prod 850 + 900 only but listed by 2215
made the correction using 900 Trans.

Told me later that he was not to use it as
he used Supervision Password - That Supervisor
Told him it was all right - That
other id bly used it also.

Asked for me not to say anything about it
or talk about it. (He said he would deny it)
Also stated when he could figure it - he
had used his - (Clock at shift change so there
would be a choice of 6 people who could
done it??

This trans + all supervisor trans. given
out on the "23"; He told me this - and
him about the 18 - he said No - Only That
Supervisor!

Again asked me not to say anything but
still, that, I will not get into trouble

7c

RERUN LOG

Yellow: Manager, Lab
 Pink: Measurement Con
 Gold: File

Technician	Sample #	Test	Test Values*						Assignable Cause/Ca
			Orig	2nd	3rd	x	c	"T"	
13	517249	ISO	2.696						QC Reworked CS 10
	517251	ISO	2.676						" "
	517255	ISO	2.575						" "
	515632	" "	1.761						In Suff airflow
	515781N	" "	3.063						" "
	513540	" "	2.103						" "
	515885	" "	1.632						Reworked by X
	524596	330-083050W	3.212						QC Reworked
100%	524599	295-0830C5R	2.870						✓
3/4	524582	295-083050	2.891						✓
0	525302	330-0830559	3.191						✓
	525304	330-0830560	3.211						✓
	525311	330-0830570	2.215	3.115					✓
	524564	295-0830011	2.863						✓
	524570	295-0830017	2.350						✓
	525312	330-0830576	3.226						✓
	525318	330-0830591	3.205						✓
	525553	3953MS1018	3.827	3.971					✓

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions FOIA 87-88

ENRICHMENT ANALYZER CALIBRATION/VERIFICATION LOG

DATE: 6-22-63
ANALYZER: 3

TIME	SAMPLES	CTD/RLSD	VERIFICATION *			CALIBRATION						
			PROD.	G218	.715 enr	2.223	3.978	REPK*D	3.978 counts	M	K	POSS.EK.SU
23:	/	/										
00:	/	/										
01:12	/	/										
02:14	/	/										
03:26	/	/										
04:	/	/										
05:05	/	/										
06:	/	/										
07:16/2	/	/										
08:	/	/										
09:	/	/										
10:18	/	210		.712		2.215		3.948		2.1521		2.1660
11:	/	/										
12:11	/	/										
13:	/	/										
14:	/	/										
15:	/	/										
16:	/	/										
17:	/	/										
18:	/	/										
19:	/	/										
20:	/	/										
21:	/	/										
22:	/	/										

* Circle OUT-OF-LIMIT values
D D

ENRICHMENT ANALYZER CALIBRATION/VERIFICATION LOG

DATE: 6-23-82
ANALYZER: 3

TIME	SAMPLES	CTD/RLSD	VERIFICATION *			CALIBRATION					
			PROD.	G218	.715 entr	2.223	3.978	REP'D	3.978 counts	H	K
23:	/	/									
00:	/	/									
01:	/	/									
02:	/	/									
03:	/	/									
04:	/	/									
05:	/	/									
06:	/	/									
07:	/	/									
08:	/	/									
09:	/	/									
10:	/	/									
11:	/	/									
12:	/	/									
13:	/	/									
14:	/	/									
15:	/	/									
16:	/	/									
17:	/	/									
18:	/	/									
19:	/	/									
20:	/	/									
21:	/	/									
22:	/	/									

* Circle OUT-OF-LIMIT values.

ENRICHMENT ANALYZER CALIBRATION/VERIFICATION LOG

DATE: 6-27-82
ANALYZER: 3

TIME	SAMPLES CTD/RLSD		VERIFICATION *		CALIBRATION	
	PROD.	G218	.715 enr	2.223	3.978	REPK*D
23:	/	/				
00:	/	/				
01:25	/	/			✓	30.7545*
02:47	/	/				30.260
03:23	/	/	.713	2.226	3.975	30.210
04:	/	/				
05:	/	/				
06:15	/	212	.712	2.222	3.972	30.220
07:	/	/				
08:	/	/				
09:14	/	212	.710	2.220	3.973	30.220
10:	/	/				
11:	/	/				
12:44	/	212	.712	2.222	3.972	30.220
13:	/	/				
14:46	/	210	.711	2.221	3.971	30.220
15:46	/					30.220
16:	/	/				
17:	/	/				
18:	/	/				
19:25	/					30.220
20:02	/					30.220
21:	/	/				
22:	/	/				

* Circle OUT-OF-LIMIT values.

ENRICHMENT ANALYZER CALIBRATION/VERIFICATION LOG

DATE: 10-25
ANALYZER: 3

TIME	SAMPLES CTD/RLSD	VERIFICATION *			CALIBRATION		
		PROD.	G218	.715 enr	2.223	3.978	REPK'D
23: 26	/	2/0	70G	2 207	2 547	/	
00: 41	-/-	5/			-	2% 0.96	6.8957 / 758
01: 41	-/-	6/			-	2% 0.2	6.575 / 747
02: 20	/	1	721	2 224	3 554		
03:	/	1					
04: 42	/	2 1/2	709	2 523	3 943		
05:	/	1					
06:	/	1					
07:	/	1					
08:	/	1					
09: 17	2/10	1	71C	2 237	3 941		
10: 58	7	1	711	2 221	3 948		
11: 14'	1	1	711	2 221	3 948		
12:	1	1					
13:	1	1					
14: 12	1210	1	700	2 212	3 976	1% 765	
15:31	1	1	774	2 140	3577	1% 6555	65116
16: 23	1	1	774	2 140	3577		
17:	1	1					
18:	1	1					
19:	1	1					
20:31	1	212	713	7321	3572		
21:55	1	111	711	2 111	3534		
22:	1	1					

* Circle OUT-OF-LIMIT values.

ENRICHMENT ANALYZER CALIBRATION/VERIFICATION LOG

DATE: 6-24
ANALYZER: 3

H 4(2)

TIME	SAMPLES CTD/RUSD	VERIFICATION *				CALIBRATION				
		PROD.	.0218	.715 enr	2.223	3.978	REP#D	3.978 counts	M	R
23:	/	/								
00:22	/	1	1	1	1					
01:29	/	1	1	1	1					
02:00	/	1	1	1	1					
03:	/	1	1	1	1					
04:	/	1	1	1	1					
05:	/	1	1	1	1					
06:	/	1	1	1	1					
07:25	1210	1	1	1	1					
08:02	✓2	1	1	1	1					
09:17	1	1	1	1	1					
10:	1	1	1	1	1					
11:54	717	1	1	1	1					
12:	1	1	1	1	1					
13:	1	1	1	1	1					
14:	1	1	1	1	1					
15:	1	1	1	1	1					
16:	1	1	1	1	1					
17:	1	1	1	1	1					
18:	1	1	1	1	1					
19:	1	1	1	1	1					
20:	1	1	1	1	1					
21:	1	1	1	1	1					
22:	1	1	1	1	1					

* Circle OUT-OF-LIMIT values.

IAIR
6

CALIBRATION DATA

STANDARD VERIFICATION LIMITS (9825A)			
ANALYZER #011	ANALYZER #012	ANALYZER #013	ANALYZER #014
.719 ± .009	.717 ± .012	.714 ± .009	.715 ± .008
2.223 ± .013	2.231 ± .015	2.220 ± .012	2.224 ± .014
3.982 ± .019	3.982 ± .024	3.976 ± .021	3.966 ± .021

REGRESSION MODEL COEFFICIENT (NP-80)		
ANALYZER	SLOPE	INTERCEPT
011	1.003968	-0.006642
012	1.005995	-0.009013
013	1.003739	-0.002588
014	1.005886	-0.003204

BIAS ADJUSTMENT FACTOR AND MINIMUM "U" COUNT LIMITS (9825A)			
ANALYZER	SLOPE	INTERCEPT	MINIMUM "U" COUNT
011	1.00593	-0.00422	285048
012	1.00901	-0.00522	289827
013	1.005711	-0.00599	290003 283264
014	1.00569	-0.01038	287744

6-23-83

NOTE: Prior each calibration, the "PP" counts for the lowest enrichment and three "P" count greater than the "B" count for the highest enrichment.

APPROVED BY

DATE ISSUED - 6/23/83

ATTN: Lab Supervisor

LMCR - STANDARD OUT OF LIMITS NOTICE, SEQUENCE # 944
INPUT VERIFICATION RESULTS
OUT-OF-ALARM

DATE: 12-FEB-84
TIME: 0415R114

TECH PAY #: [REDACTED] TECH NAME: [REDACTED] T#111
METHOD CODE: 6000 METHOD NAME: Prod Isotopic
ANALYZER #: 12 ANALYZER NAME: ATU ENRICHMENT #2
STD TYPE #: 901 STANDARD BATCH NAME: .715 ISO. VERIF.
UNIT CODE: 60 UNIT NAME: Z II-235

UPPER CTL : 0.73100
UPPER ALRM : 0.72400
NOMINAL : 0.71500
LOWER ALRM : 0.70600
LOWER CTL : 0.69900
POP SIGMA : 0.00480

VERIF STD #: 200011 RESULT: 0.72500

Was value input PROPERLY? YES ... NO -- --
If answer is NO, attach Transaction R20 report.

What do control chart and trend data show?

Have there been any changes or unusual events?
YES ... NO --
If YES, what? -----

Are there any distinctions with this analysis? (No other standards show same
Analyze different? Preparation? Tech?, Calibration?) YES ... NO ...
If YES, what? -----

Did investigation show a most likely cause? YES ... NO ...
If YES, what? -----

What corrective action was taken?
When? ----- RH WTRM? -----

Chemist approval: ----- Date: -----

Closed out on t790 ... DATE: -----

LMIA
7

TABLE FOR SCP-401

STANDARD VERIFICATION LIMITS (9825A)			
ANALYZER #011	ANALYZER #012	ANALYZER #013	ANALYZER #014
.719 ± .009	.717 ± .012	.714 ± .009	.715 ± .008
2.223 ± .013	2.231 ± .015	2.220 ± .012	2.224 ± .014
3.982 ± .019	3.982 ± .024	3.976 ± .021	3.966 ± .021

REGRESSION MODEL COEFFICIENT (HP-80)		
ANALYZER	SLOPE	INTERCEPT
011	1.003968	-0.006642
012	1.005995	-0.009013
013	1.003739	-0.002528
014	1.005886	-0.006204

BIAS ADJUSTMENT FACTOR AND MINIMUM "U" COUNT LIMITS (9825A)			
ANALYZER	SLOPE	INTERCEPT	MINIMUM "U" COUNT
011	1.00593	-0.00422	285048
012	1.00904	-0.00522	289827
013	1.00574	-0.00599	290003 283064
014	1.00569	-0.01038	287744

6-26-83

NOTE: Check each calibration. The "B" counts for the lowest enrichment must have a "B" count greater than the "B" count for the highest enrichment.

APPROVED BY

6/23/83

DATE ISSUED -

DATE 10-24-83

CC:

Clement Syporino

JECT

Concerns of Clement Syporino

Significant Concern: Personnel feel they are being pressured to meet production needs and are finding themselves running standards until they are acceptable to referee events.

Personnel also feel that the individual that runs OHS & O&S valves is looked at differently from those who seldom have OHS & O&S valves.

Personnel understand how limits are established, but feel limits are based ^{standard regula} in order to satisfy production oriented ex's.

I feel a meeting is needed to address employees' concerns and reduce self imposed pressure to produce.

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions
FOIA: 87-88

C-8

NAME PLAINLY

1-33

COMPLETE MAILING ADDRESS

PHONELINE NUMBER

GENERAL  ELECTRIC

DIAL COMM: 8*292-5461 DATE: January 19, 1984

EDITION:

H-92
K-10
J-02

DEPT: WILMINGTON MANUFACTURING DEPARTMENT

ADDRESS: M/C H-92

SUBJECT: ACTION PLANNED PER OUR 01/12/84 MEETING

TO: *JTC*

As indicated to you in our meeting of 01/12/84, actions have/will be taken in the following areas.

- Flexibility for you to change your password has been limited due to the required interactions of two computers, LMCS and HP9887, both of which require exact password matching. Specifically, [REDACTED] the only Lab LMCS authorized password accessor, will be set up by 01/25/84 to also be the only authorized HP9887 password accessor. At any time you are scheduled off for more than 48 hours, your in-process work on the HP9887 should be cleared such that you may contact John to effect a password change for you. Based on your 4-shift rotation, you should be able to change your password upon reporting to either the 3:00 - 11:00 shift or the 7:00 - 3:00 shift.
- LMCS 902/903 transaction password control, when initially set up, involved the ability to change values for standards results and, consequently, access was restricted to supervisor passwords only. Sometime ago, these transactions were modified to eliminate access to changing standards results. Use of the current 902/903 transaction by qualified lab personnel is appropriate for routine correction for input errors. However, if intentional falsification of results occurred, traceability to who/why changes were made would not be available. Therefore, transactions 902/903 will not be made available to others until they are modified to record and print out on each changed result report to document the person, reason and date of the changes. In the interim, all supervisory passwords will be changed on a continuing basis to eliminate indiscrete use by others.

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions
FOIA: *87-88*

0.9

175
January 19, 1984

-Page 2-

- When the HP9887 computer was implemented in November 1982, the pay number of the operator performing the sample solution weighing was used as the pay number to be reported on the LMCS results report. Since the possibility does exist that the shift operator who transacts the data on the HP9887 to LMCS may allow incorrect data to be transmitted, the LMCS results report will then show this incorrect results and the pay number of the non-transmitting operator. Since it is not possible to enter one pay number which will represent the performance of the total steps in completing a sample, I have decided to change the pay number reported to LMCS to be the pay number of the operator releasing the results. This change was implemented effective 01/17/84.

With respect to samples which you have identified on the rerun log as requiring rerun and subsequently being deleted by another operator, from the Lab viewpoint, I am confident that the deletions are being properly handled. Per procedure, a sample outside the expected range is to be rerun if C&R personnel indicate it is from a source which was expected to meet the range. Your C&R contact advises to rerun the sample and a subsequent C&R contact advises not to rerun. The problem lies with differing C&R decisions. From a quality viewpoint, samples outside the expected range which are not rerun will not be released by C&R. From a Lab viewpoint, we should not expend the retest labor to rerun samples not requiring reruns. As an overview of the total isotopic measurement/quality release system, no specific problems are apparent due to samples not being rerun which should have been rerun.

[REDACTED]
Manager
Chemet Laboratories, M/C H-92

TPW/dvc

WNY 51

Rumford
standard
1/10

SAMPLE	TUBE	CM	TYPE	#A	TARE	SAMPLE	WT	WT	SOL	GUG
548613	6300	.715	PROD	1	29.2573	4.5074	33.9127	60.2942	.1272	
548613	6324	.715	PROD	1	28.5060	4.5062	33.1599	58.9051	.1298	
548614	6381	2.223	PROD	1	27.8653	4.5058	32.4945	58.2618	.1291	
548614	6414	2.223	PROD	1	28.3065	4.5057	32.9353	58.6915	.1292	
548615	6398	2.999	PROD	1	27.2099	4.5031	31.8118	58.6773	.1240	
548615	6292	2.999	PROD	1	29.3095	4.5082	33.9194	60.4669	.1255	
548618	6443	3.950	PROD	1	25.1111	4.5064	29.7653	55.5568	.1236	
548618	6433	3.950	PROD	1	27.8814	4.5027	32.5317	59.5444	.1245	
548619	6261	4.000	PROD	1	27.9034	4.5041	32.5682	59.4189	.1254	
548620	6456	4.000	PROD	1	29.2327	4.5019	32.8901	58.7270	.1295	
548620	6389	2.380	PROD	1	23.8436	4.4958	28.4945	54.2973	.1295	
548620	6480	2.380	PROD	1	24.5718	4.4969	29.2250	55.0941	.1293	

Atty
Integrity is important when following STEX
another Tech. Work - 1

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions
FOIA 87-88

0-10

WK14 STD
Bad Powder

SAMPLE	TUBE	NOM	TY	E	#H	TARE	SAMPLE	IGN	SOL	GUG	U
*****	***	***	***	**	***	***	*****	**	***	***	*
40474	0	2.999	PROD	1		30.2637	4.5074	34.8710	0.0000	0.0000	2.
40474	0	2.999	PROD	1		31.1014	4.5061	35.7060	0.0000	0.0000	2.
40474	0	2.999	PROD	1		27.6084	4.5080	32.2148	0.0000	0.0000	2.
40474	0	2.999	PROD	1		27.8654	4.5029	32.4655	0.0000	0.0000	2.

E-CR14 3695

SAMPLE PREPARATION DATA

INITIAL WEIGHTS				IGNITED WEIGHTS				SOLUTION WEIGHTS			
E	TR	OPER	DATE	TARE	U02	OPER	DATE	U308	OPER	DATE	SOLUTIO
3	8		39	29.2573	4.5074		39	33.9127		39	60.2942
3	8		39	28.5050	4.5062		39	33.1599		39	58.9051
4	8		39	27.8653	4.5058		39	32.4945		39	58.2618
4	8		39	28.3065	4.5057		39	32.9353		39	58.6915
5	8		39	27.2099	4.5031		39	31.8118		39	58.6773
5	8		39	29.3095	4.5082		39	33.9194		39	60.4669
8	8		39	25.1111	4.5064		39	29.7653		39	55.5568
8	8		39	27.8814	4.5027		39	32.5317		39	59.5444
9	8		39	27.9094	4.5041		39	32.5682		39	59.4189
9	8		39	28.2327	4.5019		39	32.6901		39	58.7270
0	8		39	23.8436	4.4958		39	28.4345		39	54.2973
0	8		39	24.5718	4.4969		39	29.2250		39	55.0941

REPORT FOR ANALYZER# 3

DATE: 84042 TIME: 4:42

BEFORE STANDARDS

LOW STD	.702
MED STD	2.220
HIGH STD	3.973

POST STANDARDS

LOW STD	.712
MED STD	2.223
HIGH STD	3.980

CONSTANTS

K	61270.509585
K	1.857195
SLOPE	1.001661
INTER	-.001195

E	OPER	TUBE	NOM ENR	APP ISO	COR ISO	GUL	GUG	O/U	ENR	Avg ENR	SEN1 Y/N
3		6300	0.715	0.792	0.713	172.1	1272	2.105	0.713	0.713	YES
3		6324	0.715	0.808	0.712	175.6	1298	2.110	0.712	0.712	YES
4		6381	2.223	2.501	2.216	175.4	1291	2.199	2.218	2.218	YES
4		6414	2.223	2.482	2.190	175.5	1292	2.200	2.201	2.201	YES
5		6398	2.999	3.223	2.984	165.7	1240	2.290	2.988	2.988	YES
5		6292	2.999	3.259	2.979	159.2	1255	2.280	2.983	2.983	YES
8		6443	3.950	4.432	3.910	176.3	1296	2.110	3.916	3.916	YES
8		6422	3.950	4.350	3.970	177.2	1295	2.113	3.920	3.920	YES

VARIATION DATA

IGNITED WEIGHTS

OPER DATE U308

39 33.9127
39 33.1599
39 32.4945
39 32.9353
39 31.8118
39 33.9194
39 29.7653
39 32.5317
39 32.5682
39 32.6901
39 28.4945
39 29.2250

SOLUTION WEIGHTS

OPER DATE SOLUTION

39 60.2941
39 58.9051
39 58.2618
39 58.6915
39 58.6773
39 60.4669
39 55.5568
39 59.5444
39 59.4189
39 58.7220
39 54.2973
39 55.0941

ANALYZER 3

TIME: 4:42

STANDARDS

CONSTANTS

.712 M 61270.509585
2.223 K 1.857195
3.980 SLOPE 1.001661
INTER -.001195

GUL	GUG	O/U	ENR	Avg	SEND
			ENR	ENR	Y/N
172.1	.1272	2.109	0.713	0.713	YES
175.6	.1290	2.110	0.712	0.712	YES
175.4	.1291	2.199	2.218	2.218	YES
175.5	.1292	2.200	2.201	2.201	YES
165.7	.1240	2.290	2.988	2.988	YES
159.2	.1255	2.280	2.983	2.983	YES
178.5	.1296	2.110	3.916	3.916	YES
167.2	.1242	2.110	3.620	3.620	YES

JAN
TIME: 5:01
DATE: 94042

MPL	TUBE	NOM	TYPE	#A	TARE	SAMPLE	IGN WT	SOL WT	GUG	0
8613	6300	.715	PROD	1	29.2573	4.5074	33.9127	60.2942	.1272	2.
8613	6324	.715	PROD	1	28.5060	4.5062	33.1599	58.9051	.1298	2.
8614	6381	2.223	PROD	1	27.8653	4.5058	32.4945	58.2618	.1291	2.
3614	6414	2.223	PROD	1	28.3065	4.5057	32.9353	58.6915	.1292	2.
3615	6398	2.999	PROD	1	27.2099	4.5031	31.8118	58.6773	.1240	2.
8615	6292	2.999	PROD	1	29.3095	4.5082	33.9194	60.4669	.1255	2.
8618	6443	3.950	PROD	1	25.1111	4.5064	29.7653	55.5568	.1296	2.
3618	6433	3.950	PROD	1	27.8814	4.5027	32.5317	59.5444	.1245	2.
8619	6261	4.000	PROD	1	27.9094	4.5041	32.5682	59.4189	.1254	2.
8619	6456	4.000	PROD	1	28.2327	4.5019	32.8901	58.7270	.1295	2.
8620	6389	2.380	PROD	1	23.8436	4.4958	28.4945	54.2973	.1295	2.
3620	6480	2.380	PROD	1	24.5718	4.4969	29.2250	55.0941	.1293	2.0

IN ALL ANAs

X-X-3-✓

PRODUCTION
g/u & ISOTOPIC
& O/U RATIO

199
6
4

CONTAINER #	CRUC. POS.	TARE WGT.	UO ₂ WGT.	TARE + U ₃ O ₈ SOLUTION	TUBE #	MEAS. U 235	CALC. U 235	O/U RATIO
103/6	1				6300			
# 715	2				6324			
30957	3				6381			
2.223	4				6414			
40474	5				6398			
# 2.999	6				6292			
52435	7				6412			
3.950	8				6433			
100120J0141	9				6361			
785	10				6456			
694164	11	23.8436	4.4958		6399			
2.380	12	24.5778	4.4969		6480			
OX. IN		OX. OUT						
<u>2/18/84</u>		<u>2/18/84</u>						
<u>C.S. 0-8</u>		<u>1015</u>						

REPORT # 3682

SAMPLE PREPARATION DATA

FILE	TR	OPER	INITIAL WEIGHTS		IGNITED WEIGHTS		SOLUTION WEIGHT			
			DATE	TARE	U02	OPER	DATE	U308	OPER	DATE
348	19		40	29.3094	4.5061	0	40	33.9939	40	59.3
348	19		40	23.6955	4.5001	0	40	28.3732	40	54.7
348	19		40	23.4706	4.5066	0	40	28.1558	40	54.7
481	19		40	29.1359	4.5006	0	40	33.8143	40	60.3
481	19		40	30.1551	4.5070	0	40	34.6399	40	61.1
481	19		40	27.7979	4.5067	0	40	32.4827	40	58.8
306	19		40	26.4490	4.5034	0	40	31.1301	40	57.4
306	19		40	29.5241	4.5004	0	40	34.2023	40	60.3
306	19		40	28.7232	4.5058	0	40	33.4076	40	58.7
305	19		40	28.9073	4.5098	0	40	33.6248	40	59.8
305	19		40	28.7040	4.5052	0	40	33.3872	40	59.8

REPORT FOR ANALYZER# 1

DATE: 84041 TIME: 10:38

BEFORE STANDARDS

POST STANDARDS

CONSTANTS

LOW STD	.719	LOW STD	.715	M	66880.883209
MED STD	2.220	MED STD	2.214	K	1.697680
HIGH STD	3.961	HIGH STD	3.961	SLOPE	1.003347
				INTER	-.003590

FILE	OPER	TUBE	NOM ENR	APP ISO	COR ISO	GUL	GUG	O/U	ENR	Avg ENR	SE Y%
348		6322	3.300	3.791	3.279	180.4	.1321	1.999	3.286	3.286	YE
348		6329	3.300	3.635	3.263	172.7	.1276	2.001	3.271	3.271	YE
348		6372	3.300	4.987	4.494	172.0	.1271	1.998	4.505	0.000	NO
481		6203	0.710	0.820	0.738	172.2	.1273	2.000	0.737	0.737	NO
481		6199	0.710	0.821	0.734	173.7	.1281	2.001	0.733	0.733	YE
481		6319	0.710	0.823	0.738	173.1	.1278	2.000	0.736	0.736	NO
306		6410	0.710	1.170	1.045	173.7	.1281	2.001	1.045	1.045	NO
306		6393	0.710	0.818	0.727	174.7	.1287	2.000	0.726	0.726	YE
306		6195	0.710	0.864	0.745	181.0	.1324	1.998	0.744	0.744	NO
		6202	0.710	0.820	0.731	174.7	.1287	2.001	0.720	0.720	YE

ISI Report # 3667
Transmission Status
Released by: 22072
TIME: 10:48
DATE: 84041

Isles Transmitted To LMCC

Sample#	Tube	Iso	O/U	Avg	Avg
552481	6203	.737	2.000	.737	0.000
552481	6199	.733	2.001	.733	0.000
552481	6319	.736	2.000	.736	0.000
552305	6313	.730	2.001	.730	0.000
552305	6210	.730	2.002	.730	0.000
552305	6317	.726	2.000	.726	0.000

Isles Held For Recount

Sample#	Tube	Iso	O/U	Avg	Avg
552348	6372	4.505	1.998	0.000	0.000

Isles Withheld

Sample#	Tube	Iso	O/U	Avg	Avg
552348	6322	3.286	1.999	3.286	0.000
552348	6329	3.271	2.001	3.271	0.000
552348	6372	4.505	1.998	0.000	0.000
552306	6410	1.045	2.001	1.045	0.000
552306	6393	.726	2.000	.726	0.000
552306	6195	.744	1.998	.744	0.000

210
19
19

REGULATIONS
gU/g ISOTOPIC
6 O/U RATIO

CONTAINER #	CRUC POS.	TARE WGT.	UO ₂ WGT.	TARE + U ₃ O ₈	TARE + SOLUTION	TUBE #	MEAS. U ₂₃₅	CALC. U ₂₃₅	O/U RATIO
33072A1440	1					6322	1		
"	2					6329			
"	3					6372	0000	71.505	1016
2712299910	4					6203			
"	5					6199			
"	6					6249			
2712299912	7					6410	1.045	36.82	100.0
"	8					6393			
"	9					6195			
2712299915	10					6213			
"	11					6210			
"	12					6247			
8.84		OK. IN	OK. OUT				101.0		
1300		2/9/84	2/9/84	2.5.67			07.30		
MAG		1/45 - T# - 1600 - A40	-	19.56			02.17		

REPORT # 271A

SAMPLE PREPARATION DATA

PLE	TR	OPER	INITIAL WEIGHTS		IGNITED WEIGHTS		SOLUTION WEIGH		
			DATE	TARE	U02	OPER	DATE	U308	OPER
291	12		41	26.8801	4.5054	0	42	33.5349	42
290	12		41	26.8817	4.5071	0	42	31.5339	42
289	12		41	29.6177	4.5060	0	42	34.2654	42
280	12		41	31.3782	4.5011	0	42	36.0201	42
287	12		41	27.8660	4.5031	0	42	32.5123	42
286	12		41	27.8457	4.5009	0	42	32.4361	42
285	12		41	29.7818	4.5007	0	42	34.4375	42
281	12		41	30.1222	4.5032	0	42	34.7802	42
280	12		41	26.4491	4.5100	0	42	31.1205	42
279	12		41	25.9542	4.5002	0	42	30.5767	42
278	12		41	29.2454	4.5099	0	42	33.8821	42
277	12		41	27.6081	4.5035	0	42	32.2557	42

REPORT FOR ANALYZER # 4

DATE: 84042 TIME: 20:40

BEFORE STANDARDS

POST STANDARDS

CONSTANTS

LOW STD .708
MED STD 2.221
HIGH STD 3.976

LOW STD .709
MED STD 2.228
HIGH STD 3.986

M 62601.277938
K 1.866631
SLOPE 1.001144
INTER .000786

PLE	OPER	TUBF	NOM ENR	APP ISO	COR ISO	GUL	GUG	D/U	ENR	Avg ENR	SE Y%
291		6442	1.500	1.685	1.510	172.0	.1271	2.104	1.521	1.521	YE
90		6304	0.710	0.793	0.715	171.9	.1271	2.120	0.716	0.716	YE
90		6452	0.710	0.778	0.715	168.1	.1248	2.132	0.717	0.717	YE
		6370	0.710	0.786	0.720	168.9	.1253	2.135	0.721	0.721	YE
		6336	0.710	0.797	0.711	174.1	.1284	2.126	0.712	0.712	YE
		6387	0.710	0.801	0.716	173.6	.1280	2.139	0.718	0.718	YE
		6344	1.600	1.778	1.555	178.0	.1307	2.083	1.558	1.558	MD
		6292	2.000	2.278	2.011	177.1	.1204	2.004	2.004	2.004	MD

J REFURB 1705

SAMPLE PREPARATION DATA

SAMPLE	TR	OPER	DATE	INITIAL WEIGHTS		IGNITED WEIGHTS		SOLUTION WEIGHT			
				TARE	U02	LIPER	DATE	U308	OPER	DATE	SOL
1283	9		41	31.1012	4.5091		0	42	35.7540	42	60.
1284	9		41	28.3236	4.5083		0	42	32.9749	42	59.
1270	9		41	29.3502	4.5004		0	42	33.9974	42	59.
1271	9		41	28.3146	4.5056		0	42	32.9641	42	59.
1272	9		41	28.5341	4.5029		0	42	33.1469	42	59.
1273	9		41	29.2395	4.5053		0	42	33.9004	42	59.
1274	9		41	27.9978	4.5027		0	42	32.6036	42	57.
1275	9		41	25.4977	4.5023		0	42	30.1177	42	55.
1276	9		41	30.3573	4.5060		0	42	35.0085	42	60.
1282	9		41	27.8274	4.5033		0	42	32.4404	42	57.
1293	9		41	29.0723	4.5086		0	42	33.7321	42	58.
1292	9		41	28.7251	4.5048		0	42	33.3342	42	58.

REPORT FOR ANALYZER# 1

DATE: 84042 TIME: 19:52

BEFORE STANDARDS

LOW STD	.723
MED STD	2.226
HIGH STD	3.974

POST STANDARDS

-LOW STD	.713
MED STD	2.224
HIGH STD	3.976

CONSTANTS

M	66414.645614
K	1.765327
SLOPE	1.003347
INTER	-.003590

SAMPLE	OPER	TUBE	NOM	APP	COR	GUL	GUG	D/U	ENR	AVG	S
			ENR	ISO	ISO					ENR	S
1283		6389	0.710	0.841	0.726	180.7	.1322	2.125	0.725	0.725	Y
1284		6292	0.710	0.798	0.714	173.6	.1280	2.127	0.712	0.712	Y
1270		6480	3.300	3.738	3.287	176.9	.1300	2.113	3.296	3.296	N
1271		6398	0.710	0.836	0.751	172.5	.1274	2.124	0.750	0.750	Y
1272		6393	2.000	2.226	1.994	173.2	.1278	2.249	1.997	1.997	Y
1273		6459	1.600	1.853	1.607	179.9	.1318	2.081	1.608	1.608	Y
1274		6325	2.000	2.284	2.000	177.8	.1305	2.274	2.003	2.003	Y
1275		6458	2.400	2.731	2.418	175.6	.1292	2.220	2.422	2.422	Y

File Report# 0705
Transmission Status
Released by: 15394
TIME: 20:01
DATE: 84042

Files Transmitted To LMCS

Sample#	Tube	Iso	O/U	Avg	Avg
				Iso	O/U
441283	6389	.725	2.125	.725	0.000
441284	6292	.712	2.127	.712	0.000
441270	6480	1.296	2.113	1.296	0.000
441272	6393	1.997	2.249	1.997	0.000
441273	6459	1.608	2.081	1.608	0.000
441274	6325	2.003	2.274	2.003	0.000
441275	6458	2.422	2.220	2.422	0.000
441276	6298	1.993	2.119	1.993	0.000
441293	6372	.723	2.097	.723	0.000
441292	6294	1.970	2.270	1.970	0.000

Files Held For Recount

Sample#	Tube	Iso	O/U	Avg	Avg
				Iso	O/U

Files Withheld

Sample#	Tube	Iso	O/U	Avg	Avg
				Iso	O/U
441271	6398	.750	2.124	.750	0.000
441282	6360	1.900	2.250	1.900	0.000

1000 - STANDARDS OUT OF TOLERANCE FREQUENCY &
INPUT VERIFICATION RESULTS
OUT-OF-ALARM

DATE: 12-FF-87
TIME: 0415R114

TECH PAY #: [REDACTED]
METHOD CODE: 4000
ANALYZER #: 12
STD TYPE #: 901
INIT CODE: 40

TECH NAME: [REDACTED] 7C - T8111
METHOD NAME: Prod Testopac
ANALYZER NAME: AIU ENHANCED #
STANDARD BATCH NAME: .715 IRD. UFRIF.
LINT NAME: % II-235

UPPER CTI: 0.73100
UPPER ALRM: 0.72400
NOMINAL: 0.71500
OWER ALRM: 0.70600
OWER CTI: 0.69900
OP SIGMA: 0.00480

UFRIF STD #: 2000011 RESULTS:

0.72500

Is value input properly? YES .. NO ..
If answer is NO, attach Transaction R20 REPORT.

What do control chart and trend data show?

Have there been any changes or unusual events?

ES ... NO ..

If YES, what?

Are there any distinctions with this analysis? (In other standards show to analyze different? Preparations? Tech? Calibration?) YES .. NO ..
If YES, what?

In investigation, show a most likely cause? YES .. NO ..

If YES, what?

What corrective action was taken?

When?

Manager approval?

Locked out on t290

DATE

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions
FOIA 6+7C
87-88

0-11

SAMPLE PREPARATION DATA

FILE	TR	INITIAL WEIGHTS			IGNITED WEIGHTS			SOLUTION WEIGHTS			
		OPER	DATE	TARE	UO2	OPER	DATE	UO2B	OPER	DATE	SOL.
'955	1	[REDACTED]	65	27.6937	4.5121	[REDACTED]	65	32.5673	[REDACTED]	65	55.11
'955	1	[REDACTED]	65	26.1115	4.4996	[REDACTED]	65	30.7231	[REDACTED]	65	54.11
'967	1	[REDACTED]	65	25.6245	4.4522	[REDACTED]	65	30.4597	[REDACTED]	65	53.11
'967	1	[REDACTED]	65	27.0858	4.5054	[REDACTED]	65	32.5497	[REDACTED]	65	54.11

REPORT FOR ANALYZER*****

DATE: 03065 TIME: 12:15C

BEFORE STANDARDS

LOK STD	3.724
MED STD	3.726
HIGH STD	3.932

POST STANDARDS

LOK STD	3.718
MED STD	3.727
HIGH STD	3.936

CONSTANTS

N	70542.707754
SLOPE	1.000517
INTER	-0.005142

FILE	OPER	TUBE	NO.	APP	CCA	SUL	SLG	D/U	ENR	AVG	ENR
			ENR	XISO	ISO						
'955	[REDACTED]	7841	3.300	4.145	3.227	264.5	1445	2.061	3.300	3.224	3
'955	[REDACTED]	7696	3.300	4.017	3.226	196.2	1401	2.059	3.300	0.000	1
'967	[REDACTED]	7890	3.350	4.775	3.645	196.2	1402	2.070	3.350	2.876	2
'967	[REDACTED]	7524	3.350	5.076	3.746	211.1	1411	2.071	3.350	2.000	1

RELEASED TO LHC5

DATE

ATTN: Lab Supervisor

IMCR - STANDARD OUT OF LIMITS NOTICE: SEQUENCE # 9
INPUT VERIFICATION RESULTS
OUT-OF-ALARM

DATE: 12-FEB-84
TIME: 0415R:14

TECH PAY #: [REDACTED]
METHOD CODE: 4000
ANALYZER #: 12
STD TYPE #: 901
UNIT CODE: 60

UPPER CTL: 0.73100
LOWER ALRM: 0.72400
NOMINAL: 0.71500
LOWER ALRM: 0.70600
LOWER CTL: 0.69900
>OP SIGMA: 0.00480

VERIF STD #: 200011 RESULT: 0.72500

Has value INPUT PROPER? YES NO
If answer is NO, attach Transaction R20 report.

What do control chart and trend data show?.....

Have there been any changes or unusual events?

YES NO
If YES, what?.....

Are there any distinctions with this analyzer? (Do other standards show a
analyze different? Preparation: Tech?, Calibration?) YES NO
If YES, what?.....

Did investigation show a most likely cause? YES NO
If YES, what?.....

What corrective action was taken?.....
When? By whom?

Chemist APPROVAL: Date:

Clocked out on t790 DATE:

2/12/84 @ 2:19

In 237 Diss 1

TYPE: 2154
DATE: 1/4/84

16981 16982

AMPLE	TUBE	NOM	TYPE	#A	TARE	SAMPLE	WT	WT	GUG
49533	6445	2.200	PROD	1	28.7045	4.5065	33.5725	60.4012	1249
49529	6445	2.200	PROD	1	29.5529	4.5061	34.1566	61.0827	1233
49531	6324	2.200	PROD	1	25.5664	4.5047	30.1542	56.7380	1246
49532	6317	2.200	PROD	1	29.5770	4.4920	34.2279	60.5414	1274
50626	6201	1.200	PROD	1	29.1189	4.4914	33.7609	60.6437	1249
50626	6320	1.200	PROD	1	30.9367	4.4938	35.5819	62.5260	1247
52000	6456	1.600	PROD	1	32.4872	4.4983	37.1306	62.0211	1332
52000	6433	1.600	PROD	1	20.1129	4.4965	26.7548	61.4866	1254
35968	6485	.710	PROD	1	29.1362	4.4939	33.7714	60.4066	1267
35968	6414	.710	PROD	1	30.2853	4.5011	34.9262	62.2688	1230
49536	6443	2.700	PROD	1	27.2104	4.4922	31.7865	59.0433	1219
49537	6381	2.200	PROD	1	30.1435	4.4951	34.8014	60.3703	1307

Micro


2/12/84 C 1 9

6.6.201
6.6.269
6.6.117
6.6.20
6.6.324
6.6.301
6.6.418
6.6.430
6.6.445
6.6.446
6.6.565

* * * *

- Deleted

216 Train?



Dec 2135
1/4ed 2

about this & he said he received this in
+ also about 3 or 4 results on this same can!

cked this up @ 1/3/84 @ 2130 & thought it was strange!

Superseded! (Christine h Wk End - Com. back on Fri or S

CHNET LABORATORY SAMPLE REPORT

8 MAIL TO:

SAMPLE #: 240484 NO. OF TESTS : 2

PAGE: 1
SUBSET-21-AUG-81
CHNET 10128

SAMPLER'S FAY #: [REDACTED]
REQ. COMPONENT #: Y70
DOCUMENT NO.: L01-015
CONTAINER/LUT #: 80592

DATE SAMPLED : 23-JUL-81
TIME SAMPLED : 13100
AREA : LAB
MATERIAL CLASS : 6 SCRAP
SOURCE : RIF007

ANLS METHOD:

CODE EX NAME	RESULT	UNITS	TECH	COMPLETED
62 00 SURAF ISOTOPICS	3.921	% U-235	[REDACTED]	20-AUG-81 1710
200 00 U Titration	68.74	% U	[REDACTED]	20-AUG-81 1611

My Copy To -
I placed on his desk!

AMPL # : 240489 NO. OF TESIS : 2
PRRRRRRR

SAMPLER'S PAY #: [REDACTED]
REQ. COMPONENT #: 970
DOCUMENT NO. : LOI-015
CONTAINER/LOT #: 80592

FILED : 1
DATE : 20-AUG-81
TIME : 17108
DATE SAMPLED : 23-JUL-81
TIME SAMPLED : 13100
AREA : LAB
MATERIAL CLASS : 6 SCRAP
SOURCE : KIF007

ANLS METHOD

CODE EX NAME	RESULT	UNITS	TECH	COMPLETE
62 00 SCRAP ISOTOPICS	3.921	% U-235	[REDACTED]	20-AUG-81 1/1
200 00 U Titration	0.00	% U	[REDACTED]	20-AUG-81 1/1

this is what I
said it at he found on his deck, first -
note found my copy!

16488
16488

TESTED BY: 16488

TUBE	SAMPLE	TRAY
6121	441282	3
6283	441285	3
6334	552306	3
6348	441282	3
6360	441285	3
6389	552348	3
6394	441294	3
6398	441271	3
6405	441271	3

T1 #3 T1 240 (3rd mylay Run
Did Not Show TT # -
D input here for him &
Password:

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemption 6
FOIA 67-48

O-12

GENERAL ELECTRIC**STRICTLY PRIVATE**

MAIL CODE# 00292-5747

DATE: February 15, 1984

COPIES:

DEPT: WILMINGTON MANUFACTURING DEPARTMENT

ADDRESS: W/C 3-02

SUBJECT: ALLEGATION OF VIOLATIONS TO COMPANY PRACTICES AND PROCEDURES

(715-)

You recently indicated having knowledge and proof of co-workers and supervisory personnel violating established work practices and procedures within the Chemet Laboratory. On January 20, 1984 you stated you would be taking your information to authority other than General Electric. Moreover, you have specifically refused to share this information with me as the responsible management representative of the Company.

You, as an employee of the General Electric Company, have a responsibility to fully apprise management of any such problems of which you become aware. You are expected to comply with that obligation. Therefore, you are hereby directed to discuss with me, in full detail, your concerns relative to Laboratory and Company practices and procedures by 5:00 P.M., Friday, February 24, 1984. This meeting is to be held in my office and all evidence that you have supporting your allegations must be presented.

Failure to comply with this instruction will be considered insubordination and will result in severe disciplinary action, up to and including discharge.



Manager
Quality Assurance

EAL/HDC/bst

0-13

DIAL COMM# 8*292-5747 DATER March 13, 1984 COPIES [REDACTED]
DEPT# WILMINGTON MANUFACTURING DEPARTMENT
ADDRESS# M/C J-02
SUBJECT# CHEMET LABORATORY SAFETY REVIEW

[REDACTED], Manager
Regulatory Compliance

In accordance with our discussion, I am requesting you to conduct a safety review of the Chemet Laboratory. The basis for the review is delineated in the attached letter titled "Allegation of Violations to Company Practices and Procedures in Chemet Laboratory" submitted by [REDACTED] dated February 21, 1984.

I anticipate your review will cover both radiological and industrial safety issues that may exist in the laboratory and that you will serve as chairman of the review team consisting of both radiological and industrial safety experts.

The review should include but not be limited to:

- Specific allegations contained within [REDACTED] Letter, [REDACTED]
- Determination of the validity of the accusation that supervisors and management personnel knowingly bypass safety requirements.
- A review of the overall safety program as it applies to the laboratory (i.e., adequacy of training, radiation safety involvement, etc.).
- Adequacy of response when a safety problem is brought to management's attention.

I would like to have this review conducted PW-12. Please review your plan to conduct this review with me by Friday, March 16.

[REDACTED]
Manager

Portion Excluded

O-14

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions 6, 7C
FOIA: 87.44

February 24, 1984

To: [REDACTED] Mgr., Quality Assurance
General Electric

Re: Allegations of Violations to Company Practices and
Procedures

Dear [REDACTED]:

In reference to your February 15th letter to me concerning the above subject, I have enclosed a copy of my "Allegations of Violations to Company Practices and Procedures in Chemet Laboratory."

In it, I refer to dated documents available in the Chemet Lab for your inspection. They confirm the violations I have listed. You will note that, over a period of a year, I have brought these violations to the attention of my supervisors to no avail.

In our December 14th meeting, you said you would investigate my verbal allegations of falsification of records. When I questioned you on the outcome of your investigation, you stated that you found nothing. Had you personally followed up on these allegations, you would have found them to be valid, just as I stated them.

Since I am a concerned employee of General Electric and, up to this point, have been unable to get any recognition or resolution of this problem, I hereby present to you my allegations of violations in written form.

I sincerely hope that your investigation of these allegations will eventually insure that the Chemet Lab's practices and procedures will be above repute which, in turn, will enable the lab technicians to regain their sense of self esteem.

Yours truly,

17C

GENERAL ELECTRIC

STRICTLY PRIVATE

BILLER #: B-292-3850 DATE: March 20, 1986

REF ID:

REPT TO: HERO - Employee & Community Relations

ADDRESS: W/C A09, Wilmington, N. C.

SUBJECT: CHEMET LABORATORY SAFETY REVIEW

[REDACTED], Manager
Regulatory Compliance

SUMMARY

It is our opinion, based upon a review conducted from 3/8/84 through 3/21/84, that the safety program in the Chemet Lab is excellent. The following items were given specific review and all meet or exceed all OSHA, Company and safety guidelines.

1. Incident reports, maintenance work orders, and accident reports.
2. Job Hazard Analyses and Chemical Job Hazard Analyses.
3. Job Instruction Documents.
4. Lab ventilation and microwave oven monitoring.
5. Awareness and support of lab staff for safety and health practices.
6. Overall condition of lab facilities and processes.

DISCUSSION

Our evaluation of the subjects are discussed in more detail in the following paragraphs.

Review of injury, incident, and other records does not reveal any patterns or significant events that would indicate safety problems. There have been minimum injuries in the Chemet Lab. [REDACTED] bumped her head on 4/23/82; the plant physician [REDACTED] (See attachment A). 7
6
6

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions
FOIA: 6, 7C
87-878

O-15

March 20, 1984

Examination of job instruction documents and safety instruction sheets, including appropriate JHA's and CJBA's, substantiates that safety and health practices are in place and are administered. Employees have routinely reviewed and signed off on these documents. Among the documents in place and reviewed is GENERAL SAFETY O.2.1, Rev. 2. This document which employees sign off routinely gives thorough guidelines on the safety, use, and handling of chemical materials and associated equipment.

Evaluation of ventilation for the hoods and microwave oven was conducted. Routine surveys are conducted by the maintenance staff of the ventilation systems performance. Records indicate that ventilation performance meets or exceeds recommended and required specifications. Attention was focused on the isotopic and O/U preparation area. Each step in the process was closely observed. There was no visible emission of chemical materials; there were no noticeable odors of chemical emissions (oxides of nitrogen) in the immediate work area. Employees in the lab indicated that rarely are chemical odors noticed during this process. Possible emissions of fumes are controlled by the hood ventilation and the oven ventilation. Such fumes are exhausted into the main duct and removed from the lab. The crucibles containing hot solutions are allowed to cool in the oven or the hood. Once the solutions cool off, the emission of fumes from the crucible is negligible. Watch glasses are placed on crucibles which aid in fume control. Fumes from chemicals in the microwave oven will not be discharged into the room. Exhaust ventilation controls airborne materials. FMO Manufacturing Engineering measured the air volume exhausted from the oven. From this volume, it was calculated that with an opening as large as 7" x 7", an adequate air velocity would be maintained to evacuate all gases or vapors from the microwave oven. Cracks around the door are far less than an opening of 7" x 7". Therefore, gas or vapor from the oven will not pass through cracks into the room. The microwave oven is also equipped with an interlock that will not allow it to operate with the door open.

The purpose and design of local exhaust ventilation in hoods or ovens is to capture and control chemical emissions. This prevents fume releases into the work area. Such ventilation cannot and will not prevent buildup of materials on the inner surfaces of hoods or ovens.

The local exhaust ventilation and use of the documented and established work practices make it highly unlikely that employees would be exposed to even low level concentrations of harmful emissions.

March 20, 1984

Employees' overall perception of lab management's safety concern and safety emphasis is quite high. Comments like "goes overboard" and "safety conscious" were used by employees to describe lab management's attitude towards safety. One employee stated, "They don't always do what I think is best or as quick as I'd like, but there is no question as to their commitment."

Several indicated that most employees don't really understand effects of fumes and therefore are very cautious. They ask for roundtables in which short and long term effects of fume exposure could be discussed.

There was some concern expressed by one or two employees that even though the responses to safety concerns were adequate, there was a lack of sensitivity on management's part that she really understands all subjects. Employees indicated that they believe [redacted] has convinced herself that her concerns are legit and no amount of evidence to the contrary will change her mind.

The processes do not subject employees to significant safety or health risk. Employees are aware of procedures and documents relating to safe work practices and have easy access to them. Management activities as well as administrative systems are adequate to assure safety within the lab. Some elements which evidence these factors are:

- Chemical storage is orderly and safe
- Housekeeping - very clean and neat
- Layout of lab and equipment is safe and functional
- Emergency and safety equipment is visible and easily accessible
- Safety procedures and related documents are thorough and complete
- Compliance with OSHA standards is clearly evident.

[redacted]
Manager
Industrial Safety & Hygiene
NEBD - San Jose

[redacted]
Manager
Industrial Safety & Health
NEBD - Wilmington

CPS:TAP:cb

Attachment

87-0012

GENERAL  ELECTRIC

NUCLEAR ENERGY BUSINESS OPERATIONS
GENERAL ELECTRIC COMPANY • 173 CLETHOR A VENUE • SAN JOSE, CALIFORNIA 95113

STRICTLY PRIVATE

QUALITY ASSURANCE & RELIABILITY OPERATION
San Jose, California

April 26, 1984

[REDACTED] Manager
Quality Assurance
Wilmington Manufacturing Department
M/C J02 WLM

Subject: EMPLOYEE ALLEGATIONS OF VIOLATIONS TO COMPANY PRACTICES AND PROCEDURES

The enclosed report is issued as a result of an employee's "Allegations of Violation to Company Practices and Procedures" review which was held during the period of March 26 through March 30, 1984.

[REDACTED]
Consulting Engineer
Analytical & Test Methods
M/C 860, Ext. 56263

PJW:lls

Attachment

Information in this record was deleted
in accordance with the Freedom of Information
Act, exemptions 6 + 7C
FOIA: 87-88

O-16

~~STRICTLY PRIVATE~~

Quality Assurance Review of
Employee Allegations of
Violations to Company Practices and Procedures

1. Purpose and Scope

This review was performed as a result of [REDACTED] request to investigate allegations which were presented to [REDACTED] Manager, Quality Assurance, Wilmington Manufacturing Department (WMD), in a letter dated February 26, 1984, and a copy of a document titled, "Allegations of Violations to Company Practices and Procedures in Chemet Laboratory," dated February 21, 1984, by [REDACTED]. This document is attached as Attachment 1.

2. Summary of Conclusions

The four alleged violations of Company Practices and Procedures related to Quality presented to [REDACTED] were investigated by the review team and were found, for the most part, to have substance.

3. Conduct of Review


The review covered only the specific allegations presented by [REDACTED] related to Quality and was in response to request letters from [REDACTED] Manager, Chemet Laboratories, dated February 1, 1984, and [REDACTED] Manager, WMD Quality Assurance, dated March 13, 1984. The review team did not investigate beyond the scope of [REDACTED] Quality allegations.

The review team members consisted of the following:

[REDACTED] Employee and Community Relations, Team Member
[REDACTED] Quality Audit and Customer Service, Team Member
[REDACTED] Consulting Engineer, Quality Assurance &
Reliability Operation, Team Leader

-2-

An examination and investigation of WMD Instructions, Procedures, Logs, and Analytical System Readout Tapes were conducted during the week of March 26 through March 30, 1984. Interviews were conducted with the following personnel:



Pre-review and post-review meetings were held with the following management personnel:

[REDACTED] Manager, Hourly Employee Relations
[REDACTED] Manager, Qual. Audit & Cust. Serv. - Acting for E.A. Lee
[REDACTED] Manager, Employee & Community Relations
[REDACTED] Manager, Regulatory Compliance
[REDACTED] Manager, Fuel Quality
[REDACTED] Manager, Licensing & Nuclear Materials Management
[REDACTED] Manager, Chemet Laboratory

4. Allegations and Review Team Comments

a) Allegation I.A.: "Analyzers are not properly calibrated following detector change."

Comments:

- This allegation appears true if one uses the Claimant's interpretation of the applicable procedure.
- Interviews of individuals involved with analyzer operation revealed a variety of opinions as to the possible meanings of the same procedural instruction.
- Differentiation between the definitions of "Calibration" and "minimum-U-count" was not clear in the Procedural Instructions COI 411, Rev. 2 and Rev. 3.

-3-

a. (Cont'd)

- Varied interpretation of these procedures could easily create:

- inconsistencies among operators
- possible violations of procedures
- possible questionable validity of results
- possible perceptive management countenance of operator deviation from procedure
- needless loss of time because work would have to be reperformed

- b) Allegation I.A.1.d: I.A.2.e: "Calibration and verification were not completed before samples were run and material released" during 8/20/82 and 6/22/83.

Comments:

- Allegation appears correct. The investigation revealed that, during 8/20/82 and 6/22/83, six acceptable calibrations or six verifications were not run after a detector change as required by Instruction COI 411, Rev. 2 and Rev. 3.

- c) Allegation I.B.1-d: "Results from isotopics are accessible in computer and can be altered."

Comments:

- Except for one supervisor and one operator, seven individuals interviewed who work in the area near confirmed this statement.
- The investigation determined that any "altering of results" was merely the correcting of data input errors.

- d) Allegation: 1. "Use of 902/903 computer transactions are condoned by supervisor."

Comments

- Use of 902/903 transactions by someone other than a supervisor occurred when a supervisor was not present and that supervisor authorized access to the system by shift personnel. Thereafter, such individuals continued to use the system.

-6-

4. (Cont'd)

Comments: (Cont'd)

r (Even though instructions were issued on 2-22-83 and 1/25/86 to reemphasize that there should be limited access to 902/903, non-supervisor access continued) A

- a) Allegation I.C: ? ("Supervisors input data under technician's password and create false data.") P

Comments:

- Supervisors acknowledged using another person's password but only with that person's consent and in that person's presence.
- Supervisors who use another person's password merely make corrections in the input to make it more accurate. There was no falsification of data.

5. Conclusion

As the result of this review, the team has concluded that, as discussed above, () have substance.

[REDACTED]
Consulting Engineer
Quality Assurance Audit
Review Team Leader
M/C 860, Ext. 56243

[REDACTED]
Employee & Community
Relations, WMD
Team Member
M/C A09-WLM

[REDACTED]
Quality Audit
Customer Services
Team Member
M/C J28 WLH

Date	Technician	Sample #	Test	Test Values*						Assignable Cause/C
				Orig	2nd	3rd	X	S	"T"	
10/16/93		517249 ^{4000 mg/L}	ISO	2.696						QC R Standard S
SAT		517251 ^{4000 mg/L}		2.676						" "
		517255 ^{4000 mg/L}		2.675						" "
	SS	515632	" "	1.761						2nd suff. in lab
10/17/93		515718	" "	3.063						" "
		513540	" "	3.113						" "
		513485	" "	1.532						(Released by X)
		515717								

11/7/93		524596	3.30 ± 0.3124	3.212						+ QC R Standard
Mond		524599	2.90 ± 0.2058	2.890						-
11/13/93		524582	2.15 ± 0.2610	2.090						-
11/14/93		524502	3.30 ± 0.3124	3.191						-
11/15/93		525304	3.12 ± 0.3056	3.111						-
		525311	3.30 ± 0.3124	3.215	3.215					-
		524564	2.95 ± 0.2611	2.862						-
		524570	2.97 ± 0.2617	2.850						-
11/8	Tue	525312	3.30 ± 0.3124	3.226						-
		525318	3.30 ± 0.3124	3.215						-
		525553	3.95 ± 0.3118	3.827	3.971					-

This is what I copied of "Re runs".
 They were released on another
 Strange - up to the truck & re-
 removed from the lab. This
 the only time work like this
 reflected and then by some
 reason

Information in this record was deleted
 in accordance with the Freedom of Information

Act, exemptions

FOIA:

87-88

11/17/93

11/17/93

O-201701

Mr. Vice.

I am sending this copy for your information, as it is
a source from which the samples were drawn.

This was the only Gops work I could find so this
was already "filed" & stored under lock & key.

RERUN LOG

Yellow: Manager, Lab
 Pink: Measurement Control
 Gold: File

Date	Technician	Sample #	Test	Test Values*							Assignable Cause/Comment
				Orig	2nd	3rd	X	s	"T"		
10-8-83		512382	ISO	7.28	7.21	7.16	7.25		.0115		All Values accept
10-8-83		516274	"	7.27							Drift inform
10-8-83		516279	"	7.64							Drift inform
10-11-83		512470	ISO	0.743	7.58	7.60	7.537	.0093	1.15		All Values except
		512488	"	3.357	3.323	3.355	3.341	.0124	1.15		3.325 OUTLIER
		514831	"	2.258	2.221	2.207	2.229	.0091	1.15		NO OUTLIER
10-12-83		516446	"	3.734	3.83	3.742	3.707	.0103	1.15		NO OUTLIER
10-13-83		514804	"	1.349	1.331	1.326	1.333	.0121	1.15		NO OUTLIER
10-13-83		513504	"	3.267	3.161	3.178	3.200	.0569	1.15		
		514897	"	2.041	2.015	2.003	2.007	.0057	1.15		
		514824	"	1.336	1.304	1.322	1.301	.0108	1.15		All Values accept
2-13-83		516276	"	1.886	1.904	1.893	1.897	.0057	1.15		1.80 OUTLIER
10-17-83		516174	"	3.670	3.115	3.671	3.155	.0297	1.15		3.715 OUTLIER
10-18-83		513278	"	2.718	2.415	2.435	2.398	.0197	1.15		2.708 OUTLIER
		513572	"	1.347	1.335	1.322	1.324	.0125	1.15		NO OUTLIER
10-19-83		513957	"	2.198	2.157	2.201	2.235	.0045	1.15		NO OUTLIER
		5111803	"	2.046	2.120	2.112	2.109	.0122	1.15		NO OUTLIER
		513941	"	2.104	2.204	2.162	2.190	.0122	1.15		2.162 OUTLIER
10-20-83		515620	H ₂ O	5547	5523	4941	5223	513.9	1.1500		4561 OUTLIER
10-21-83		511616	"	6.284	6.158	6.20	6.153	.0061	1.1500		All Values acceptable
		513982	"	3.591	3.415	3.925	3.910	.0175	1.15		All Values acceptable
10-22-83		517249	"	2.696							OC Release
		517251	"	2.670							OC Release
		515725	"	2.577							OC Release
		515612	"	1.261							Drift inform
		515214	"	3.063							Drift inform
		513590	"	3.103							Drift inform
		515785	"	1.483							OC Release
		511531	"	6.172	6.22	6.243	6.257	.0061	1.15		6.215 OUTLIER
10-23-83		515964	"	2.014	2.13	2.20	2.184	.004	1.15		2.214 is OUTLIER
		517308	"	1.794	1.79	1.79	1.797	.0026	1.15		NO OUTLIER
10-24-83		51493	"	3.141	3.263	3.239	3.217	.0270	1.15		NO OUTLIER
2-17-83		515875	"	2.77							Not completely digest
10-18-83		517550	"	2.77							C/R Release
		51321	"	3.11	3.73	3.68	3.155	.0371	1.15		no outl/er

Date	Technician	Sample #	Test	Test Values*							Assignable Cause/Comment
				Ortg	2nd	3rd	X	s	"T"		
1/1/83		520680	Fro	1.742	1.732	1.739	1.737	.0051	1.15		NO OUTLIER
1/1/83		521754	Fro	2.251	2.223	2.237	2.220	.0140	1.15		NO OUTLIER
1/2/83		519989	150	3.725	3.734	3.722	3.720	.0063	1.15		NO OUTLIER
		522642		0.815	0.802	0.799	0.809	.0106	1.15		No outlier
		519979		3.660	3.678	3.671	3.673	.0025	1.15		No outlier
1/3		524432	150	2.347	2.328	2.318	2.324	.0137	1.15		NO OUTLIER
		520359		3.144	3.165	3.171	3.164	.0139	1.15		NO OUTLIER
		517259		3.947	3.969	3.958	3.957	.0124	1.15		Outlier was 3
		521654		1.569	1.563	1.564	1.565	.0056	1.15		NO OUTLIER
1/3/83	(516767)	150		3.834	3.833	3.845	3.823	.0104	1.15	1-990	NO OUTLIER
	521054			3.143	3.152	3.163	3.157	.0104	1.15	3.873	NO OUTLIER
	521607			3.148	3.166	3.167	3.154	.0155	1.15	NO OUTLIER	-
	519836			1.270	1.252	1.274	1.260	.0121	1.15	3.4750	1.970 is 9.8% off
	519992			3.788	3.794	3.781	3.783	.0050	1.15		No outlier
1/4/83	519319	Hg		6713	6725	6716	6740	1.915	1.15		No outlier
1/4/83	503448	Hg		5353	5150	5717	5467	2.918	1.15		NO OUTLIER

		516717		1990							
		521514		2873							
1/15/83		520564	110	3.948	3.911	3.925	3.961	.0185			NO OUTLIER
1/15/83		512270	150	2.520	2.554	2.568	2.561	.0325			NO OUTLIER
1/15/83		525711		3.042	3.049	3.037	3.044	.0482	1.15		NO OUTLIER
"		524141		3.392	3.913	3.969	3.9313	.0325	1.15		NO OUTLIER
"		526164		3.355	3.283	3.257	3.332	.0485	1.15		3.885 is 6.1%
1/16/83		526152	150	2.890	2.940	2.917	2.916	.0150	1.15		NO OUTLIER
1/16/83		524596		3.242	✓						OC. Rel [REDACTED]
<i>decided to add sample released to customer</i>		524774		2.890	✓	Wrong Ans					
		524582		2.870	✓						
		525302		3.791	✓						
		525304		3.411	✓						
		525311		2.812	2.819						
		524564		2.812	✓	Wrong Ans					
		524570		2.850	3.850						
		525751	L	2.070	2.095	2.016	2.076	.0110	1.15		2.075 is outlier

Note: I could not find "printouts" on other samples - but 360 Trans. showed they were all 100% ANSWERS

CL.

RERUN LOG

Yellow: Manager, Lab
 Pink: Measurement Control
 Gold: File

Date	Technician	Sample #	Test	Test Values*							Assignable Cause/Comment
				Orig	2nd	3rd	X	S	"T"		
11/8/83	[REDACTED]	525753	150	2.027	2.025	2.024	2.028	0.017	1.15	2.0630	IS OUTLIER
	L	525772		2.026						QC REL	
		526975		2.025						QC REL	
11/8/83	[REDACTED]	525555	150 ^{sec}	1350.62						Acceptable	
		525553		2.691	3.971					QC REL	
		525556		3.366	3.718	3.940	3.934	3.615	1.15	3.303	IS OUTLIER
11/8	[REDACTED]	514708	11	1.981	1.978	1.976	1.977	0.036	1.15	No outlier	
		525559	11	3.707	3.615	3.708	3.700	3.013	1.1535	3.611 is outlier	
11/9	[REDACTED]	524226	11	3.138	3.037	3.018	3.023	0.063	1.150	3.138 is outlier	
11/9	[REDACTED]	524314	150	2.460	2.481	2.493	2.469	0.03	1.15	1.20 outlier	
11/10/83	[REDACTED]	525766	150	2.122	2.169	2.061	2.103	0.056	1.15	No outlier	
		525224	11	2.832	2.837	2.912	2.878	0.065	1.15	No outlier	
		524305	11	2.959	2.916	2.920	2.971	0.011	1.15	No outlier	
		525231	11	2.387	3.17	2.996	2.9626	0.0656	1.15	No outlier	
11/10/83	[REDACTED]	527861	R-30	3.223	X	K	X	X	X	Single sample	
		519264	(112.30)	2.443	X	Y	Y	X	Y	Single with R-30	
11/11/83	[REDACTED]	525286	150	3.211	3.110	3.24	3.207	0.0076	1.15	No outlier	
11/12	[REDACTED]	517236	11	2.102	2.13	2.101	2.103	0.067	1.15	2.111 outlier	
		524102	11	2.214	1.930	1.921	1.923	1.910	1.15	2.214 outlier	
		524153	11	1.924	1.944	1.935	1.926	0.053	1.15	1.924 outlier	

SAMPLE #: 325553 NO. OF TESTS: 2

DATE: 7-NOV-87
TIME: 10150
TIME: 10153

SAMPLER'S PAY: [REDACTED]
TEST COMPONENT: 782
DOCUMENT NO.: PROD10.80
MATERIAL/LOT #: 39E3H5001B

DATE SAMPLED: 7-NOV-87
TIME SAMPLED: 10150
APPEL: PPU
MATERIAL CLASS: 1 UQ2
SOURCE: 1 133

NAME
CODE/ANZ: NAME
80 11 FROM ISOTOPICS
202 701 O/U IN U

RESULT: 3.971 2.057
UNITS: 3.4-238 2.0-238
TECH: [REDACTED] [REDACTED]

COMPLETED: 9-NOV-87 10-NOV-87

3.871

6

6
Wally did some reports released to him & were in office for release -
Measuring # 524591 under
done it & he seen by
that I released it - The
told him that I did not
the paper work - I knew
it was with him & for him
I had all day the
brought to [REDACTED] office
also told him about Po as I was
& Sir. Pelle sent which
said he did not want to
say so over mistake! I had to show him
second page of ISO report is will show
also released it

Here are 3 copies of
reports released to him
[REDACTED] OTR.

Talked to [REDACTED] about this
Talked to [REDACTED] about this
material - Told him that there are
some errors in this one
But let's settle this
first now then, but
some of the other is
wrong. (After you review the
these correct?)
Also, I worked on the
site at the time of

CHEMTECH LABORATORY TEST REPORT

44021 1
DATE 21-NOV-67
TIME 1015Z

SAMPLE #: 52E311 NO. OF TESTS : 2

SAMPLER'S PAY #: 997
REQ. COMPONENT #: 997
DOCUMENT NO.: FROD 00.01
CONTAINER/LOT #: 3300BS0072

DATE SAMPLED : 9-NOV-67
TIME SAMPLED : 07130
AREA : 4001
MATERIAL CLASS : 1002
SOURCE : 31289

ANALYST: DDC ANZ
607 11 FROD ISOTOPICS
205 701 O/U 12 U

RESULT	UNITS	TECH	COMPLETED
(3.215) 2.215	U-215	✓	9-NOV-67 10
	U-215		9-NOV-67 10

2.215

CHEMET LABORATORIES SAMPLE REPORT

TEST NUMBER: [REDACTED]
S. MAIL TO: [REDACTED] FOR [REDACTED]PAGE: 1
DATE: 08-NOV-83
TIME: 10152

SAMPLE #: 524570 NO. OF TESTS : 2

REMARKS:
4. SAMPLER'S PAY #1 [REDACTED]
5. REQ. COMPONENT #997 A
6. DOCUMENT NO. #1 PROD 80.96
7. CONTAINER/LOT #1 2950850017DATE SAMPLED : 8-NOV-83
TIME SAMPLED : 00130
AREA :
MATERIAL CLASS : 1 U02
SOURCE : DL31204

ANLS

CODE ANL	NAME	RESULT	UNITS	TECH	COMPLETED
60 11	PROD 1507 CFCG	1.850	% U-235	[REDACTED]	8-NOV-83 10
205 701	O/U SF U	2.070	O/U Ratio	[REDACTED]	8-NOV-83 10

V
2.850

3.95 ENR

Do not have a legible
copy



↓
(3571) In ReRun Log

6

Wiley did blame me for
missing it, but I don't think
Don is the one by the organization
that produced it - The Producer - from
Told him that he was in my class, when
the paper was. I served that a reward) released that were the
it was written and for -
I told him & he said he was writing
Brought to [unclear] about this -
Also told him already - Don didn't
& Sir. Peltier went which was very odd -
said God used his eyes (so far as he
saw) he did not want to admit his
own material! I had to show him parts
second page of ISS report, it will follow
15 minutes and it

~~2.95 ENR~~



~~(2.850)~~ In Return Log

do not have a legible copy

330

2215 In Remembrance

Do not have a legible copy

EX-391769-963

SEARCHED
INDEXED
SERIALIZED
FILED
APR 1 1964
FBI - NEW YORK

