

NOV 17 1982

Cass County Memorial Hospital
ATTN: Anton W. Kovack
Chief Executive Officer
1800 East Mechanic
Marrisonville, MO 64701

License No. 24-20234-01

Gentlemen:

Thank you for your letter dated November 5, 1982, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated October 13, 1982.

This letter also refers to the telephone conversation between Mr. W. P. Reichhold of my staff and R. A. Morgan, Medical Physicist for Clinishare on November 10, 1982, concerning your corrective actions.

It is our understanding that your corrective actions will include the following:

1. An evaluation of the technologist's extremity exposure will be made for the month of September 1982. The evaluations will be added to the technologist's permanent exposure records.
2. A quarterly review of the personnel exposure records will be performed to ensure that all the records are available.
3. A quarterly review of the dose calibrator quality control checks will be performed to ensure that:
 - (a) Linearity checks are performed quarterly.
 - (b) Linearity checks are performed for the required time period.
 - (c) Accuracy checks have been performed annually.

8211230205 821117
NMS LIC30
24-20234-01 PDR

IE HQ FILE COPY

| | | | | | | | |
|-----------|--|--|--|--|--|--|--|
| OFFICE ▶ | | | | | | | |
| SURNAME ▶ | | | | | | | |
| DATE ▶ | | | | | | | |

DMB:IE 07

NOV 17 1982

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

Robert E. Burgin
 for D. G. Wiedeman, Chief
 Materials Radiation Protection
 Section 1

cc: Robert A. Morgan
Medical Physicist

cc w/ltr dtd 11/5/82:
DMB/Document Control Desk (RIDS)

| | | | | | | | |
|---------|--------------|------------|------------|--|--|--|--|
| OFFICE | <i>RRH</i> | <i>RRH</i> | <i>RRH</i> | | | | |
| SURNAME | Reichhold:rr | Gibbons | Wiedeman | | | | |
| DATE | 11/15/82 | | | | | | |