

NOV 17 1982

03018138

Calumet Public Hospital  
ATTN: Dion Paquette  
Administrator  
205 Osceola Street  
Laurium, MI 49913

License No. 21-20242-01

Gentlemen:

Thank you for your letter dated October 29, 1982, informing us of the steps you have taken to correct the noncompliance which we brought to your attention in our letter dated October 7, 1982. With regard to the wipe test procedures described in the enclosure dated October 4, 1982, to your letter, please be advised that they are not consistent with your required procedures. As stated in License Condition No. 16 which references the letter dated July 26, 1982, you are required to decontaminate the tested area if the wipe test results exceed background (0.05 mR/hr).

We will examine these matters during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

D. J. Sreniawski, Chief  
Materials Radiation Protection  
Section 2

cc w/ltr dtd 10/29/82:  
DMB/Document Control Desk (RIDS)

8211230217 821117  
NMS LIC30  
21-20242-01 PDR

IE HQ File Copy  
DMB IE-07

OFFICE ▶	RIII	RIII	RIII			
SURNAME ▶	Douglas/cw	Matson	Sreniawski			
DATE ▶	11/10/82	11/10	11/16/82			