

Appendix

NOTICE OF VIOLATION

Chicago Osteopathic Hospital

License No. 12-04390-01

As a result of the inspection conducted on November 8, 1982, and in accordance with the NRC Enforcement Policy, 47 FR 9987 (March 9, 1982), the following violation was identified:

License Condition No. 16 requires that licensed material be possessed and used in accordance with statements, representations and procedures contained in certain applications and letters.

The referenced letter dated December 6, 1977 states in Item No. 5 that procedures for opening packages will follow that as outlined in Appendix F of the NRC NUREG-0338 Guide dated November 1, 1977.

Appendix F of NUREG-0338 Guide states in the procedures for opening packages containing radioactive materials that the exposure rate at 3 feet, the surface, a wipe of the external surface of the final source container and a survey of the packing material and packages must be performed.

Contrary to this requirement, it was found through statements of licensee representatives and the NRC inspectors review of records that the final source containers of packages containing radioactive materials were not wipe tested and that the packing material and the packages were not surveyed prior to waste disposal.

This is a Severity Level V violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each item of noncompliance: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further noncompliance; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

Dated

11/18/82

*Robert E. Burger*  
for D. G. Wiedeman, Chief  
Materials Radiation Protection  
Section 1