



INDUSTRIAL GAMMA INSPECTION

BOB SMITH — PHONE (816) 474-6252 — P.O. BOX 4041

1300 W. 7TH ST. — KANSAS CITY, MO. 64101

License No. 24-19850-01

November 8, 1982

Mr. D. G. Wiedeman, Chief
Materials Radiation Protection, Section 1
United States Nuclear Regulatory Commission
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

Reference: D. G. Wiedeman letter dated Oct. 13, 1982
Appendix, Notice of Violation

Dear Sir:

1. Violation of 10 CFR 34.28(b):
 - (1) Quarterly inspection and maintenance report forms have been procured and will be maintained in a newly created file.
 - (2) Check list will alert licensee at time each report is due.
 - (3) Full compliance achieved October 4, 1982.
2. Violation of 10 CFR 34.24:
 - (1) Calibrated survey meter intended to cover the period in question, June 7 to July 21, 1982 arrived in inoperable condition from Gamma Industries, and the time required for return, repair, and shipment to the job site carried over past the expiration date of the instrument on hand. The meter used in the interval was fully operable.
 - (2) Survey meters will be calibrated and returned within the time limit or no-work will be performed.
 - (3) Full compliance achieved August 2, 1982.
3. Violation of 10 CFR 20.401(b):
 - (1) File created to maintain records. Receiving and shipping documents for 10 CFR 71 will be used to record monitoring results.
 - (2) Check list will ensure monitoring and recording of results.
 - (3) Full compliance achieved October 13, 1982.

Sincerely,

Robert E. Smith, RSO
Industrial Gamma Inspection

RES: jlm
cc: file

8211230287 821118
NMS LIC30
24-19850-01 PDR

NOV 12 1982