

Docket No. 030-07517

University of Hawaii  
Office of the President  
Bachman Hall 202  
Honolulu, Hawaii 96822

JAN 20 1991

Attention: Mr. Ralph Horii  
Vice President for Finance and Operations

Gentlemen:

We thank you belatedly for your letter dated April 19, 1990 informing us of the steps you have taken to correct the items which we brought to your attention in our letter dated February 21, 1990. The letter was a Notice of Violation regarding the findings of our inspection conducted on January 10 and 11, 1990.

As mentioned in our telephone conversation of January 22, 1991, your proposed corrective action of amending your license to change the removable contamination limit from 500 dpm/100 cm<sup>2</sup> to 5000 dpm/100 cm<sup>2</sup> is not acceptable. The NRC guidelines that you cite as a reference gives a limit of 1000 dpm/100 cm<sup>2</sup> for removable contamination rather than 5000 dpm/100 cm<sup>2</sup>. For these reasons your current limit of 500 dpm/100 cm<sup>2</sup> must still be used. Your corrective actions will be verified during our next inspection.

Your cooperation with us is appreciated.

Sincerely,

*RJ Pate*

Robert J. Pate, Chief  
Nuclear Materials Safety and  
Safeguards Branch

bcc w/copy of letter 4/19/90:  
docket file  
State of Hawaii  
A. Johnson  
G. Cook  
B. Faulkenberry  
J. Martin  
J. Zollicoffer  
M. Smith

REGION V  
JFPang *JFP*  
1/24/91

*RJP*  
RJPate  
1/29/91

REQUEST COPY [ YES / NO ]    REQUEST COPY [ YES / NO ]

SEND TO PDR [ YES / NO ]

059087

9102060170 910129  
REG5 LIC30  
53-00017-23    PDR

*DCS*  
*JE07*