

JAN 28 1991

Docket Nos. 030-00247  
030-13728

License Nos. 20-13916-01  
20-13916-02

Holy Family Hospital and Medical Center  
ATTN: William L. Lane  
President/Chief Executive Officer  
70 East Street  
Methuen, Massachusetts 01844

Gentlemen:

Subject: Combined Inspection Nos. 20-13916-01/90-001 and 20-13916-02/90-001

This refers to your letter dated October 31, 1990, in response to our letter dated October 3, 1990.

Thank you for informing us of the corrective and preventive actions you documented in response to our correspondence. However, your letter has been reviewed and additional information is found to be needed.

In response to item F, you have proposed a corrective action which will not completely assure your compliance with the current regulations in 10 CFR 35.50(b)(3). Your proposed procedure for checking linearity is similar to those currently being considered by our Office of General Counsel as being in compliance with our regulations, however, to date no such position has been established. Please provide your corrective actions which will assure compliance with our regulations by checking linearity to 10 microcuries.

Please submit to this office within seven (7) days of receipt of this letter a written statement containing the requested information.

Your cooperation with us is appreciated.

Sincerely,

**Original Signed By:**  
**Ronald R. Bellamy**

*for*  
Mohamed M. Shanbaky, Chief  
Nuclear Materials Safety Section A  
Division of Radiation Safety  
and Safeguards

9102060185 910128  
REG1 LIC30  
20-13916-01 PDR

cc:  
Public Document Room (PDR)  
Nuclear Safety Information Center (NSIC)  
Commonwealth of Massachusetts

OFFICIAL RECORD COPY

RI HOLY - 0001.0.0  
01/28/91

RETURN ORIGINAL TO  
REGION I

050066

1E-07

Holy Family Hospital and Medical Center 2

bcc:  
Region I Docket Room (w/concurrences)

*Thompson*  
RI:DRSS  
Thompson/bj

01/28/91

RI:DRSS  
White

01/ /91

*McClan*  
*1/28/91*  
OFFICIAL RECORD COPY

RL HOLY - 0002.0.0  
01/09/91