

## LAWRENCE & MEMORIAL HOSPITAL

365 Montauk Avenue, New London, CT 06320 (203) 442-0711

December 22, 1993

Jenny M. Johansen, Chief Medical Inspection Section Division of Radiation Safety and Safeguards U S Nuclear Regulatory Commission Region 1 475 Allendale Road King of Prussia, Pennsylvania 19406-1415

Dear Ms. Johansen:

This letter refers to the letter and notice of violation which we received following an inspection of our activities carried out under our U S NRC License No. 06-09261-01. The Docket No. is 030-01275.

The following statements are in response to Appendix A, Notice of Violation. The paragraph letters correspond to the letters in that Notice.

- A. We believe that the survey meters have been appropriately calibrated on an annual basis. The 99mTc source used for calibration are calibrated against NIST-traceable sources by the vendor who supplies them. The requirement for the use of 137 Cs or 60Co sources with multimillicurie activities seems inappropriate for small medical institutions in several respects: first, the radiation exposure to personnel during calibration would be unnecessarily high; second, the survey meter would not be calibrated at the most common gamma-ray energy used in medical licensed activities (140 keV). Finally, the expense and inconvenience of acquiring a substitute meter during expensive offsite calibration is an unnecessary burden for small hospitals. Nevertheless, with regard to survey meter calibration, we will seek the services of a licensed provider of these services, or establish and implement the model procedure that was published in Appendix B to Regulatory Guide 10.8, Revision 2.
- B. Annual tests of dose calibrator accuracy tests will be reviewed and signed by the Radiation Safety Officer. This was simply an oversight.

\*

All these corrective steps have been accomplished as of December 10, 1993. The RSO will monitor performances of each of the items A and B, above.

We intend to maintain all of our activities under our NRC license in full compliance with all the terms and conditions of the license.

Sincerely,

Peter in mollay M.D.

Peter M. Molloy, M.D. Radiation Safety Officer

/dd