

NOTICE OF VIOLATION

Smithkline Bio-Science Labs., Inc.  
Farmington Hills, Michigan

License No. 21-16923-01  
Docket No. 030-11878

As a result of the inspection conducted on January 22, 1991, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1991) (Enforcement Policy) the following violations were identified:

1. License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced documents.

The referenced August 17, 1981 letter requires that weekly radiation surveys be performed of waste storage areas.

Contrary to the above, weekly radiation surveys have not been performed of the laboratory waste storage area since at least 1989.

This is a Severity Level IV violation (Supplement VI).

This is a repeat violation.

2. License Condition No. 15 requires that licensed material be possessed and used in accordance with the statements, representations, and procedures contained in certain referenced documents.

The referenced March 13, 1981 application requires that survey instruments be calibrated on an annual frequency.


Contrary to the above, the laboratory survey instrument in use has not been calibrated since November 1989.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

FEB 05 1991

Dated \_\_\_\_\_

  
\_\_\_\_\_  
Roy J. Caniano, Chief  
Nuclear Materials Safety  
Section 2

9102120042 910205  
REG3 LIC30  
21-16923-01 PDR