

APPENDIX
NOTICE OF VIOLATION

Oklahoma Civil Defense Agency
Oklahoma City, Oklahoma

Docket No. 030-07330/91-01
License No. 35-11468-02

During an NRC inspection conducted on January 7, 1991, a violation of NRC requirements was identified. In accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (1990), the violation is listed below:

License Condition 18 requires, in part, that the licensee conduct its program in accordance with the statements, representations, and procedures contained in the application dated May 14, 1990.

Item 10 under wipe test procedures, Sub-Item 6, states, in part, that the licensee will evaluate the smear with an appropriate G-M counter or a gas-flow proportional counter at a known counting efficiency.

Contrary to the above, from May 31, 1990, to January 7, 1991, efficiency of the counter had not been determined.

This is a Severity Level IV Violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Oklahoma Civil Defense Agency, is hereby required to submit to this office, within 30 days of the date of the letter transmitting this Notice, a written statement or explanation in reply, including for each violation: (1) the reason for the violation if admitted, (2) the corrective steps which have been taken and the results achieved, (3) the corrective steps which will be taken to avoid further violations, and (4) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending the response time.

Dated at Arlington, Texas,
this 2nd day of February 1991

9102130071 910201
REG4 LIC30
35-11468-01 PDR