

DEC 6 1982

APPENDIX A

NOTICE OF VIOLATION

Beebe Hospital of Sussex County, Inc.
Lewes, Delaware 19958

Docket No. 030-13331
License No. 07-17792-01

As a result of the inspection conducted on October 7, 1982, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

- A. 10 CFR 19.12 requires that all individuals working in restricted areas be instructed in the applicable provisions of the Commission's regulations and licenses.

Contrary to the above, as of October 7, 1982, individuals who were working in the restricted area had not been properly instructed in the applicable provisions of the regulations and the conditions of the license.

This is a Severity Level IV violation. (Supplement VI).

- B. Condition 14 of License No. 07-17792-01 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in an application dated September 30, 1977, and letters dated December 15, 1977, and June 19, 1978, November 7, 1978, and May 18, 1981.

Item 3 of Attachment 5, page 8, to the application dated September 30, 1977, requires daily constancy checks on dose calibrators.

Contrary to the above, as of October 7, 1982, the required daily constancy tests on the dose calibrator were not performed for periods exceeding one month.

This is a Severity Level IV violation. (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, Beebe Hospital of Sussex County, Inc. is hereby required to submit to this office within 30 days of the date of this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8212210114 821206
NMS LIC30
07-17792-01 PDR