

NOTICE OF VIOLATION

Calvin College
Grand Rapids, Michigan

License No. 21-08721-01
Docket No. 030-00820

As a result of the inspection conducted on January 16, 1991, and in accordance with the "General Statement of Policy and Procedure for NRC Enforcement Actions," 10 CFR Part 2, Appendix C, (1990) (Enforcement Policy) the following violations were identified:

1. License Condition No. 13(C) requires that each sealed source be tested for leakage and/or contamination at intervals not to exceed six months.

Contrary to the above, the licensee failed to test each sealed source for leakage and/or contamination at intervals not to exceed six months. Specifically, during the period January 1989 to September 1990, two sealed sources containing approximately 10 millicuries and one millicurie of cesium-137, respectively, were only tested once for leakage and/or contamination.

This is a Severity Level IV violation (Supplement VI).

2. License Condition No. 18 requires that the licensee conduct its program in accordance with the statements, representations, and procedures contained in certain referenced documents.

Item 14 of the referenced application dated July 31, 1978, states that the Radiation Protection Officer keeps an inventory on the amount of radioactivity presently possessed.


Contrary to the above, on the date of the inspection, the licensee failed to maintain an inventory on the amount of radioactivity presently possessed.

This is a Severity Level IV violation (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of this Notice a written statement or explanation in reply, including for each violation: (1) the corrective steps that have been taken and the results achieved; (2) the corrective steps that will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

FEB 1 1991

Dated _____



Roy J. Caniano, Chief
Nuclear Materials Safety
Section 2

9102130230 910201
REG3 LIC30
21-08721-01 PDR