

DEC 6 1982

APPENDIX A  
NOTICE OF VIOLATION

West Park Hospital  
Philadelphia, Pennsylvania 19131

Docket No. 30-3146  
License No. 37-11658-01

As a result of the inspection conducted on October 15, 1982, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

- A. 10 CFR 20.105(b) requires that radiation levels in unrestricted areas be limited so that an individual who was continuously present in the area could not receive a dose in excess of 2 millirems in any hour or 100 millirems in any seven consecutive days.

Contrary to the above, on October 15, 1982, radiation levels of 3 milliroentgen per hour existed adjacent to the wall of the Nuclear Medicine Laboratory, an unrestricted area, and had existed in that area continuously for three hours.

This is a Severity Level IV violation. (Supplement IV).

- B. 10 CFR 19.12 requires that all individuals working in restricted areas be instructed in the applicable provisions of the Commission's regulations and licenses.

Contrary to the above, as of October 15, 1982, individuals who were working in the restricted area had not been instructed in the applicable provisions of the regulations and the conditions of the license.

This is a Severity Level IV violation. (Supplement VI).

- C. 10 CFR 20.301 requires that no licensee dispose of licensed material, except by certain specified procedures.

Contrary to the above, as of October 15, 1982, licensed material on alcohol wipes had been disposed of to the normal trash, a method not authorized by 10 CFR 20.301.

This is a Severity Level IV violation. (Supplement IV).

- D. Condition 15 of License No. 37-11658-01 requires that licensed material be possessed and used in accordance with statements, representations, and procedures contained in an application dated March 10, 1978, and letters dated December 12, 1978, and December 19, 1978.

OFFICIAL RECORD COPY

8212210407 821206  
NMS LIC30  
37-11658-01 PDR

DEC 6 1982

1. Item 10 of this application requires that dose calibrators be calibrated in accordance with procedures contained in Appendix D, Section 2, of Regulatory Guide 10.8.

Item A.1 of Appendix D, Section 2, requires that the dose calibrator linearity be determined at installation and quarterly thereafter.

Contrary to the above, as of October 15, 1982, dose calibrator linearity had not been determined since June 25, 1982, a period of more than a calendar quarter.

2. Item 15 of this application requires that personnel monitor (survey) their hands for contamination after handling isotopes.

Contrary to the above, on October 15, 1982, personnel failed to adequately monitor their hands after handling isotopes. This item was identified by the inspector through independent measurements which indicated that one individual had detectable contamination on his hands and clothing of which he was unaware.

This is a Severity Level IV violation. (Supplement VI).

Pursuant to the provisions of 10 CFR 2.201, West Park Hospital is hereby required to submit to this office within 30 days of the date of this Notice a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.